

McMaster University

Report Generated: 2025-11-27 19:33:12

Device Model: Pacemaker

Device Serial Number: UNC

DCM Serial Number: Group28

Application Model: Pacemaker

Application Version: 1.0

Report Name: AAIR Parameter Report

Report Parameters

Lower Rate Limit: 60

Upper Rate Limit: 120

Maximum Sensor Rate: 120

Atrial Amplitude: 5.0

Atrial Pulse Width: 1.0

Atrial Sensitivity: 0.0

ARP: 250.0

PVARP: 250.0

Hysteresis: 0

Rate Smoothing: 0.0

Activity Threshold: 4

Reaction Time: 30

Response Factor: 8

Recovery Time: 5