

File Name: 1_D_P1_MIX_PRIME.LHC
File Location: C:\ProgramData\BioTek\Liquid Handling Control
2.22\Protocols\automation_v3.1
Last Saved: 17/09/2021 11:32:08

LHC Version: 2.19.1

Instrument: MultiFloFX
Port: USB MultiFloFX sn:19041612
Settings: Non-Autoclavable Syringes
16-tube 7° Syringe manifold
Peri-pump Dispenser
Peri-pump Dispenser 2
Random Access Dispenser

Plate Type: 384 Well Plate
Protocol Name: P2_RAD-384
Protocol Version: 1.0
Archive Revision: 47
Comments: 20 = 10 ul
Step Details: P-Purge
Peri-pump: Primary
Volume: 3000 µL/tube
Flow Rate: High
Required Cassette: Any

P-Prime
Peri-pump: Primary
Volume: 1200 µL/tube
Flow Rate: Low
Required Cassette: Any

<end>