

HD 2/12/2026

(C1)

Page 1



The United States Department of Labor Occupational Safety and Health Administration

Quality Control Data Sheet

1460 Hydrogen Fluoride (as F)

Analyst Last Name: Dinh
 Analyst First Name: Hanh
 Date Reported: 12-FEB-26
 Prepared By: KR
 Date Prepared: 29-JAN-26
 Media: MCEF, BUP Na2
 Instrument: SIE
 Method: OSHA ID-110
 Unit of Measurement: X (MICROGRAMS)
 BLK Corrected: No

1460

QC# 131268

Found: 88.000
 Theor: 90.419
 Found/Theor: 0.973
 In control: Y

QC# 131269

Found: 167.000
 Theor: 180.838
 Found/Theor: 0.924
 In control: Y

QC# 131270

Found: 341.000
 Theor: 361.677
 Found/Theor: 0.943
 In control: Y

QC# 131271 Blank

Found: 0.000

Analytes in Control: 3 Analytes out of control: 0

Average: 0.947
 LCL: 0.838*
 UCL: 1.100*
 MEAN: 0.969
 SD: 0.044
 SAE: 0.111

Precision 0.050
 Mean 0.036
 UCL 0.100*
 In Control Y*

* Indicates that the control limits are administratively set.

+ Indicates that there are no statistical control limits for this analyte, method and media and limits are defaulted to MEAN = 1.000, SD = .100 and precision UCL = .200.

This sample is beyond the statistical control limits but within administrative control limits. Administrative control limits are +/- 10% of the mean or +/- 10% of 1.000, which ever is wider.

H indicates the Found/Theor ratio is above the upper control limit, L indicates the Found/Theor ratio is below the lower control limit.

Comments: