



The United States Department of Labor
Occupational Safety and Health Administration

Page 1

Quality Control Data Sheet

1280 Fluorides (as F)

Analyst Last Name: Dinh
Analyst First Name: Hanh
Date Reported: 12-FEB-26
Prepared By: KR
Date Prepared: 11-JUL-25
Media: MCEF 37
Instrument: SIE
Method: OSHA ID-110
Unit of Measurement: X (MICROGRAMS)
BLK Corrected: No

1280

QC# 131052

Found: 96.000
Theor: 90.237
Found/Theor: 1.064
In control: Y#

QC# 131053

Found: 183.000
Theor: 180.474
Found/Theor: 1.014
In control: Y

QC# 131054

Found: 360.000
Theor: 360.947
Found/Theor: 0.997
In control: Y

QC# 131055 Blank

Found: 0.000

Analytes in Control: 3 Analytes out of control: 0

Average: 1.025
LCL: 0.819*
UCL: 1.100*
MEAN: 0.939
SD: 0.040
SAE: 0.108

Precision 0.067
Mean 0.033
UCL 0.100*
In Control Y*

* Indicates that the control limits are administratively set.

+ Indicates that there are no statistical control limits for this analyte, method and media and limits are defaulted to MEAN = 1.000, SD = .100 and precision UCL = .200.

This sample is beyond the statistical control limits but within administrative control limits. Administrative control limits are +/- 10% of the mean or +/- 10% of 1.000, whichever is wider.

H indicates the Found/Theor ratio is above the upper control limit, L indicates the Found/Theor ratio is below the lower control limit.

Comments: