

HD 2/12/2026

Description

Precision Y# on Fluorides (as F) (1280) result for OSHA ID-110.

Investigative Steps with Results

The Found/Theoretical ratios for Fluorides (1280) on QC131052 – QC131054 are 1.064, 1.014 and 0.997 respectively.

Precision results were outside statistical limits but within administrative limits. While the precision result appears to be acceptable, statistically it is outside the control limit.

Identified Cause (s)

High recovery of QC131052 might contribute to broadening precision results.

The reason is unknown. All QCSMs and samples are treated the same.

Remedy & 91b Qualification

Continue analyzing and keep eyes opened for similar occurrences in the future. No data qualification.

Resolution

Attached report to QC Data sheet in data packet.

HD 2/12/2026