

pH Adjustment for Fluorides by Ion Selective Electrode

Method OSHA ID-110

IMIS 1280 as Fluoride IMIS 1460 as HF

ate Analyzed: 2/10/2026	Room Temperature: 71° F
Analyst: HD	Slope: 96.0%
Instrument: "Patty" - Thermo Scientific Orion Dual Star pH/ISE meter	pH Electrode: Fisher Scientific Accumet gel-filled combination electrode

IMIS Code	Standard ID/Sample No.	Adjusted pH or Calib. pH	% Recovery
N/A	pH 4	4.01	
N/A	pH 10	10.01	
N/A	CCV pH 7	7.104	101.5%
1280	QC131052	7.644	
1280	QC131053	7.435	
1280	QC131054	7.062	
1280	QC131055	7.118	
1460	QC131268	7.321	
1460	QC131269	7.484	
1460	QC131270	7.747	
1460	QC131271	7.451	
1280	F58027	7.062	
1280	F58028	7.706	
N/A	CCV pH7	7.101	101.4%
1280	* F58729 F58029	7.368	
1280	F58030 * F58730 F58028	7.761	
1280	F58031 + F58731 F58029	7.829	
1280	F58032 + F58732 F58030	7.226	
1280	F58033 * F58733 F58031	7.398	
1280	F58034 * F58734 F5803	7.535	
1460	F58027 F58027	7.603	
1460	HD 2/18/26 } F58028 F58028	7.512	
1460	F58029 F58029	7.742	
1460	F58030 F58030	6.986	
N/A	CCV pH7	7.110	101.6%
1460	HD 2/18/26 } F58031 F58031	7.507	

* HD 2/18/2026.

pH Adjustment for Fluorides by Ion Selective Electrode

Method OSHA ID-110

IMIS 1280 as Fluoride IMIS 1460 as HF

IMIS Code	Standard ID/Sample No.	Adjusted pH or Calib. pH
1460	F58032	7.697
1460	F58033	7.646
1460	F58034	7.085
N/A	CCV pH7	7.145

% Recovery

102.1%