

Results of Fluoride Analysis by Ion Selective Electrode

Method OSHA ID-110

IMIS 1280 as Fluoride IMIS 1460 as HF

Date Analyzed: 2/12/2026	
Analyst: HD	Standard Addition Volume: 500 µL
Instrument: "Patty" - Thermo Scientific Orion Dual Star pH/ISE meter	F ⁻ Electrode: Thermo Scientific Orion Fluoride electrode – 9609BNWP
Beginning Room Temperature: 71° F	Ending Room Temperature: 71° F

IMIS Code:	Sample/Standard Volume (mL)	Standard or Sample No.	E ₀ mV	E _F mV	Conc. mg/L	Comments	% Rew very
-	40	Blank	123.7	21.3	0.220		
-	40	STD 1 µg /mL	83.3	19.0	1.06	Slope= -59.0 mV/dec	106%
-	40	STD 10 µg /mL	26.2	5.9	9.94	(-54 to -60 mV/dec)	99.4%
-	40	ICV 5.0 µg /mL	43.8	12.3	4.97		99.4%
-	40	CCV 5.0 µg /mL	43.4	12.2	5.06		101.2%
-	40	RLV 0.5 µg/mL	99.2	20.0	0.565		113.0%
1280	50	QC131052	73.0	27.0	1.92		
1280	50	QC131053	56.1	23.0	3.66		
1280	50	QC131054	38.8	17.0	7.20		
1280	50	QC131055	135.5	31.3	0.164	< RLV	
1460	50	QC131268	70.6	22.8	1.76		
1460	50	QC131269	54.1	19.3	3.34		
1460	50	QC131270	36.0	13.4	6.82		
1460	50	QC131271	133.7	27.0	0.148	< RLV	
1280	50	F58027	138.2	31.3	0.147	< RLV	
1280	50	F58028	136.2	31.5	0.161	< RLV	
N/A	40	CCV 5.0 µg /mL	43.8	12.4	5.00		100.0%
1280	50	F58029	133.2	31.3	0.180		
1280	50	F58030	136.7	31.5	0.158		
1280	50	F58031	135.5	31.4	0.165		
1280	50	F58032	136.2	31.2	0.158		
1280	50	F58033	136.0	31.6	0.163	< RLV	
1280	50	F58034	135.9	31.4	0.162		
1460	50	F58027	137.1	27.0	0.130		
1460	50	F58028	139.5	27.0	0.118		
1460	50	F58029	139.7	27.1	0.117		
1460	50	F58030	139.3	27.3	0.120		
N/A	40	CCV 5.0 µg /mL	43.9	12.4	4.97		99.4%
1460	50	F58031	138.3	27.2	0.124		
1460	50	F58032	138.9	27.2	0.121	< RLV	
1460	50	F58033	140.8	27.2	0.113		
1460	50	F58034	139.4	27.2	0.119		
N/A	40	CCV 5.0 µg /mL	44.0	12.6	5.00		100.0%
N/A	40	RLV 0.5 µg/mL	99.8	20.2	0.556		111.2%

* End every 10 sample runs with CCV 5.0 ug/mL. End total analysis with CCV 5.0 ug/mL and RLV 0.5 ug/mL.