

Fluoride Stock Standard

OSHA ID-110

Analyst Initials	HD
Preparation Date	7/10/2025
Expiration Assigned Date:	7/10/2026
Diluent	18 MΩ DI H ₂ O
Reviewer	7/11/2025 X Jan Ottowicz
	Signed by: JAN OTTOWICZ

STD ID	Final Stock Volume (mL)	NaF Mass Weighed (g)	Manufacturer	Bar Code	Lot #	Chemical Expiration Date	Stock Conc. (ppm)
F-STD-04	500	1.10563	Fisher	STD02189-0002-001	214832	10/14/2027	1000

Intermediate Standard is prepared for each analysis from the Stock Standard as described below:

Std	HAD Program	Dilution Volume (mL)	Final Concentration (ppm F-)
Intermediate Standard	10X_10ML	50	100

Working Standards are prepared for each analysis from the Intermediate Standard as described below:

Std	HAD Program	Dilution Volume (mL)	Final Concentration* (ppm F-)
Cal 10	5X_10ML	20	10
CCV	10X_10ML	20	5
Cal 1	50X_10ML	20	1
RLV	100X_10ML	20	0.5

*Final concentration shown on the Working Standards table is after the addition of T-T Buffer. All working standards will have a final volume of 40 mL.

* Multiple sets of the same working standard will need to be prepared because of standard addition procedure that requires a fresh sample for analysis. Standard addition is done from the 1000 ppm F- Stock Standard solution.