

Fluoride Stock ICV

OSHA ID-110

Analyst Initials	HD
Preparation Date	7/10/2025
Expiration Assigned Date:	7/10/2026
Diluent	18 MΩ DI H ₂ O
Reviewer	7/11/2025 X Jan Ottowicz
	Signed by: JAN OTTOWICZ

STD ID	Final Stock Volume (mL)	NaF Mass Weighed (g)	Manufacturer	Bar Code	Lot #	Chemical Expiration Date	Stock Conc. (ppm)
F-ICV-04	500	1.10594	Sigma Aldrich	STD02190-0001-001	MKCK9071	6/21/2026	1000

Intermediate ICV is prepared for each analysis from the Stock Standard as described below:

Std	HAD Program	Dilution Volume (mL)	Final Concentration (ppm F-)
Intermediate ICV	10X_10ML	50	100

Working ICV is prepared for each analysis from the Intermediate ICV as described below:

Std	HAD Program	Dilution Volume (mL)	Final Concentration* (ppm F-)
ICV	10X_10ML	20	5

* Final concentration shown on the Working Standards table is after the addition of T-T Buffer. All working standards will have a final volume of 40 mL.

* Multiple sets of the same working standard will need to be prepared because of standard addition procedure that requires a fresh sample for analysis. Standard addition is done from the 1000 ppm F- Stock Standard solution.