

Sterile Lot Record

Sterile Lot Number: M26-179

Contract Sterilizer: ISL

B.I Lot Number: BAD-026

Chamber: C

Part Number	Description	Lot #	COC QTY	Test QTY	Released QTY
MTUUX400-K	Tibial Base Plate, Size 4	V11022719	38	-	38
MTUUX200-K	Tibial Base Plate, Size 2	V11022713	72	-	72
MTUUX200-K	Tibial Base Plate, Size 2	V09022702	46	-	46
MTUUX700-K	Tibial Base Plate, Size 7	V11022725	27	-	27
MTUUX200-K	Tibial Base Plate, Size 2	V11022712	55	-	55
IN-UFPSRB00	FREEDOM Infinia FEMORAL SIZE B, PS RIGHT	V13012718	38	-	38
IN-UFPSLE00	FREEDOM Infinia FEMORAL SIZE E, PS LEFT	V13012717	46	-	46
RTUUX300-RK	Stemmed Tibial Base Plate, Size 3	V10162730	36	-	36
RTUUX300-RK	Stemmed Tibial Base Plate, Size 3	V10162729	41	-	41
RTUUX200-RK	Stemmed Tibial Base Plate, Size 2	V10162715	48	-	48
RLPSXE517-RK	Tibial Articular Surface PCK, Size E5-6, 17mm	V08232705	28	-	28
UPUUX837-K	All Poly Patella, Size 37mm	V07052711	198	-	198

Total Rejected: 0

Tested: 0

Total Released: 673

Comments: *Released under deviation 25.17*

Verification following shipment from Sterilizer: JK Date: 15-Apr-2026

(Part # / Lot # / Quantity)

Authorized Signature, Operations or Designee

Lot Approved for Release: [Signature] Date: 13 May 2026

Authorized Signature, Operations or Designee

**DOCUMENTATION RELEASE VERIFICATION FORM
(Packaging Documentation)**

Sterile Lot/Batch Number: M26-179

Supplier is to attach this form to documentation send to Maxx for sterile release purposes.

Verify that the...	Check box and add initials next to the box at the time of completion for each item.
- Labels are correct per MXO-PP00001 and internal labeling procedure. (Qty, Description, Exp date)	☒ <i>am</i>
- CoC review (matches labels, total qty is correct lot #s Part #s)	☒ <i>am</i>
- Assembly Record is attached (If applicable)	☐ N/A ☒ <i>am</i>
-IFUs (correct IFU or IFUs were used and are listed on CoC)	☒ <i>am</i>
AMV or Deviations related to this specific sterile batch (If applicable)	☐ N/A ☒ <i>am</i>
NCs related to this specific sterile batch. (If applicable)	☐ N/A ☒ <i>am</i>

Documentation released to Maxx by: *Pantash. Pannar, Pannar*
Authorized Signature, Quality Assurance Representative or designee

Dated: 08.04.2026

STERILE PRODUCT RELEASE VERIFICATION FORM

Sterile Lot/Batch Number: M 26.179

Attach this form to substantiate the release of the above noted sterile lot.

Document	Verify that the...	Check box to verify bulleted items are complete.
Assembly, Packaging and Labeling.	- Labels are correct per MXO-PP00001	☒
	- IFU Part number and revision are correct per MXO-PP00002	☒
	- COC matches labels and IFU	☒
	- Assembly Record (If applicable)	☐ N/A ☒
Summary of process sterilization EO EO	- Sterile Lot Number is correct	☒
	- Batch number is recorded (and matches) on every page	☒
	- Chamber ID is recorded	☒ N/A ☐
	- BI Lot Number is recorded	☒ N/A ☐
EO- Conditioning Record	- Conditioning times are within validated ranges	☒ N/A ☐
	- Humidity injection/evacuation is within the validated range	☐ N/A ☐
EO- Aeration Record	- Dwell times are within validated ranges	☒ N/A ☐
	- Product has been aerated for at least the minimum x hours	☒ N/A ☐
	- Min and Max temps are within the validated range	☒ N/A ☐
EO- Biological Indicator Sterility	- Test Results are found to be: Sterile.	☒ N/A ☐
EO- Limulus Amebocyte Lysate (LAL) Test Per FDA Requirements	- Test Results are found to be: Acceptable	☐ N/A ☒
GAMMA- Product dose mapping GAMMA- Sterilization dose	- Product dose mapping was performed and within range	☐ N/A ☒
	- Sterilization dose is within validated range.	☐ N/A ☒
Process Report	- Report has been signed and dated	☒
	- All phases have been processed	☒
	- Check for non-conformities or process deviations	☒
Post Sterilization Inspection	- Check for any open non-conformances.	☒
	- Final lot quantities on Sterile Lot Record and COC are correct	☒

Sterile lot released for distribution by: 
Authorized Signature, Quality Assurance Representative or designee

Dated: 13 May 2020