

SENSORS & PUMPS

An hour of someone else's hands near your devices.

Compression lows

Pressure on a CGM can show a sudden low that isn't real. Confirm with a finger prick

Don't dismiss it

A real low can look like compression. When in doubt, prick

No oil

Dexcom and Abbott warn lotion stops adhesive holding. Keep a hand's width clear

Show them

Point out the sensor and infusion set before you lie down

Pump in reach

If you disconnect, agree with your team how long — and reconnect before you leave

Alarms

A sensor alarm mid-session is not an interruption to apologise for