Scaling and Normalization

Pietro Franceschi | pietro.franceschi@fmach.it

Variable Scale

Metabolomics aims at measuring the concentration of :

1. The larger (as possible) number of metabolites
2. ... over the larger (as possible) range of concentration

We typically are dealing with huge differences in signal across the different variables

- potential issues in the reliability of some measurement

Variable Scale

Metabolomics aims at measuring the concentration of :

1. The larger (as possible) number of metabolites
2. ... over the larger (as possible) range of concentration

We typically are dealing with huge differences in signal across the different variables

- potential issues in the reliability of some measurement
- no direct effect on a univariate analysis

Variable Scale


Metabolomics aims at measuring the concentration of :


1. The larger (as possible) number of metabolites
2. ... over the larger (as possible) range of concentration

We typically are dealing with huge differences in signal across the different variables

- potential issues in the reliability of some measurement
- no direct effect on a univariate analysis
- strong effect on multivariate analysis

Same data different scales





Take home


- In the multivariate space the "shape" of the sample cloud depends on the absolute scale of the variables

Take home

- In the multivariate space the "shape" of the sample cloud depends on the absolute scale of the variables
- All multivariate methods based on the distance in that space will be affected by the variable scale (e.g. PCA, Clustering, PCoA, ...)

How to correct for that

- Variable transformation (*e.g. log or sqrt*)
- Variable scaling (*e.g. unit variance*)



Notes

1. The best choice depends on what you want to see

Notes

1. The best choice depends on what you want to see
2. Increasing the weight of small signal is not always the best option

Notes

1. The best choice depends on what you want to see
2. Increasing the weight of small signal is not always the best option
   - low signals are less reliable

Notes

1. The best choice depends on what you want to see
2. Increasing the weight of small signal is not always the best option
   - low signals are less reliable
   - missing values pop-up there

Notes

1. The best choice depends on what you want to see
2. Increasing the weight of small signal is not always the best option
   - low signals are less reliable
   - missing values pop-up there
3. Log transformation (and sqrt to a lower extent), often also correct for the non-normal distribution of data

Sample Normalization

The overall metabolite concentration in one or more samples can be different from the others. The sample will show-up as an **outlier**.

1. There is a real difference in that sample. **Good!**
2. The samples show different levels of dilution (e.g. urine). Bad!
3. The signal is lower for "analytical" reasons. Bad!
   - lower extraction of metabolites
   - reduce response of the instrument (in particular for MS based techniques)
   - ...

Normalization is used to compensate for the unwanted difference in sample response


In terms of data matrix

1. Normalization is performed acting on the **rows**
2. Scaling is performed acting on the **columns**

Normalization strategies

1. Use **chemical standards** to compensate for analytical issues
2. Quality controls!
3. Wisely plan (**randomize!**) your analytical sequence to avoid biases
4. Normalize to the overall signal ... mmm ...)-:
5. Probabilistic Quotient Normalization (PQN)

Compositional Data



Sum normalization creates a new (...fake..) biomarker!

PQN

The idea behind Probabilistic Quotient Normalization is to identify a **consensus** normalization factor taking into consideration the distribution of the variable specific factor for each sample

The Recipe


1. Identify a reference sample
2. For each one of the other samples and each variable calculate the ratio with the reference
3. For each sample use the median of the distribution of the ratios to estimate the consensus normalization factor

In action

- Consider 4 samples where we measure 100 metabolites
- For three samples out of the four the value of each metabolites is extracted from a gaussian distribution with mean 200 and variance 10
- For the fourth sample, the mean is 100 and the variance is 10


This design simulates a drop in the response of my pipeline of a factor of 2 while measuring the forth sample

Simulated data



Let's take the first sample as reference ...

Distribution of ratios



Code ...

```
PQNratios %>%
  group_by(name) %>%
  summarise(median = median(value))
```

```
## # A tibble: 3 x 2
##   name  median
##   <chr>  <dbl>
## 1 s2     1.00
## 2 s3     1.01
## 3 s4     0.498
```

- So to make S4 comparable to the others I should divide the signal by roughly 0.5
- The use of the median ensures that the system will work also in presence of a few biomarkers