



VRIJE  
UNIVERSITEIT  
BRUSSEL



Master thesis submitted in partial fulfilment of the requirements for the degree of  
Master of Science in de Ingenieurswetenschappen: Computerwetenschappen

# BRAIN CONTROLLED WHEELCHAIR

An introduction to brain-computer  
interfaces and a viability study of  
affordable, generalised  
brainwave-controlled systems

Lennert Bontinck

2021 - 2022

Promotors: Prof. Dr. Geraint Wiggins & Prof. Dr. Kevin De Pauw

Advisor: Arnau Dillen

**Faculty of Sciences and Bioengineering Sciences**





VRIJE  
UNIVERSITEIT  
BRUSSEL



Proefschrift ingediend met het oog op het behalen van de graad van Master of Science in de Ingenieurswetenschappen: Computerwetenschappen

# HERSENGESTUURDE ROLSTOEL

Een inleiding tot brein-computer interfaces en een verkenningstudie van betaalbare, gegeneraliseerde hersengolf-gestuurde systemen

Lennert Bontinck

2021 - 2022

Promotors: Prof. Dr. Geraint Wiggins & Prof. Dr. Kevin De Pauw

Copromotor : Arnau Dillen

**Faculteit Wetenschappen en Bio-ingenieurswetenschappen**



# Abstract

This master thesis explores the field of brain-computer interfaces (BCIs). First, it aims to provide a great foundation for the knowledge required for working in the BCI field as a computer scientist. To accomplish this, an exhaustive literature review in the introductory chapter aims to provide a great general introduction to the field and current state-of-the-art as well as challenges and promises of the field. A chapter on biomedical signals (biosignals), the source of data for BCI systems, is also provided. It discusses how electroencephalography (EEG) can be measured and provides an overview of common hardware, issues and more.

Next, the viability of real-world applications using classification algorithms on live EEG measures collected from affordable BCIs hardware is explored. This is done by first introducing a general BCI pipeline and discussing all of its components. Afterwards, a three-signal control system is proposed as proof of concept (POC) based on this general BCI pipeline. Special care is given to include all important details of the system, in an attempt to improve reproducibility. The system is also evaluated taking into account best-practice techniques whilst also realising the BCI field lacks standardized testing strategies.

**TODO: this abstract should be further completed after the thesis is finished.**



# Acknowledgements

I'm extremely grateful to both my supervisor Prof. Dr. Geraint Wiggins and advisor Mr. Arnau Dillen for allowing me to write this thesis with a subject tailored to my interests. This thesis would not have been possible without their helpful advise, constructive criticism and insightful suggestions. Their patients cannot be underestimated and the organised biweekly meetings were simply invaluable. I must also thank Prof. Dr. Jef Vandemeulebroucke for teaching both the biomedical signals and imaging course as well as the clinical decision support systems course. The contents of these courses has contributed to a better understanding on some of the physical phenomena and used technologies addressed in this thesis. I'd also like to recognize the statistical and writing expertise gained through the methods of scientific research course by Prof. Dr. Bart de Boer. Lastly, I gratefully acknowledge the unparalleled support of my girlfriend, parents and friends throughout the sometimes stressful process of this master thesis.

**Todo: complete with additional people if need be.**





# Contents

<b>Abstract</b>	<b>iii</b>
<b>Acknowledgements</b>	<b>v</b>
<b>I Understanding Brain-computer interfaces</b>	<b>1</b>
<b>1 Brain-Computer interfaces</b>	<b>3</b>
1.1 Introduction to this chapter . . . . .	3
1.2 Growing scientific and commercial interest in BCIs . . . . .	4
1.2.1 BCIs have gained big-tech interest and funds . . . . .	5
1.2.2 Improved brain-signal measuring facilities . . . . .	8
1.2.3 More powerful, affordable and portable equipment . . . . .	14
1.2.4 Specialized data processing techniques . . . . .	16
1.2.5 Summarizing the cycle of increasing popularity . . . . .	22
1.3 Common use cases for BCIs . . . . .	23
1.3.1 Preventing, monitoring and controlling diseases with BCIs . . . . .	23
1.3.2 Using BCIs to extend existing medical systems . . . . .	24
1.3.3 Using BCIs as replacement for unusable interaction methods . . . . .	25
1.3.4 Commercial and other use cases . . . . .	29
1.4 Opportunities and obstacles for BCI research . . . . .	29
1.4.1 Motivating examples using consumer-grade BCIs . . . . .	29
1.4.2 The potential of an AutoML variant for BCI pipelines . . . . .	35
1.4.3 A lack of standardized testing . . . . .	35
1.4.4 Challenges from the highly interdisciplinary nature of BCI systems . . . . .	35
1.4.5 Difficulties with repeatability and reproducibility of experiments . . . . .	35
1.4.6 Complex data variability and user-training . . . . .	36
1.5 Ethical challenges for BCIs . . . . .	36
1.5.1 The return of the Luddites . . . . .	36
1.5.2 Advertisements based on your thoughts . . . . .	36
1.5.3 Hacking BCI systems . . . . .	36
1.5.4 Changing peoples personal identities . . . . .	36
1.5.5 Painfully confronting users with their brain . . . . .	36
1.5.6 E-waste inside your skull . . . . .	36
1.6 Chapter conclusions and proposing a three-signal system for basic controls . . . . .	36

<b>2</b>	<b>Origin and acquisition of biomedical signals</b>	<b>39</b>
2.1	Introduction to this chapter . . . . .	39
2.2	Origins of biosignals . . . . .	39
2.2.1	Electrical biosignals . . . . .	39
2.2.2	Non-electrical biosignals . . . . .	39
2.3	Biosignals from the brain . . . . .	39
2.3.1	Anatomy of the brain . . . . .	39
2.3.2	Brain waves . . . . .	39
2.3.3	Measurable brain-activity . . . . .	40
2.3.4	Generalisation issues of brain activity . . . . .	40
2.4	Measuring brain-signals . . . . .	40
2.4.1	Measuring modalities . . . . .	40
2.4.2	Motivation for using non-invasive EEG . . . . .	40
2.4.3	Standards for EEG measuring systems . . . . .	40
2.4.4	Comparison of available EEG measuring equipment . . . . .	40
2.4.5	Common EEG artefacts . . . . .	40
2.5	Chapter conclusions . . . . .	41
<b>3</b>	<b>Processing brain-signals and taking actions from their interpretation</b>	<b>43</b>
3.1	Introduction to this chapter . . . . .	43
3.2	A general EEG-based BCI pipeline . . . . .	43
3.2.1	Data acquisition . . . . .	43
3.2.2	Preprocessing . . . . .	43
3.2.3	Windowing . . . . .	43
3.2.4	Feature engineering . . . . .	43
3.2.5	Classification model . . . . .	43
3.2.6	Performing an action . . . . .	43
3.3	Alternative BCI pipelines . . . . .	44
3.3.1	Using existing open-source data for training and testing . . . . .	44
3.3.2	Neglecting preprocessing and feature engineering . . . . .	44
3.3.3	No action performing step . . . . .	44
3.3.4	Offline vs online BCI systems . . . . .	44
3.4	The role of machine learning and deep learning . . . . .	44
3.4.1	Difference between traditional machine learning and deep learning . . . . .	44
3.4.2	Supervised, Semi-Supervised, Unsupervised, and Self-Supervised Learning . . . . .	44
3.4.3	Common regular machine learning classifiers . . . . .	44
3.4.4	Common deep learning classifiers . . . . .	44
3.5	Common issues when processing brain-signals and how to avoid them . . . . .	44
3.5.1	Biased data . . . . .	44
3.5.2	Incorrect or ambiguous evaluation . . . . .	45
3.5.3	No explainability or interpretability . . . . .	45
3.5.4	Overfitting . . . . .	45
3.6	Chapter conclusions . . . . .	45
<b>II</b>	<b>Development an EEG-based brain-computer interface that clas-</b>	<b>47</b>
	<b>sifies motor imagery tasks</b>	
<b>4</b>	<b>EEG-based offline classification system for motor imagery tasks</b>	<b>49</b>

<b>5</b>	<b>Moving from an offline classification system towards an online BCI system</b>	<b>51</b>
<b>III</b>	<b>Reflection on the results of this thesis</b>	<b>53</b>
<b>6</b>	<b>Using the system and verifying the results</b>	<b>55</b>
<b>7</b>	<b>Self-reflection and conclusions</b>	<b>57</b>
	<b>List of abbreviations and acronyms</b>	<b>63</b>
	<b>References</b>	<b>65</b>



# List of Figures

1.1	Funding of newer BCI related companies depicted in millions (USD). Figure based on data by Rao (2020) from 2019. It is noted this data is limited to companies that were created after 2010 where funding information is made available. . . . .	9
1.2	General components of non-invasive EEG measuring equipment. . . . .	10
1.3	The contrast between EEG measuring equipment focused on the best possible data quality and one that favours user comfort. . . . .	11
1.4	The anatomy of the human head, specifically of the cerebrospinal system. Non-invasive EEG measuring equipment is placed on the scalp, causing signals from the brain to be blocked by the skull and cerebrospinal fluid (CSF) among other structures. Free to use Figure by Blausen.com staff (Blaus, 2014). . . . .	12
1.5	Number of BCI-related papers over time. Based on data by Saha et al. (2021) obtained by searching PubMed using the keyword: "brain-computer interface". . . . .	17
1.6	General pipelines of a computer-aided diagnosis (CADx) system used for classification. A machine learning (ML) approach is called a two-step approach as it exists from feature extraction and classification. A deep learning (DL) approach is called a one-step approach as it combines both of these steps into a singular classification algorithm. . . . .	18
1.7	Summary of some of the potential reasons BCI research has seen its gain in popularity discussed in section 1.2. There is no clear chronological order for these events and they likely influence each other all at the same time. . . . .	22
1.8	Contrast between a complete control design of a robotic grasping arm that is less efficient to operate and one that is more efficient to operate but has less control. Considering the task of grasping items, both systems would succeed but the more efficient one would likely be more pleasant to use. . . . .	34
2.1	Early analog EEG equipment. . . . .	42



# List of Tables





## Part I

# Understanding Brain-computer interfaces



# Chapter 1

## Brain-Computer interfaces

### 1.1 Introduction to this chapter

Brain-computer interfaces (BCIs) are systems, consisting of hardware and software, that aim to read or even stimulate a user's brain signals for a wide variety of applications. Whilst many of these applications for BCIs revolve around providing novel interaction methods for computer applications, they are capable of fulfilling more general tasks as well. Because of this, BCIs are also referred to as brain-machine interfaces (BMIs) and can be seen as a special type of the more general human-machine interfaces (HMIs) and biological signal control (biosignal control) systems. A well-known Professor in this field is Jonathan R. Wolpaw who was also the guest editor for the first international meeting devoted to BCI research and development as part of the IEEE conference on Rehabilitation Engineering. During that meeting, a first formal definition for BCIs was given:

A brain-computer interface is a communication system that does not depend on the brain's normal output pathways of peripheral nerves and muscles.

---

*J. Wolpaw et al. (2000)*

Since then, Jonathan R. Wolpaw has (co-)authored a lot of influential papers in the field of BCIs (Daly & Wolpaw, 2008; Shih et al., 2012; J. Wolpaw et al., 2000) and created a great introductory textbook to the field (J. Wolpaw & Wolpaw, 2012). As a board-certified neurologist, Wolpaw's work is often centred around applications in a more medical setting rather than a commercial one. In this medical setting, his opinion on what defines a *perfect* BCI is often strived for and can be summarized as follows:

The perfect [medical] BCI is a safe and affordable system which works all the time, does not require the permanent assistance of a technician or a scientist, restores communication at "normal" speed, is aesthetically acceptable, is reliable and, for the same function, does not require more concentration for a patient user than what it does for an able-bodied person.

---

*Peterson et al. (2020) and J. Wolpaw and Wolpaw (2012)*

One of the things this thesis aims to study is how far BCIs have come concerning this definition of a perfect BCI. It is noted that the term communication in these definitions simply depicts the exchange of information rather than specific human communication such as speech. For example,

a computer mouse could be seen as a communication device that exchanges information about the user's intent to the computer. Many of these properties for a perfect medical BCI system would also be beneficial for commercial BCI systems.

Especially the commercial interest in BCIs has seen a recent spike, through multiple big-tech companies such as Meta (formerly known as Facebook), Valve (a major gaming company) and Neuralink (an Elon Musk company) showing interest in the field (Bernal, 2021; Facebook, 2021; Musk & Neuralink, 2019). This has given rise to the public interest for potential life-improving BCI applications as well as some public outrage on more ethical aspects that challenges these systems.

This first part of the thesis aims to provide all required background information needed for understanding BCIs. The remainder of this chapter further introduces the main rationale behind BCIs research by discussing the rise in popularity of both medical and commercial BCIs, some practical examples of BCI systems that have been developed and some of the opportunities and obstacles in the field. The chapter ends with a note on some of the ethical challenges for BCI systems and a discussion of the proposed system for this thesis. Chapter 2 gives more depth on the origins and measurability of brain signals while Chapter 3 addresses the techniques and technologies needed to process brain-signals and take actions based on their interpretation. As research on BCIs is highly multidisciplinary, entering the field as a computer scientist can be rather intimidating due to the steep learning curve of the ideas, technologies, challenges and terminology used in such research. To lower this initial learning curve, these first three chapters aim to introduce the most important concepts in an easy-to-understand manner for a typical computer scientist student with some artificial intelligence (AI) background. Although no specific procedure for systematic review was followed for these first chapters, special attention was paid to favor papers from reputable sources which have been influential based on both the amount of times the work itself is cited in other papers and the performance of connected papers. The latter was determined by using the connected papers tool<sup>1</sup>.

The interested reader is also referred to the great introductory book on BCIs by J. Wolpaw and Wolpaw (2012) and the review article by Nicolas-Alonso and Gomez-Gil (2012) when more introductory insight is desired. Whilst these resources have dated a little and state-of-the-art has changed since then, the main ideas discussed in them remain unchanged. A more recent, systematic review article by Dillen et al. (2022) focuses on current deep learning (DL) techniques for use with BCIs among other biosignal control systems.

## 1.2 Growing scientific and commercial interest in BCIs

With brain signal measuring modalities such as electroencephalography (EEG) being over 100 years old, the idea of using those brain signals for a wide variety of use cases has been explored for many decades (Berger, 1929; Haas, 2003; Kübler, 2020). With feasibility studies of using BCIs already existing in the 1970s (for example by Vidal, 1973) showing that most of the ideas explored today are not new, a clear spike in both scientific and commercial interest can be seen after the 2000s. It is no coincidence that the first international meeting devoted to BCI research and development as part of the IEEE conference on Rehabilitation Engineering discussed in Section 1.1 was also from this period.

This rise in popularity can be explained by several events. Perhaps most importantly is the improvement of both brain signal measuring equipment and computational processing equipment in both efficiency, accuracy and portability. Recent improvements in machine learning (ML) and DL after some AI winters between the 1970s and the 1990s are bound to also have played an

---

<sup>1</sup><https://www.connectedpapers.com/>

important role. The interest of big tech companies such as Neuralink, Meta and Valve have also introduced unseen amounts of funds further accelerating BCI research.

This section focuses on discussing these most important contributing factors to the new rise of interest in BCI research. These factors are discussed in an arbitrary order, as most of them have influenced each other and it is hard to name a singular reason that explains this recent rise in interest. For a more in-depth overview of the rich history that BCI research has, the reader is referred to the work by Kübler (2020).

### 1.2.1 BCIs have gained big-tech interest and funds

Big tech has been catching on with the possibilities BCIs bring, and the amount of money they can earn from it. Although profitability is an important factor in most medical applications as well, the focus of medical applications lies on improving the life of a patient, whilst the focus of commercial applications can differ greatly. Since commercial BCIs are still in their early stages and the idea of constantly wearing a brain-signal recording headset has not been accepted by the wide public yet, many commercially oriented companies start with products that are a cross between medical and commercial applications.

Most noteworthy of these more commercially oriented companies is Neuralink, an Elon Musk company. Neuralink’s initial white paper discusses its aim to create a scalable high-bandwidth BCI system, focusing on its mechanical achievements (Musk & Neuralink, 2019). These mechanical achievements are rather impressive, with state-of-the-art robot surgery inserting ultra-thin sensors directly into the skull allowing for a sleek and visually pleasing package that is mostly hidden from the human eye. Comparing this to non-invasive methods of recording brain signals, which are methods that don’t require inserting machinery into the human body, the signal quality is also expected to be far greater. However, an invasive approach currently introduces added health risks and more ethical challenges making non-invasive methods often more suited for general use (Dadia & Greenbaum, 2019; Dillen et al., 2022; Jawad, 2020). Since the publication of the Neuralink white paper, the company has held live demos of their BCI implanted directly into the skulls of animals such as pigs and monkeys. A video by Neuralink of a monkey playing pong using brain signals as input<sup>2</sup> has gathered over 6 million views on YouTube already. Combined with many news articles, the kind of exposure that Neuralink has gotten is unseen compared to the regular exposure of literature in the field. This can be questioned, as earlier work by Ifft et al. (2013) demonstrated monkeys taking control over two avatar arms simultaneously, a task that is arguably even harder to accomplish than simply playing pong. Adding to this, the experiment by Ifft et al. (2013) has an appropriate peer-reviewed paper backing it whilst Neuralink among other commercially oriented companies in the field often lack scientific backing for the claims they make. Thus, the scientific value of these more commercial demos and applications can be argued for, but the funds for research introduced by these companies and the exposure to the field have accelerated research in the field and helped popularize the field. Adding to this, the proposed system by Neuralink is one of the most aesthetically pleasing compared to alternative invasive or non-invasive systems on the market, which is one of the properties of Wolpaw’s perfect BCI system given in Section 1.1.

Besides Neuralink, companies like Meta, Valve, Neurable, InteraXon and many more are exploring the commercial possibilities of BCIs as well. Some of the companies do this through direct internal research whilst others might provide funds for external projects (Alcaide et al., 2021; Cuthbertson, 2021; Moses et al., 2021; Stockman, 2020). There seem to be two main focuses of the technology in the commercial space. Either using the new interaction method to perform work more efficiently or using it for recreational purposes.

---

<sup>2</sup><https://youtu.be/rsCul1sp4hQ>

## Using BCIs to boost work efficiency

Meta, formerly known as Facebook, has been playing with the idea of BCIs for quite a while but has been relatively quiet about it publicly. In 2021, Meta publicly announced it had provided funds for research on the use of a BCI-system to restore speech functionalities for people suffering from anarthria (Facebook, 2021; Moses et al., 2021). The system by Moses et al. (2021) achieved an average of 15 words per minute, decoded with a median error of 25%. Whilst this might not sound impressive, anarthria is a disease which causes patients to not be able to articulate speech at all due to lost control of the muscles required for making sounds. Adding to this, people suffering from anarthria often suffer from other lost muscle control as well, making alternatives such as keyboard typing or writing impossible. Taking this into account, these results should be seen as very impressive and such a system can be life-changing for certain patients. Whilst the system by Moses et al. (2021) was invasive just like Neuralink’s system, it was far from visually pleasing. The patient was fixed in a chair and physically connected to a bulky processing unit in the form of a small server rack, which makes the system non-mobile and makes the user stand out if it were to be used in the real world where discreteness is often desired.

The system by Moses et al. (2021) is an example of one that is backed by the funds that big-tech companies have and which is mainly focused on medical applications whilst the final intention of the funding company is most likely of commercial nature. Indeed, it is not hard to imagine the commercial interest of Meta in developing a more general *virtual keyboard* to enable fast *thought to speech* or *thought to text* applications usable by the masses. In fact, during the F8 conference (Facebook, 2017) a couple of years before the paper by Moses et al. (2021), Meta stated the following:

Specifically, we have a goal of creating a silent speech system capable of typing 100 words per minute straight from your brain – that’s five times faster than you can type on a smartphone today.

---

*Facebook (2017)*

Such a virtual keyboard could replace certain speech-to-text applications already broadly used for commercial purposes. In the same blog post by Facebook (2021) discussing the funding for the project by Moses et al. (2021), it is also mentioned that Meta has interest in using BCIs for high-bandwidth interactions in AR/VR. However, Meta has been subject to multiple privacy concerns lately (Fuller, 2019; Hu, 2020). The company’s reputation has been damaged from this which doesn’t help in selling the concept of them having a BCI which allows them to read the brain activity of the users. This could explain why they have recently started to shift their focus from BCIs towards muscle-based interfaces using electromyography (EMG) (Facebook, 2021).

A recent example of a BCI being used to boost work efficiency is covered by S. Chen (2022). He discusses how Chinese researchers have been working on a non-invasive and portable system that aims to detect if a user is watching pornographic content through brain signals. When presenting fifteen male participants aged between 20 and 25 with erotic content and regular content, an accuracy of over 80% was obtained for determining whether the user was watching erotic content or not. Such a system should aid in China’s content regularisation which often bans such erotic content on their domestic social-media platforms. Current systems rely on manual evaluation by a reviewer for removing or keeping content that is flagged as inappropriate by either the community or an algorithm. Further automating this task through a brain-controlled system could boost the efficiency of this process significantly.

### Using BCIs for recreational use

Whilst some promising results have been obtained when using BCIs in commercial settings to boost work efficiency, many of the systems still lack the desired performance to become truly viable. And thus, the most prominent type of commercial BCIs are those focusing on recreational use. In this regard, the headsets by InteraXon, produced under the Muse brand, are one of the earliest examples, with their first version being released in 2014. The first iteration of this product was advertised as a meditation aid. This headset relies on measuring Theta waves in the brain, which are lower frequency waves that suggest a user is meditating. Section 2.3.2 talks about these brain wave frequencies in more detail. The actual accuracy and usefulness of these types of systems are debated, as discussed by Stockman (2020). More recently, a newer version of the InteraXon headset came to market, named Muse 2, which also aids in sleep monitoring. This relies on detecting Delta waves among other patterns to determine the sleep quality of a user. Like before, this accuracy and usefulness can be argued for. A similar commercial product for sleep tracking is available from the company Dreem, under the name Dreem 2. Dreem has received more funding, as can be seen in Figure 1.1, but InteraXon, the company behind the Muse headset, has arguably contributed more to the field. Not only was it one of the first commercial BCIs that gained media attention, but the company also plays an important role in the commercialisation of BCIs as their headsets are cheap, non-invasive and visually pleasing whilst also being widely available. Adding to this, these headsets have pretty good supporting libraries in Python among other programming languages that allow developers to use these headsets for other purposes as well. Besides InteraXon, some other companies that specialize in providing commercially usable brain-signal recording headsets exist, as will be further discussed in Section 2.4.4.

Perhaps the most promising short-term commercial use of BCIs is in combination with virtual reality (VR) and augmented reality (AR). Besides Meta's interest in this region, as discussed earlier in this section, Valve has also said it is actively researching how to use BCIs as a novel interaction system in VR games (Cuthbertson, 2021). Valve is the company behind Steam, one of the world's largest game marketplaces and they are specialized in creating games and gaming hardware as well. To achieve the goals of this project code-named *Galea*, Valve is working together with OpenBCI, a well-respected company in the BCI research field that has provided open-source hardware and software for use in BCI systems. Tobii, a company that specializes in eye-tracking software, is also working on project Galea. With a final goal of creating an open-source BCI that can be used in gaming, the anticipation for the headset has been high. However, just as with deadlines from other companies such as Meta and Neuralink, the project has been postponed multiple times. This is not surprising as the promises of what a BCI can do are near endless and initial trials often offer promising results but going to a final product has been proven to be incredibly hard due to several open issues (Dillen et al., 2022).

A final stream of money that is important to mention, is coming from militaries around the globe. The U.S. military among others is known to invest a lot of money in any form of innovation, especially related to devices that can give them a strategic edge when fighting in a war (Center for Security and Emerging Technology et al., 2020; Hunter Christie, 2022). Whilst most of this information is classified, it is known that the U.S. Department of Defense and others have shown interest in a wide variety of applications using BCIs (Binnendijk et al., 2020). Whilst one can only guess what these government organisations are developing, it is likely that over time these applications might become public knowledge and aid the research field of BCIs in creating even better systems.

### Summary of big-tech using BCIs for commercial applications

To summarize, there have been a lot of big companies showing interest in commercial BCI applications in the past few years. Some might contribute directly to the field by funding scientific research, which is often still focused on medical applications but whose results can show potential for certain commercial applications (Facebook, 2021; Moses et al., 2021; Musk & Neuralink, 2019). On the other hand, some companies are working on commercial products internally, mostly for improving work efficiency (S. Chen, 2022; Facebook, 2017) or for recreational use (Muse and InteraXon headband, Cuthbertson, 2021). These commercial products have yet to see truly successful examples, as they are either questionable in delivering what they promise (Stockman, 2020), experience delayed deadlines or are even cancelled in their entirety. Nonetheless, these companies focusing on commercial applications often have high amounts of funding and a focus on the user experience (UX) of BCIs which could help accelerate research in the field and make BCI systems more visually pleasing and accepted by the broader public. In this way, they also contribute to Wolpaw’s vision of a perfect BCI system as discussed in Section 1.1.

Figure 1.1 shows the funding of BCI-related companies founded after 2010 as a rough indication of how much money is spent on start-up companies in the field. Interestingly, from the companies Neurobionics, Muse and Neuralink mentioned in this thesis, the funding amount is in proportion to the overall popularity of that company to the wider public, although this is by no means a proven relation. It also shows that whilst academic research on BCIs doesn’t require huge funding, with open-source datasets and relatively cheap hardware available as is further discussed in Section 1.2.4, 2.4.4, and 3.3.1, creating an effective commercial product can become an expensive affair rather quickly.

### 1.2.2 Improved brain-signal measuring facilities

As brain-computer interfaces (BCIs) are a type of human-machine interface (HMI) relying solely on brain signals to operate, the measuring facilities for acquiring data of those brain signals have a direct impact on the capability of those systems.

Most BCIs rely on non-invasive measuring equipment that uses EEG as a source of data and this paper will focus mainly on such measuring equipment as well. Chapter 2 explains in greater detail what EEG and some of its alternatives are, the equipment used for acquiring brain-signal data and more. For this introduction, it suffices to know that non-invasive EEG measuring equipment measures the electrical potential difference, often in microvolts ( $\mu V$ ), between electrodes placed on the scalp.

Following Wolpaw’s definition for a perfect BCI given in Section 1.1, the recording hardware should ideally be aesthetically acceptable and shouldn’t require the assistance of a professional to install. In recent years, new developments in this hardware have made meeting these criteria more plausible, which are addressed in this section.

#### Hardware improvements in non-invasive EEG measuring equipment

Three major hardware distinction made between the electrodes used in non-invasive EEG measuring equipment is whether they are wet or dry electrodes, whether they are active or passive electrodes and whether communication to the processing unit happens wirelessly or not. When considering Wolpaw’s definition of a perfect BCI described in Section 1.1, dry-electrodes with passive amplification that connect wirelessly to the processing unit would be ideal. However, when looking at data quality, a wired wet-electrode with active amplification is best. Luckily, recent advancements have made these differences in data quality more acceptable, as will shortly be discussed in what follows.






Figure 1.1: Funding of newer BCI related companies depicted in millions (USD).  
 Figure based on data by Rao (2020) from 2019. It is noted this data is limited to companies that were created after 2010 where funding information is made available.

Wet EEG electrodes are electrodes which require an electrolytic gel to be applied between the electrode and the scalp. This gel functions as a conductor and, as discussed further in Section 2.4.4, currently allows wet electrodes to have better data quality compared to dry electrodes (Cruz-Garza et al., 2017; Mathewson et al., 2017; Tseghai et al., 2021). However, wet electrodes require the assistance of a professional to correctly apply the gel and are far less aesthetically acceptable than dry electrodes. Adding to this, the electrolytic gel could also cause allergic effects for the user. Due to the viscosity of the electrolytic gel changing over time, artefacts in measurements may also appear (Tseghai et al., 2021). These are unwanted properties and conflict with Wolpaw's vision of a perfect BCI.

Advancements in dry electrodes are making the gap with wet electrodes smaller and smaller (Cruz-Garza et al., 2017; Mathewson et al., 2017; Tseghai et al., 2021). These dry electrodes don't require the use of an electrolytic gel and given the use of an appropriate headset can be installed on the scalp without the assistance of a professional. Both of these properties are in favour of Wolpaw's properties for a perfect BCI. The main reason dry-electrodes are becoming more viable to be used in real-life environments is due to improvements in active electrode technology (Mathewson et al., 2017).

Active electrodes are electrodes which do more than just forwarding their measured voltage




Figure 1.2: General components of non-invasive EEG measuring equipment.

fluctuation to the main controller board whilst passive electrodes do just that. This is often necessary since the measured signal is of such low strength that even a short distance cable from the electrode to the main board can cause a lot of noise due to electromagnetic interference (J. Xu et al., 2017). To reduce this noise, a preamplifier is used which additionally amplifies the signal before transmission over the wire as opposed to only being amplified in the main controller board. This makes the final system less compact and more expensive but is often required in anything but lab environments, especially for wireless dry electrodes, as further discussed by Mathewson et al. (2017).

When talking about wireless electrodes, it is not the effective electrode itself that is wireless but rather the communication between the main controller board, a board to which all electrodes are connected by wire, and the processing unit such as a computer. Whilst a wireless approach allows for the creation of an aesthetically more pleasing system where the measuring hardware and processing hardware are physically separated, a wired connection will always remain more efficient and reliable. However, as discussed by Tosi et al. (2017), Bluetooth, an open standard for wireless communication, has seen extensions that are more reliable, power efficient and capable of higher transmission speeds. This has made wireless solutions more appealing in BCI systems but overall issues with wireless solutions, in general, will prevail. Most important is the risk of connection loss and a higher latency resulting in a longer time between the point a signal is measured and it is received by the computational unit.

All of these advancements have enabled companies such as Muse, Dreetm and OpenBCI to develop non-invasive, dry-electrode based EEG measuring equipment with active amplification in an affordable and often aesthetically acceptable manner. An example of such an aesthetically pleasing system is given in Figure 1.3b. As BCIs become even more popular, a heavier focus on affordability and visuals with EEG measuring equipment is to be expected. As these two properties were less important in previous medical settings where a patient would wear such equipment only when undergoing a test in the hospital. Figure 1.3 shows the contrast between a high-end medical-grade EEG recording system and one that focuses on user experience (UX).




Figure 1.3: The contrast between EEG measuring equipment focused on the best possible data quality and one that favours user comfort.

### Algorithmic improvements for non-invasive EEG measuring equipment

Whilst hardware improvements has made the collection EEG data more affordable, reliable and accurate, one important issue still remains. Even with the best active wet electrodes, The contrast between spatial and temporal resolution is enormous. EEG is known to have a good temporal resolution but rather poor spatial resolution. A good spatial resolution would mean that the measurement from electrodes corresponds only to a small, known region of the brain, typically underneath that electrode. Such a correlation is helpful as it reduces noise and increases interpretability of the signal. It also allows for fewer electrodes to be used if only the activity of certain areas of the brain is of interest.

Thus, many attempts have been made at improving spatial resolution of EEG but it has been proven to be a challenging task (Ferree et al., 2001). Besides potential noise of the measurements, this is also caused by the anatomy of the human head. Remember that the electrodes used for non-invasive EEG measuring are placed on the scalp, the skin of the human head. As shown in Figure 1.4, besides the scalp, different structures such as the skull and cerebrospinal fluid (CSF) are in between the electrodes and the actual brain. These components *blur* and *disperse* the perceived brain-signal, making it hard to track where the measured signal came from when looking at the electrical activity on the scalp.

Whilst increasing the number of electrodes placed on the skull physically limits the region under one single electrode, it doesn't guarantee an improve in spatial resolution. Indeed, clever




Figure 1.4: The anatomy of the human head, specifically of the cerebrospinal system. Non-invasive EEG measuring equipment is placed on the scalp, causing signals from the brain to be blocked by the skull and cerebrospinal fluid (CSF) among other structures. Free to use Figure by Blausen.com staff (Blaus, 2014).

processing algorithms are required to correct for overlapping signals between electrodes so that the overlapped signal is correct for and the effective spatial resolution is improved. Besides this, there is also the issue that decreasing the distance between electrodes introduces the need for placing more electrodes to cover the entire region of the brain. This increases cost, lowers user comfort and decreases the visual acceptance of the system. Another issue with increasing the number of electrodes and decreasing the spatial resolution means that the alignment of the electrodes on the skull is now even more prone to errors and change over time, e.g. due to movement of the user. This makes the need for a professional higher, which is also detrimental with respect to Wolpaw's criteria for a perfect BCI.

Ferree et al. (2001) has found 19-electrode EEG systems to have a highly varying spatial resolution in the 20 to 40  $\text{cm}^3$  range. Systems with 129 electrodes were found to have a spatial resolution of around 6 to 8  $\text{cm}^3$  (Ferree et al., 2001) when also using algorithmic tricks to further improve spatial resolution. However, according to Nunez and Cuttillo (1995), around  $10^7$  parallel pyramidal neurons reside in each  $\text{cm}^3$  of the brain cortex. This means the acquired data is still obtained from a incredibly large number of neurons even in the best spatial resolutions.

Whilst hardware improvements in both electrodes and headsets for better placement might improve the spatial resolution further, the spatial resolution improvements possible through hardware have been plateauing. As was the case for the comparison between few and many electrode systems by Ferree et al. (2001), appropriate algorithms have to be used to effectively increase the spatial resolution. Recently, these techniques often rely on using Laplacians (Liu et al., 2020; Srinivasan, 1999; Srinivasan et al., 1996), although other approaches using for example convolutional neural networks have been proposed (Kwon et al., 2019). These techniques also have the added benefit of cleaning the time-varying signal as described by Yao et al. (2019).

### **Invasive BCIs try to combat the issues of EEG**

The previous parts discussed how non-invasive EEG measurements have been improved. However, alternatives to EEG exist for measuring brain signals, some of which are further discussed in Section 2.4.1. Most notable in recent years is measuring modalities that rely on capturing brain signals by equipment directly inserted into the human body, making it an invasive approach. One such example of an invasive measuring modality is electrocorticography (ECoG) and the most popular invasive BCI at the moment is the one proposed by Musk and Neuralink (2019). The white paper by Musk and Neuralink (2019) has shown that invasive BCIs could greatly exceed the data quality and visual aesthetics of even the best non-invasive alternatives. As further discussed in 2.4.2, this invasive method places flexible electrodes directly inside the skull. These electrodes are invisible to the human eye with the only visual component being a rechargeable wireless transmitter that is magnetically attached to the skull. Neuralink's final aim is to make the brain-signal measuring equipment completely invisible to the human eye.

Musk and Neuralink (2019) has built robots to insert the electrodes inside the skull in a very precise location without the need for an open-skull operation or even anaesthesia. This allows for a magnitude more electrodes to be installed and is expected to suffer far less from noise resulting in a far greater temporal and spatial resolution compared to EEG. This suggests that invasive systems are superior to non-invasive alternatives, but the fact that they are more permanent, far more expensive and invasive gives rise to technical and ethical questions. These ethical questions are further discussed in Section 1.5. From a technical standpoint, maintaining and upgrading a non-invasive BCIs is far simpler and cheaper. The fact that you are inserting foreign objects into the brain also introduces far more health risks than non-invasive systems do. Convincing the user to put on a headset that can be removed will also be far easier than convincing the user to get a BCI permanently implemented in their skull.

It has also been shown that the theoretical more precise temporal and spatial resolution doesn't linearly correlate with improved BCI accuracy/control, rather it seems to plateau relatively quickly with current state-of-the-art signal processing and classification techniques (Aflalo et al., 2015; Lebedev, 2014). Some critics point to the dropping curve found by Aflalo et al. (2015) to conclude that the increased electrode amount and reachable neurons achieved by Musk and Neuralink (2019) don't have a direct impact on the usability of BCIs in real-world applications. Because of these aspects, the ease-of-use appeal and far cheaper price for non-invasive alternatives still outweigh the benefits offered by invasive methods for almost all but highly medical applications, at least in the opinion of the writer of this thesis. Nevertheless, future improvements in signal processing and classification techniques could prove invasive methods to be far superior for BCI applications and the mechanical achievements so far are not to be underestimated. An invasive system is also promising concerning Wolpaw's definition of a perfect BCI discussed in Section 1.1. Once installed, it would ideally require no more assistance from a professional, is aesthetically acceptable as it can be invisible to the human eye, has signs of being far more reliable than EEG and more.

### **Summarizing the improvement of measuring facilities**

Since BCIs rely solely on brain signals to operate, the measuring facilities for acquiring data of those brain signals have a direct impact on the capability of those systems. As was discussed in this section, the most commonly used modality for non-invasive data acquisition, EEG, has benefited from both hardware and software improvements. From a hardware point of view, the switch to dry electrodes using active amplification and wireless connection to a computational unit has made BCIs more favourable concerning Wolpaw's criteria for a perfect BCI (Mathewson et al., 2017; Tosi et al., 2017; J. Xu et al., 2017). From a software perspective, clever algorithms

have enabled preprocessing of the signal to improve spatial resolution (Kwon et al., 2019; Liu et al., 2020; Srinivasan, 1999; Srinivasan et al., 1996). Improving the spatial resolution can also positively affect the temporal resolution due to inherent noise reduction as discussed by Yao et al. (2019). As is further discussed in Section 3.2.2, other preprocessing techniques have also been introduced and refined further aiding in improving the data quality.

### 1.2.3 More powerful, affordable and portable equipment

The improvements in brain signal measuring equipment have likely been influential in the gaining popularity of BCIs as it provides more precise data more affordably. However, having the possibility of obtaining clean data is only part of the way to a perfect BCI system. Other improvements concerning computational power, affordability and portability have also played an important role in BCI research, contributing to the rise of popularity in the process.

#### The emergence of faster and cheaper hardware

As chapter 3 will discuss in greater detail, working with EEG data, or other forms of brain signal data can require computationally very heavy operations to achieve desired processing results of that data. Luckily, together with the improvements in state-of-the-art measuring equipment, there is also an emerging supply of less accurate but far more affordable and portable EEG measuring equipment. Due to Moore's law (Schaller, 1997) and other advancements, central processing units and other computational hardware have also seen massive improvements in computational power. This has made algorithms previously requiring expensive specialized computational hardware possible on the average personal computer. All of these factors have made BCI applications, which were previously limited to lab environments with a high financial cost, accessible to a far broader public. The availability of open-source datasets for common tasks related to brain signals has also allowed computer scientists to experiment in the field without additional hardware cost (Kaya et al., 2018).

#### Splitting BCIs into multiple major components for portability and reusability

Early attempts at making BCIs more portable and affordable include those by Lin et al. (2008) and Shyu et al. (2010). In essence, these applications rely on separating the data acquisition process and data processing into two standalone systems connected over Bluetooth. Remember from Section 1.2.2 that Bluetooth is an open standard for wireless communication that has seen improvement in the last couple of years. Dividing a BCI system in a data acquisition and data processing system allows for creating a lightweight measuring device to be placed on the user's head, with a heavier and bulkier computational unit to process the signals which ideally is still pocket-able. The latter was not a trivial task and introduced the need for custom hardware at the time. Lin et al. (2008) used a custom-made digital signal processor (DSP) for the task whilst Shyu et al. (2010) opted for a more general field programmable gate arrays (FPGA) based DSP. Whilst these were great demonstrations of how the technology could be used outside the lab, the actual usage for a bci detecting driver's drowsiness (as proposed in the paper by Lin et al., 2008) and allowing multimedia control (as proposed in the paper by Shyu et al., 2010) was rather limited. The idea of custom-made and possibly proprietary processing hardware which focuses on a single task is also very limiting, although it does have commercial benefits.

What did stick, was the idea of splitting the hardware into two standalone parts, a wireless EEG measuring device and a processing unit. As discussed in Section 1.2.2, a wireless connection between these two components is also favoured when taking into account Wolpaw's criteria on a perfect BCI. It also makes it possible for smaller research teams or even individuals with a

certain specialisation to take part in the highly interdisciplinary field by not requiring knowledge of all components but just the one that is of interest. As an example, it enables computer scientists to purchase off-the-shelf affordable EEG measuring hardware and communicate with it through provided libraries for their favourite programming language. In most cases, the personal computer they already own is powerful enough for the experiments, especially for offline systems. This allows for reusing existing hardware which is great from a financial perspective. Section 2.4.4 discusses some of the EEG measuring equipment available on the market. It is noted that EEG measuring hardware is not strictly needed for a computer scientist as researchers such as Kaya et al. (2018) have made excellent free-to-use EEG datasets available.

With the introduction of the iPhone in 2007, it didn't take long for researchers to explore the idea of using a mobile phone as a processing unit for a BCIs. Wang et al. (2011) were one of the first to explore this idea, with a steady-state visual evoked potential (SSVEP)-based BCI. Section 2.3.3 will go into further detail on the types of measurable brain-signals. In essence, such a system relies on a category of brain signals that are often easy to detect but require a specific stimulation. This type of system can be used for a wide variety of applications. Imagine an audio-guided tour in a museum where visitors only need to stare at a screen next to an item of interest to start hearing the explanation of that item. This could be achieved with only a couple of dry electrodes placed on the skull in a headset that also provides the audio to the visitor. This headset could then be connected over Bluetooth to the visitor's phone running an app for the museum tour. The technology needed for such a system would lean close to that of so-called *P300 spellers*, which have already been heavily studied (Capati. et al., 2016; Hussein et al., 2020; Won et al., 2019). Such a system would also fit perfectly with Wolpaw's definition of a perfect BCI, albeit oriented to a commercial setting rather than a medical one.

### Making BCIs a one-in-all device again for profitability

Whilst the advantages of using the computational power of devices a customer already owns are clear, it also imposes some disadvantages. For one, the varying type of computational devices is bound to give varying performance results, compatibility issues and overall limits the guarantee of a pleasing user experience (UX). Adding to this, the measuring equipment and processing equipment can't be connected from the factory resulting in an experience that is not plug-and-play. From a commercial perspective, it would be easier if the system was all-inclusive and possibly patentable.

Recent trends in computing hardware where manufacturers are shifting away from general all-purpose CPUs and then developing their own custom CPU architectures have shown that custom chips can outperform their general counterparts. Patenting the architecture of those chips is possible making it commercially interesting. Apple's mac M series processors announced in 2020 are one such recent example. These M series processors have a neural engine that is stated to accelerate the time needed for ML tasks<sup>3</sup>. Graphics processing units (GPUs) used for autonomous driving systems also differ from general-purpose GPUs.

Because of this, the author of this paper believes custom-made chips could create a future where the headset has a directly integrated processing unit once again. Whilst this would make for a more attractive package for the customer and give commercial advantages to the manufacturer, it would be disadvantageous for research purposes. The manufacturer could limit the possibilities of using the BCI for different purposes, patent promising hardware and more. Another possible route the author of this paper sees is the use of cloud computing and fast 5G connections to also create a more simple user experience that doesn't require Bluetooth tethering to a close-by processing unit. This approach would still leave a separation between measuring

---

<sup>3</sup><https://nr.apple.com/dH8i4U3v2w>

hardware and processing hardware making changes to any of the two independently easier. Concerning Wolpaw’s criteria of a perfect BCI, these approaches would also be acceptable. This belief of switching back to all-in-one devices or using a cloud service for processing the data is further endorsed by the findings of Dillen et al. (2022). In their systematic review of biosignal control systems, eight of the 46 studied papers used embedded hardware and one used cloud solutions.

### **Summarizing the improvements on computational power, affordability and portability**

To summarize, due to Moore’s law (Schaller, 1997) and other advancements, CPUs among other computational hardware have seen massive improvements in computational power. This increase in computational power has enabled more advanced processing of the data on more affordable and portable hardware. Early attempts at making BCIs more portable and affordable focused on splitting the brain signal measuring equipment from the data processing equipment (Lin et al., 2008; Shyu et al., 2010). The system by Wang et al. (2011) was one of the earliest examples of a true portable BCI-system that was affordable and relied on a smartphone as a processing unit. It showed how working with BCIs can be done using cheap and general-purpose hardware. The research was published at a turning point for BCIs where publication numbers on BCI-related papers started rising. This hints that the increased affordability and portability combined with more computational power played an important factor in the rise of interest in BCI. The rise of BCI-related papers is illustrated in Figure 1.5 based on data by Saha et al. (2021). Dillen et al. (2022) found that papers on biosignal control systems using DL have seen a steady increase over the last five years as well.

In the future, as BCIs see more commercial applications, this separation of a BCI in a measuring component and processing component might reverse to an all-inclusive device. This has potential downsides for scientific research but makes commercial sense. The replacement of physical computational units in close proximity to cloud solutions is another possible evolution.

### **1.2.4 Specialized data processing techniques**

The previous sections 1.2.2 and 1.2.3 discussed how both measuring and computational hardware have seen recent improvements. Another important part of the puzzle is the algorithms that convert data from the now more user friendly measuring devices to useful actions using the now more powerful, affordable and more portable computational hardware. Most of these algorithms are data-driven classifiers that use ML techniques. In more recent years, deep learning (DL) techniques and alternative approaches have been incorporated in the BCI pipeline as well, which has been proven to be very successful. Chapter 3 discusses commonly used techniques in more detail and multiple ML and DL based BCI pipelines are discussed in later chapters of this thesis. This section gives a more high-level summary of recent developments in the AI field that have likely contributed to the rise in popularity of BCIs.

### **Postponing another AI winter**

Machine learning (ML) and deep learning (DL) are techniques that fall under the AI umbrella. These techniques are being used as buzzwords in a whole suite of applications and it seems as if every week there is yet another big promise or threat related to AI discussed in major news outlets. Recent examples that have shown the world what new techniques in this field are capable of include the Go champion beating computer algorithm by Silver et al. (2016), the impressive text generation model GPT-3 by Brown et al. (2020) and the image generation model DALL-E






Figure 1.5: Number of BCI-related papers over time. Based on data by Saha et al. (2021) obtained by searching PubMed using the keyword: "brain-computer interface".

by Ramesh et al. (2021). This abundance of new achievements and an overall high public interest in anything that mentions buzzwords from the AI umbrella has caused a long lasting AI summer since the last AI winter of the late 1980s and early 1990s. Such an AI summer means that there is incredible amount of funding available for improving ML and DL techniques among others. This in turn causes further advancements in the field of ML and DL which results in more impressive achievements.


However, an AI summer also implies that an AI winter will inherently return. An AI winter is a period of time where the interest in the field is reduced and thus funding and research is limited. As discussed by Floridi (2020), such an AI winter may be relatively close. This is in part due to new regulations and public backlash on the more questionable but highly profitable applications DL is involved in. A recent example of this is the controversy surround Clearview AI. Here, state-of-the-art DL image recognition algorithms are used on billions of images collected from all over the internet, including social-media platforms, to recognize almost anyone with a public profile linked to them. As further discussed by Rezende (2020), this technology conflicts with many EU laws yet was used by multiple police departments. Adding to this, new regulatory changes are being proposed to limit the use of algorithms which lack explainability and interpretability (European Commission. Joint Research Centre., 2020; The Royal Society, 2019). This challenges many ML and DL approaches currently used as explained further in Section 3.5.3.

Nevertheless, there is still a high amount of resources being put into ML and DL research. Throughout history, these technologies have been linked with the biomedical setting a lot. As explained by Baldi (2018), DL and biomedical data have directly influenced each other's evolution's since the 1980s. Because of this, applications that process biomedical data have been an important factor at prolonging the current AI summer. Since BCIs use biomedical data as well, they have been one of the applications keeping interest in ML and DL research high. This is in part due to the science-fiction properties BCI systems have creating a lot of public interest


as already discussed when talking about Elon Musk’s Neuralink in Section 1.2.1. Thus, BCI systems, which rely heavily on ML and DL, are one of the research areas in these technologies that are so promising they help prolonging the current summer of AI.

### Improved and new ML and DL concepts have enabled more capable BCI systems

Most of the main concepts from both machine learning (ML) and deep learning (DL) are already multiple decades old. As DL is a subset of the ML techniques, pipelines for using these techniques are very similar. To illustrate this, a general pipeline of a computer-aided diagnosis (CADx) system used for classification is given in Figure 1.6 and commonly used techniques are discussed below. It is noted that besides classification tasks, some regression problems for CADx systems exist as well. However, such regression problems are far less common in BCI systems relying on EEG with the systematic review article of Dillen et al. (2022) only finding articles on classification problems for such systems. Because of this, this thesis focuses on classification problems. CADx systems are used extensively in hospitals for the interpretation of biomedical images and have been studied ever since computers were invented. The most common example of a CADx system is the classification of lung images as being either from a lung cancer patient or not, often also highlighting the nodules used for this classification. These pipelines are very similar to the ones used for BCIs, which is further discussed in Section 3.2.



(a) General pipeline for ML based CADx system.



(b) General pipeline for DL based CADx system.

Figure 1.6: General pipelines of a CADx system used for classification. A ML approach is called a two-step approach as it exists from feature extraction and classification. A DL approach is called a one-step approach as it combines both of these steps into a singular classification algorithm.

Traditional ML is a commonly used name for the collection of ML techniques that are not DL. Such traditional ML rely on a *two-step* process where there is both a feature extraction step and a regular classification step as is visualised in Figure 1.6a. Feature extraction is the process of representing the often highly dimensional and unstructured raw data using characteristic properties. These representations are often derived from expert knowledge and are chosen by the designer of the system rather than learned from data. With respect to a CADx system for classifying if a patient has lung cancer, these features might be the size of nodules in the lung, a metric of how round these nodules are, the Hounsfield unit (HU) of certain parts of the lungs and more. These features are then used for learning by a traditional ML classifier. The true challenge in these two-step traditional ML systems lies in finding appropriate features, which

can be a very time-consuming task that requires a lot of domain knowledge. It isn't uncommon for features in CADx systems to be refined in the span of multiple years (van Ginneken, 2017). Section 3.2.4 discusses the feature engineering step in more detail and Section 3.4.3 discusses traditional ML in more detail.

Alternatively, a *one-step* approach in CADx systems denotes the use of DL in the pipeline, such as shown in Figure 1.6b. DL generally differentiates itself from the previously discussed two-step traditional ML approach by working directly on the preprocessed data rather than a feature representation of the data. However, this isn't necessarily the case as feature extraction can still be beneficial for some DL methods whilst others might even work better on the raw, non-preprocessed, data. Section 3.4.1 discusses the difference between traditional ML and DL in more detail. Intuitively, DL models exist of multiple layers which can be seen as a graph-like structure which becomes deeper as the number of layers increases, hence the name deep learning (DL). Typically, multiple types of layers are present in a DL approach, with the earlier layers focusing on what could be seen as gradual feature extraction steps whilst the later layers are often more tailored towards the classification of those data-driven feature outputs from the previous layers. Section 3.4.4 discusses DL and the common variations in more detail. A DL approach is interesting as it can allow for skipping the feature extraction process where good features have to be found by the developer of the system. Finding such good features can not only be time-consuming but limited expert knowledge can also make creating features that carry enough differentiating information impossible, meaning that even the best classifiers can't learn from them. Adding to this, DL approaches in CADx systems have also been proven to outperform state-of-the-art ML approaches (van Ginneken, 2017). However, DL models are more challenging in terms of explainability and interpretability as further discussed in Section 3.5.3. The challenge in DL approaches is finding a suitable model and training it in a way such that it doesn't overfit, a phenomenon further discussed in section 3.5.4. In a way, DL allows a person with limited domain knowledge of the data to obtain a better result as it bypasses the need for manual feature extraction. However, this is a double-edged sword as limited knowledge of the data increases the risk of bias, overfitting and other potential issues.

Relating this to BCIs, which have a very similar pipeline, and how the ML and DL techniques used often originate from older concepts, some of the most common BCI pipeline components are considered. A typical traditional ML BCI pipeline often relies on a form of the common spatial pattern (CSP) technique for feature extraction. This technique is quite old being introduced by Koles et al. (1990) around 30 years ago. Likewise, the traditional ML classification used is often a type of support vector machines (SVM). Once again, this technique was first introduced by Boser et al. (1992) around 30 years ago. Over the years, CSP has evolved and many extensions such as filter bank common spatial pattern (FBCSP) by Kai Keng Ang et al. (2008) have been introduced. Likewise, SVM has seen many extensions and improvements (Chervonenkis, 2013; V. Utkin, 2019). This has resulted in the combination of these two relatively old techniques, but with recent extensions, still performing among the state-of-the-art in BCI applications using traditional ML approaches (Rasheed, 2021; Tangermann et al., 2012). Likewise, when using a DL approach in the BCIs pipeline, CNNs are most commonly used (Dillen et al., 2022; Rasheed, 2021). This technique is again a rather old one, the foundations of which were first described by Fukushima (1980) over 40 years ago. Just as the CSP procedure, CNNs have seen multiple extensions and improvements over the year, just as other DL approaches. Most recent and noteworthy developments are works such as that by Schirrmeister et al. (2017) which focuses on proposing CNN architectures that can learn from very few samples and for which a described method exists to visualize the model's early layers. Other improvements, more general to DL, have also been made such as the development of new activation functions. Lee (2020) discuss how changing the activation function from CReLU to ReLU6 offered a 35% performance increase

while keeping other components fixed for certain experiments relying on a neural network (NN) in a reinforcement learning (RL) setting. Thus, these improved versions of older concepts have enabled far better performance making it possible to create more capable BCI systems.

This doesn't mean that all approaches used for processing the data in BCIs rely on decades-old techniques that have improved over the years. One interesting and relatively new approach is the use of Transfer learning (TL) from drastically different domains. Previously, TL was mostly used in BCIs to train a model on data which may originate from different users performing similar but not necessarily identical tasks. This general model is then further refined on a specific patient and task, transferring the knowledge acquired from the previous data to the new data. When done correctly, this can provide far better performance compared to learning on the new data alone for problems where data is limited (Dillen et al., 2022). As available data specific to BCIs applications remains limited, some recent research has gone into transferring knowledge from completely different domains to BCI specific data. G. Xu et al. (2019) used a model pretrained on images and transferred it to EEG data for a motor imagery (MI) task with promising results. Other attempts at transferring knowledge from other domains, such as natural language, have also been made (De Wulf, 2022).

### More open-source datasets and BCI related libraries

Whilst more affordable and portable measuring hardware has enabled a low-cost solution for researchers to acquire their own data, as discussed in Section 1.2.3, the process of acquiring BCI related data remains significantly time-consuming and can impose multiple challenges. As addressed by Dillen et al. (2022) many publications don't properly report the data acquisition process, leaving a lot of ambiguity in both the meaning of data labels and how representative of the real-world the data is. Whilst two papers might discuss a MI task as being *imagined left-hand movement*, one might have collected the data in a lab-like environment from a trained user envisioning a single squeezing shut movement of the hand whilst another might have it correspond with any envisioned movement of the fingers on an untrained and non-focused user. On a legal aspect, as data on brain signals is biomedical data, heavy regulations are in place on how this data can be shared and used, next to the general data protection regulation (GDPR) (Malin et al., 2013; Vlahou et al., 2021).

Considering that many researchers in the field often focus on one specific component of the BCI pipeline, it is not feasible for them to go through the trouble of collecting data themselves. Adding to this, when a researcher wants to reuse a certain component of the BCI pipeline from another author's work, the source code of their project is often not present, not well documented, or not compatible with their programming environment (Dillen et al., 2022).

Luckily, BCI specific data and coding libraries have been made available in recent years. Some of the earliest BCI related datasets that are still commonly used to this day are from the BCI competitions organised by the Berlin Brain-Computer Interface research program (BBCI)<sup>4</sup>. Of the four different competitions, two were in part organized by Jonathan R. Wolpaw, from who this thesis took the definition of a BCI and a perfect BCI, further demonstrating his importance in the field (Blankertz et al., 2004; Blankertz et al., 2006). The fourth and final competition provides three EEG datasets labelled with MI tasks, making it a popular choice in literature. More recent datasets include those by Kaya et al. (2018), who have put a tremendous focus on discussing all necessary details of the data acquisition process, including testing the MI skills of each subject and providing the software and instructions given to the subjects. Not only do these publicly available datasets allow researchers to skip the data acquisition step, but it also improves the reproducibility of their work and makes comparing it to other work using the

---

<sup>4</sup><https://www.bbc.de/competition/>

same dataset easier. However, there is still far from an abundance of data that can be used for training BCI pipelines compared to other fields of research. Indeed, collecting datasets of books for natural language processing (NLP) learning applications or cat images for computer vision applications is a far easier task than collecting EEG datasets for identical MI tasks with equal data acquisitions methods.

From a code perspective, many authors still fail to deliver a copy of their source code along with their article (Dillen et al., 2022). Some websites, such as Papers With Code<sup>5</sup> has been created to more easily find papers that do provide their code, but for BCI research this is still limited. Luckily, advanced libraries have been emerging which provide a multitude of common operations from the BCI pipeline. Perhaps the most famous of which is the Python MNE library by Gramfort (2013) which provides tools for organizing, visualizing and processing EEG data such as windowing the EEG data, performing baseline correction and determining the CSP features. Other famous Python libraries include Braindecode by Schirrmester et al. (2017) and EEG-DL by Hou, Jia, et al. (2020, 2022), Hou, Zhou, et al. (2020), Jia et al. (2020) among others. Whilst Python is the most commonly used programming language for implementing BCI pipelines and has by far the most supporting libraries, libraries for MatLab (e.g. EEGLab by Delorme and Makeig, 2004), C++ (e.g. Brainaccess by Neurotechnology<sup>6</sup>) and other popular programming languages are starting to emerge as well. Besides these general libraries, some of the most famous articles whose source code isn't provided have also seen open-source implementation based on the original author's description. Most noteworthy is the Army Research Laboratory (ARL) EEGModels Project which provides implementations of the EEGNet model proposed by Lawhern et al. (2016) and the ShallowConvNet and DeepConvNet models proposed by Schirrmester et al. (2017) written in Keras and Tensorflow (Python DL libraries by Chollet et al., 2015; Martín Abadi et al., 2015).

### Summarizing the emergence of specialized data processing techniques

Besides improvements on a hardware level, both for the measuring equipment and the processing equipment, significant software improvements have also been made. Most importantly are the improvements made to two-step traditional ML approaches and single-step DL approaches. This includes improvements of the CSP feature extraction algorithm such as FBCSP by Kai Keng Ang et al. (2008), improvements to the SVM classifier (Chervonenkis, 2013; V. Utkin, 2019) and general DL improvements (Dillen et al., 2022; Lee, 2020; G. Xu et al., 2019). The introduction of more open-source datasets and libraries also aids in far faster development of new BCI pipelines and allows researchers to focus on a specific component of the BCI pipeline to improve. The fact that BCIs still provide a science fiction feeling to the general public has also made it one of the technologies that aid in postponing another AI winter.

Relating this to Wolpaw's definition of a perfect BCI, it becomes apparent that these software improvements play an important role in achieving the goal of a perfect BCI. Better pipelines should allow a BCI system to function more reliably and in more challenging environments. Improved techniques could also improve the number of classifications a system can handle in a certain period. This can enable the BCI system to match or even surpass the conventional interaction method it wishes to replace. Whilst these are very important aspects of a perfect BCI, it should be remembered the user of a BCI will most likely not appreciate this evolution to the same degree a computer scientist will. At the end of the day, an ideal processing pipeline should be one the user never notices is there, whilst a wrong classification that results in the wrong command being executed or other issues with the pipeline will result in unpleasant experiences

<sup>5</sup><https://paperswithcode.com/>

<sup>6</sup><https://www.neurotechnology.com/brainaccess-documentation/C++Api>

the user will remember.

### 1.2.5 Summarizing the cycle of increasing popularity

When looking at the number of papers published on PubMed with the keyword *brain-computer interface* visualised in Figure 1.5, a clear upwards trend is visible over the years. This upward trend seems to have started in the 2000s and the jump was most significant around 2012. Whilst proving which elements are responsible for this upward trend in both scientific and commercial interest isn't directly possible, this section has highlighted multiple potential reasons. It is unlikely any one of these potential reasons is the sole reason for the increased popularity of BCI systems. Rather, all of the discussed reasons likely influence each other as portrayed in Figure 1.7.




Figure 1.7: Summary of some of the potential reasons BCI research has seen its gain in popularity discussed in section 1.2. There is no clear chronological order for these events and they likely influence each other all at the same time.

The potential reasons that were discussed can be separated into four major categories. First, an increased interest by big-tech companies causes big funds to get into the field. Secondly, improvements in measuring equipment allow for higher quality data and a more pleasant user experience (UX). Third, processing equipment has seen increased affordability whilst also improving the computational power and portability allowing systems to be more affordable and capable. Finally, more optimised software allows for better classification results and faster devel-

opment of new systems. It is highly unlikely all of these reasons have equal influence or that they are the only reasons for the improved popularity, but they should give the reader a better understanding of the context BCIs are currently in without becoming too technical. The next section focuses on effective use cases for BCIs to let the reader understand the different capabilities of these systems even more.

## 1.3 Common use cases for BCIs

Previous sections addressed some of the recent commercial applications from BCIs. Whilst these are great examples of how the technology can be used by the general population, the medical applications of BCIs remain very important to the field, if not of most importance. EEG signals originate from the medical field and many of the successes in the field are related to helping people with certain diseases and/or disabilities. This section will highlight some of the major areas where BCIs can improve the quality of life for these people.

### 1.3.1 Preventing, monitoring and controlling diseases with BCIs

EEG trials are often used to diagnose brain disorders. Especially seizure-related disorders such as epilepsy are often diagnosed through EEG readouts. But also other neurological disorders such as the locked-in syndrome rely on EEG amongst others to be correctly identified. Even the psychological field uses EEG as a physiological measuring tool to aid in the diagnosis and treatment of patients, although psychiatrists should be aware of the limitations EEG has for them (Badrakalimuthu et al., 2011).

As EEG is used for a wide variety of diagnoses and as input data for most BCIs, much research has gone into how BCIs can be used as an aid for medical diseases. Abdulkader et al. (2015) gives an overview of the most common use cases for BCIs. In general, BCIs is used for three different reasons in medical applications: prevention, detection/diagnosis and rehabilitation/restoration.

One popular example of prevention relates to traffic accidents. According to World Health Organization (2018), traffic accidents were the number 1 cause of death for children and young adults. Whilst many types of equipment are already in place to prevent traffic accidents from occurring, such as speed cameras and breathalyzers, it has been studied how BCIs can be used as a prevention measure as well. The earlier discussed research by Lin et al. (2008) that was one of the first attempts at a truly portable BCI, measured driver's drowsiness levels. Such information can be used to alert or even enforce drivers to take a break from driving when drowsiness levels are too high. Other research includes that by Fan et al. (2012) who has discovered that emergency situations can be classified faster from EEG data than the user's response. This was done through a driving simulation and was found to have an accuracy of about 70%. Chin-Teng Lin et al. (2013) developed a system that could estimate motion sickness levels, which in turn could again be used as an alert for drivers to take a break from driving. These three examples show how BCIs can be used in a variety of ways for the prevention of traffic deaths. Many more applications exist for the prevention of this phenomenon and others using BCIs.

Computer-aided detection (CADE) and CADx using medical imagery are widely used for diagnosis and treatment monitoring in oncology and other fields. Comparable systems exist using EEG data amongst others for diagnosis aids. Whilst these do acquire and process brain signals to output a classification, the classification itself is not directly used nor is the output directly connected to other systems. This makes calling diagnostic systems using EEG data a stretch of the definition in some regards. Nonetheless, the previously given example of the commercial Muse headset for sleep tracking is one example of a detection and diagnostic system that can be classified as a BCI. It detects irregular sleep patterns, allows for diagnosing certain

sleep disorders and can interact with the user to prevent or suppress the found phenomena in the future. Taran and Bajaj (2020) have developed a similar system to detect sleep apnea from EEG, but don't call their system a BCI. Poorvitha et al. (2020) also developed a system to detect sleep apnea but they do call it a BCI. This shows that the definition of a BCI is not strict. However, all of these systems make use of similar methodologies compared to more clear BCI systems.

Restoration of lost mobility and environmental control is where BCIs shine. As most prostheses rely on electromyography (EMG), a muscle-based biomedical signal (biosignal), for controlling them, they do not apply to patients who don't have muscle control anymore. Luckily, BCI are often used as novel interaction methods using brain signals, which enables such patients to regain previous levels of mobility and control. Section XXX give some more insight on how BCIs are used regain lost mobility and environmental control. As Daly and Wolpaw (2008) discuss, BCIs can also be used to guide patients in rehabilitation through brain plasticity. Brain plasticity, or neuroplasticity, is the brain's ability to adapt itself based on experience. BCIs could show patients which regions of the brain are used and which types of brain signals are present. This information could then be relayed to the patient through various means so that it induces neuroplasticity. Such systems are still in development and require sophisticated neurological expertise that falls outside the scope of this master thesis. The overview provided by Daly and Wolpaw (2008) provides a great starting point for further literature on this manner. In a similar manner, pedaling motor imagery (MI) tasks have been studied recently as they seem to have a great potential for lower-limb recovery. The works by Cardoso et al. (2021) and Cardoso et al. (2022) discuss such systems in greater detail and are also an example of how BCIs can be used for rehabilitation.

It is noted that many more BCI applications exist for preventing, monitoring and controlling diseases than those discussed. The work by Adama and Bogdan (2021) is another great example of how BCI can be used for rehabilitation. The work by Shih et al. (2012) also highlights some related BCI work in medicine.

### 1.3.2 Using BCIs to extend existing medical systems

BCIs can be used in a wide variety of applications. Even in non-trivial domains, they can find their usages as a standalone system or as an extension to existing systems. To demonstrates the latter, this section discusses how BCIs can be used as an extension to classical hearing aids for an improved user experience. A more trivial extension to existing robotic prostheses and exoskeletons is also addressed.

According to Seol et al. (2020) over 450 million people suffer from disabling hearing loss. Most solutions to hearing loss rely on a microphone to capture environment audio which is then amplified and played through a speaker that is placed in or near the ear. This microphone can be integrated inside the speakers and thus form a stereo setup located at the ears of the patient. This is not always ideal when there is a lot of ambient noise. Sometimes using an external directional microphone placed close to the audio source of interest can form a solution. This can be a microphone placed on the desk of a professor teaching in a filled room or the speakers connected directly to an audio source such as a television. However, this solution is not applicable in all situations. Thus, most hearing aids include some noise suppression on the microphones directly to filter out ambient noise and amplify noise coming from human speech. Wong et al. (2018) evaluated such noise suppression for Mandarin-speaking users and found the results to be good but not ideal. Da Silva Souto et al. (2016) have shown that a BCI can be used to determine which speaker a user is listening to by analysing directional queues. This information is useful, as it can optimize the microphones to pick up speech from that area and algorithms could optimize for the sounds the user is focusing on. It is noted da Silva Souto et al.



(2016) discuss how a long waiting time to determine the area of interest challenges the practical usability of their system as of now. Nonetheless, it shows one of many non-trivial ways a BCI could be used as an extension of existing systems to improve them.

Perhaps the most studied and promising extension BCIs can fulfil in existing medical systems lies in the interaction with robotic prostheses or exoskeletons. Most of the current robotic prostheses and exoskeletons rely on muscular activity in the body. This muscular activity can then be measured by electromyography (EMG), the data of which can be used to control the robotic prosthetic. For example, patients who have lost (part of) their arms but have functioning muscular activity in the remaining body part, can use this remaining muscular activity to control robotic prostheses. Sudarsan and Sekaran (2012) discuss the design and development of such a system based on EMG. Some of the processing techniques are similar to those of EEG driven algorithms. Alternatively, when the limb is still intact but the control over (part of) the limb is lost or extra support is needed, an exoskeleton may be used. Just like robotic prostheses, most exoskeletons rely on EMG. A thesis by the German Fleischer (2007) highlights the fundamentals of EMG based exoskeletons.

As was already touched upon in Section 1.3.1, EMG measurements are not applicable for all patients. In particular, people who have neurological diseases limiting the production of the required muscular based biosignals fall outside the scope for these solutions. However, due to the developments in BCIs, the viability of robotic prostheses and exoskeletons for these patients has been steadily on the rise. AL-Quraishi et al. (2018) give an in-depth systematic review of upper and lower limb exoskeletons and robotic prostheses controlled by EEG-based BCI. AL-Quraishi et al. (2018) address the high risk associated with failed instructions for robotic prostheses and exoskeletons. Indeed, compared to a misclassification with P300 spellers, the risks that can follow from misinterpreted instructions of exoskeletons and robotic prostheses are of such a degree that even high accuracy systems might not be good enough. AL-Quraishi et al. (2018) also highlight that whilst multi-label classification of EEG is possible with considerable accuracy in an offline lab setting, the number of detectable classes is limited in a real-time and real-life environment. Because of this, EEG-based systems in these applications still have some challenges to overcome to match the precision and reliability of EMG counterparts. Whilst improvements regarding these aspects have been made since the work of AL-Quraishi et al. (2018) was published, the main challenges remain to this day, especially when using affordable systems. Because of this, widespread adoption of EEG-based exoskeletons and robotic prostheses is still very limited.

### 1.3.3 Using BCIs as replacement for unusable interaction methods

The previous section highlighted how BCIs can be used to extend existing medical systems. When working with exoskeletons and robotic prostheses, the BCI systems provide a novel interaction method for these devices that normally rely on EMG or other sources for input. Many of the successful BCI applications consist of providing novel interaction methods with existing systems. It could be argued that almost all of the literature proposed BCI applications boil down to making a novel interaction method of some sort. Opposed to the discussed EEG-based exoskeletons and robotic prostheses, whose risk currently leads to limited real-life applications, less risk imposing BCI applications do find their use as novel interaction methods in real-life already.

The P300 spellers discussed earlier are examples of novel interaction methods that aim to replace keyboards, especially for those who don't have the required capabilities to operate them (Hussein et al., 2020; Won et al., 2019). Whilst wrong classifications in a speller application could result in unpleasant situations, it is clear that the risk involved is far smaller than exoskeletons for example. As the name suggests, P300 spellers make use of *P300 signals* which are a type of event-related potential (ERP). As is further discussed in Section 2.3.3, a P300 signal is a

positive bio-electrical wave measurable with EEG around 300ms after a stimulus occurred. The stimulus used is often a flashing pattern on a monitor, of which there are multiple shown in a matrix form. The users have to focus on the element of the matrix that they want to select and appropriate algorithms can extract this selection from EEG data. Guy et al. (2018) performed a usability study on 20 amyotrophic lateral sclerosis (ALS) patients in a real-life like environment. According to Guy et al. (2018), most participants achieved over 70% accuracy, which is in line with the findings of Hussein et al. (2020) and Utsumi et al. (2018) in similar studies amongst other types of patients. More interestingly, even though the accuracy wasn't extremely high, all participants of the experiment by Guy et al. (2018) succeeded in the given tasks. This is in part due to our ability as a human to understand typo's in words and sentences relatively easily. Another important factor is that most of these systems make use of auto-correct software to help combat faulty classifications. Besides auto-correct software, state-of-the-art text prediction is also a crucial part of these systems as it can almost double the efficiency of P300 spellers. Guy et al. (2018) has found that the use of word predictors raised the mean number of correct symbols per minute from 3.6 to just over 5. Whilst 5 symbols per minute is still a lot slower compared to regular keyboard input, it enables useful communication for those who can't communicate through regular means. As Utsumi et al. (2018) have shown, P300 spellers can be used by ALS patients with satisfactory results in a pleasant manner for the user. Utsumi et al. (2018) have shown equal results for duchenne muscular dystrophy (DMD) patients. In general, P300 spellers are often used as they have a low learning curve that can quickly enable a patient to regain some communication skills. The reason these systems have a low learning curve and can be quickly deployed is due to their simple user interface and a combination of a system that generalizes well and that has been studied thoroughly for the use of TL. Early examples of using TL for P300 related BCIs include those by Kindermans et al. (2012). It is noted that eye-tracking is a viable alternative for P300 spellers in most cases and research has been put into combining the two, such as the work by Kalika et al. (2017).

The focus on user experience by Utsumi et al. (2018) is an important factor of a real-life case study. User experience is often overlooked by initial BCI system proposals, where the focus is often on numerical measures such as accuracy, speed and false negative (FN) or false positive (FP) rates. However, good working BCIs can cause psychological burden and other side-effects on the patients using them which an initial proof of concept (POC) might not reveal. This gives rise to some ethical challenges, some of which are discussed in Section 1.5. As time passes, the user might move to a more capable and sophisticated system that has a steeper learning curve, higher cost and is more demanding for the user.

As P300 signals are relatively easy to detect through EEG, it is one of the most studied signals from event-related potentials (ERPs). Many other BCI applications are based on P300 signals then just spellers. One such example is the *Facebrain* application by Warren and Randolph (2019). Facebrain provides an EEG-based novel interaction method with the social media platform Facebook. In essence, it's a regular P300 speller with the first screen(s) representing possible actions to take on the platform. When text input is required, a regular P300 speller user interface is presented. This allows a user to operate almost all of Facebook's functionalities with only a P300-based BCI. The application by Warren and Randolph (2019) is one of many that shows the same strategy and classification algorithms as P300 spellers can be used for a wide variety of applications by changing the meaning and functionalities of the shown matrix elements. The matrix elements could even be overlapped on top of an image to have intuitive motion control, a methodology used by Kapgate (2022) for their P300 controlled quad-copter.

ERPs and the measurable signals they produce, such as the P300 signal, are only one of many sources for detectable brain signals. In general, ERP related signals are easier to detect reliably, as the stimulus can be controlled, giving a hint when and where to look for signals and what to

look for. However, P300 and ERPs in general also have their issues and limitations. The BCI handbook by Nam et al. (2018, Chapter 26) discusses the crowding effect, adjacency problem, repetition blindness and user discomfort amongst other issues ERPs have. Most problems arise from the often limited space for sending stimuli without overlap and the changing behaviour of both the brain and participant's experience after a prolonged session where many stimuli have been applied.

An alternative to ERPs is using a mental phenomenon called MI as source of signals for a BCI system. MI is the process in which a person generates brain activity in the motor cortex merely by imagining motor movements. Section 2.3.3 explains in further detail how MI is not dependent on an external stimuli nor actual motor movements. This makes MI-based BCIs extra appealing as they don't require external stimuli and are applicable for people with motor disabilities. Pfurtscheller et al. (1997) were the first to experiment with the idea of using MI in an EEG classification task. Since then, many MI-based BCIs have been proposed. Cong Wang et al. (2011) proposed a MI-based BCI to control a robot arm system. Their research is interesting in two ways. First, they use only three distinct MI classifications: imagined right-hand movement, imagined left-hand movement and imagined foot movement. These three controls enable the user to select eight different possible actions through a menu where two options are always shown that can be controlled using either an imagined left-hand movement or an imagined right-hand movement. Scrolling through the menu to show two other possible actions is possible through the imagined foot movement. This shows that with the right system design few controls can still allow for many actions to be taken. Secondly, they found that experienced users have better overall classification performance which indicates that MI is something that can be trained.

A more recent and more complex MI-based BCI system is the vehicle control system by Zhuang and Yin (2017) which recognizes four possible actions: left, right, throttle and brake. There are three very interesting aspects in the work by Zhuang and Yin (2017). First, they use two distinct classifiers for the EEG data. One makes a distinction between left and right through a typical MI-based back propagation neural network (BPNN) whilst the other classifies throttle and brake behaviour using the subject's threshold value of the average band power. Secondly, an additional system is in place to reduce the risk of wrong classifications in the system. This additional system is a type of collision detection and avoidance system that uses four ultrasonic wave radars and a camera. The rationale behind this additional system is that the car would operate more like a semi-autonomous system that is responsible for a safe ride whilst the input of the BCI system is used to steer this semi-autonomous car in the right direction. This additional system is required as the accuracy of around 84% for the throttle and brake classification and 89% for the left and right classification is not enough for a reliable system. Finally, they use an interesting data collection method to train the classifiers on a user-per-user basis. They configured a driving simulator where the user has the freedom to perform any action they want through a classical steering wheel and pedal setup. The user should synchronously think about the action they want to perform to generate MI data and they have to perform the effective action, as to be able to label the data. Whilst this is an interesting approach, it is limited in the fact that it requires the user to be able to operate a steering wheel and pedals at the same time, which is not the case for classical target users of these systems.

Far more types of brain signals exist for use in BCIs than those discussed in this section. Section 2.3.3 will go into further detail on the different types of brain-signals that can be used. For a more exhaustive list of all different approaches to using BCIs as novel interaction the reader is advised to consult one of many systematic reviews, books and other materials already available (Kübler, 2020; Nam et al., 2018; Nicolas-Alonso & Gomez-Gil, 2012; Panoulas et al., 2010; Shih et al., 2012; Sonam, 2018). From the works these materials discuss, it should become apparent that most short-term goals of BCIs still lie in improving the quality of life for people

with disabilities. However, the rise in popularity of BCIs in the gaming industry and amongst other big tech companies as discussed in Section 1.2.1 shows there is a potential future where BCIs find more real-life use-cases in other fields as well. Many of the current research also looks into combining multiple technologies to limit risk, increase the number of classification classes, create a more pleasant training experience and more. It is important to stress that the research field of BCIs is not as new as some of the more commercial companies and publishers would like to portray it is. Because of this, some of the claimed breakthroughs should be taken with a grain of salt. One such example is the use of BCIs with monkeys. Neuralink, the earlier discussed Elon Musk company, has published a YouTube video<sup>7</sup> in 2021 demonstrating a monkey playing pong using the BCI by Neuralink. The video quickly gained millions of views, far more than most papers in the BCI field will ever reach. As a result, this video without a backing paper or proper explanation of the process was covered as highly innovative and groundbreaking by many news outlets. However, similar setups have already been made in the past. For example, Ifft et al. (2013) demonstrated monkeys taking control over two avatar arms simultaneously, a task that is arguably even harder to accomplish and has an appropriate peer-reviewed paper backing it.

It was argued in Section 1.3.2 that the risk involved when working with exoskeletons and robotic prostheses is too high for current BCI systems to be reliably used. The BCI controlled car from Zhuang and Yin (2017) discussed in Section 1.3.2 demonstrated how risk of a BCI system can be greatly reduced by using an additional system responsible for only allowing safe actions. Similar ideas could be used for robotic prostheses to reduce the risk involved. B. Xu et al. (2022) proposed such a hybrid system to control a robotic arm not only through a MI-based BCI but also by using obstacle avoidance algorithms to reduce the risk of harmful contact, computer vision for object detection to get a better idea on the wanted interaction and eye-tracking to gather extra information surrounding user's intention. Hybrid systems like the one by B. Xu et al. (2022) are very promising as they can greatly reduce the risk involved in many BCI systems, such as prostheses related applications, whilst also increasing the overall accuracy of the system.

Whilst limb prostheses such as robotic arms are one of the most common types of prostheses, they are only a fraction of all prostheses in existence. Everything from dentures and hearing aids to artificial breasts can also be labelled as a type of prostheses. Visual prostheses such as bionic eyes are another type of prostheses and they are being studied heavily in the BCI field. Not only can BCI systems improve visual prostheses, many of the existing visual prostheses could be seen as a special type of BCI system as a whole. Both the works by Ptito et al. (2021) and Niketeghad and Pouratian (2019) give an overview on the progress in visual prostheses in the BCI field. These BCIs are often invasive that can stimulate the brain and other parts of the body, as opposed to only reading brain activity. Through the simulations or other means, the user can regain some form of vision from these BCI systems. Second Sight is one of few companies that has commercially made visual prostheses with Food and Drug Administration (FDA) approval. It is discussed in the overview on BCI-related vision restoration systems by Niketeghad and Pouratian (2019). The international trial by Humayun et al. (2012) on the products of Second Sight shows promising results, although it is noted the study is performed by Second Sight employees and not by an independent research team. The exact working of visual prostheses or, more specifically, Second Sight products is not of interest for this work, but the recent decisions of Second Sight company reveal one of the largest risks of invasive BCIs and BCIs in general. Due to the discontinuation of some of the Second Sight products, hundreds of users are left without product support for a system that shaped their everyday life. Besides this, the now non-functioning product is still present inside their body. The issues and ethical questions this brings to the table are discussed further in Section 1.5.6.

---

<sup>7</sup><https://www.youtube.com/watch?v=rsCul1sp4hQ>

### 1.3.4 Commercial and other use cases

TODO

## 1.4 Opportunities and obstacles for BCI research

As technology evolves, it has often been the case that people with certain disabilities benefit from the evolution as well. Take the evolution of live captions as a recent example, it has found its way directly into the Android smartphone operating system. Whilst initially implemented for users to enjoy video content in situations where they can't listen to the audio, people with hearing difficulties now have a direct way to enjoy more content too. The push for autonomous cars will allow those that are traditionally unable to drive a car to finally enjoy the freedom a car can offer as well without being dependent on others to do so. Object and scene recognition algorithms made for optimizing smartphone cameras can also be used as a way of describing what is on a photograph for those who have limited vision.

Whilst it is clear from the previous sections that research in BCIs has its roots in medical applications, more commercial applications are currently being explored as discussed in Section 1.2.1. These commercial applications have a larger potential audience, which will probably result in more sales. In general, hardware and software that is produced and used on a larger scale cause market competition which in turn often lowers prices whilst increasing quality. This is bound to also make medical-oriented applications of BCIs more accessible and of better quality. It would also allow for smaller creators with a limited budget to create meaningful, quality-of-life improving applications. This has been the case with smartphone applications for a long time. Simple applications such as a colour picker in the camera app can aid people who have colour blindness in determining whether a banana is ripe or not. Small applications by individual developers relying on location changes can remind those with short-term memory loss if they have not forgotten essential things such as locking the door.

More and more open-source deep learning (DL) libraries tailored towards EEG such as the ones by Schirrmester et al. (2017) and Lawhern et al. (2016) have been introduced. These DL libraries can often work on raw EEG data directly, which allows developers who don't have much domain experience to make interesting and good performing classification pipelines or even make new discoveries in the field. Publicly available EEG dataset such as the one by Kaya et al. (2018) makes it possible for developers to work on a pipeline without needing the monetary investment in EEG measuring equipment. If desired, the affordable headsets by OpenBCI and others, further compared in Section 2.4.4, make it possible to have EEG measuring equipment with good open-source libraries for well under 1000 euros. These smaller projects do not only have the potential to be of big impact for very specific users, they often reveal common issues and interesting alternative approaches in the BCI field.

### 1.4.1 Motivating examples using consumer-grade BCIs

As discussed above, good open-source libraries, datasets and even measuring hardware have made the entry to the BCI field easier than ever before. Some of the open-source DL libraries tailored towards EEG classification work well on raw EEG data, which allows for meaningful classification accuracy without the need of understanding all domain-specific properties of EEG data required for classical preprocessing. The question remains whether or even if, real-life applications are possible on consumer-grade BCI systems. In an ideal world, such applications should be fully configurable and usable by the end-user without the help of a domain expert and the system should be maintainable without exhaustive preliminary knowledge. Many works have

demonstrated the potential of these BCI systems but the lack of widespread adoption suggests there are still hurdles to overcome for them to become a reality. For this reason, this thesis focuses on providing a solid foundation to the BCI field, a POC application to demonstrate how a working system can be implemented and a thorough viability discussion of these systems in their current state.

Multiple studies comparing different aspects of these cheaper consumer-grade systems to the more traditional medical-grade BCI systems have already been done, as is further discussed in Section 2.4. In general, cheaper systems work with a lower electrode count and those electrodes also have a lower signal quality. Whilst this means the hardware is of lower quality, cheaper consumer-grade systems can reach usable levels of accuracy as shown by Frey (2016) and Rashid et al. (2018) for the Texas Instrument ADS1299 chip, commonly found in consumer-grade hardware such as the offerings from OpenBCI. Similarly, studies have been done on different types of electrodes as already touched upon in Section 1.2.2 and further discussed in Section 2.4.4. These electrode comparison studies have shown that the gap between easy-to-use and affordable dry electrodes and the more superior but harder to use and often more expensive wet electrodes is shrinking (Cruz-Garza et al., 2017; Mathewson et al., 2017; Tseghai et al., 2021). However, these findings often follow from very controlled experiments in lab-like environments. This is to be expected, as most comparisons aim to eliminate as many random factors as possible. However, this means that these experiments don't give great insight into the real-life usability and applicability of these cheaper BCI systems. This section will focus on some work that demonstrates which real-life applications are possible on consumer-grade BCI hardware and how they are possible. These works were motivational examples for this master thesis. They give a more thorough vision of what types of applications could be possible with consumer-grade systems in their current form, often highlighting the issues that are still present with them. Most of these works are centred around OpenBCI hardware, the manufacturer of some of the more successful low-cost biosignal measuring devices as further discussed in Section 2.4.

### **Step-by-step implementation of a binary MI classification model by Peterson et al**

Peterson et al. (2020) aimed to show the feasibility of a complete low-cost consumer-grade BCI system. Peterson et al. (2020) did this by discussing the steps required to make an offline binary MI classification system using common low-cost consumer-grade hardware. The classification distinguished the MI task of a grasping movement with the participants' dominant hand and a rest condition. They compare three different approaches in growing complexity: CSP, penalized frequency band common spatial pattern (PFBCSP) and penalized time-frequency band common spatial pattern (PTFBCSP). The work by Peterson et al. (2020) has five interesting aspects worth highlighting here.

First, whilst they did use the OpenBCI Cyton and Daisy board they did not use the 3D printable Ultracortex Mark IV headset from OpenBCI. They argued that this is due to the Ultracortex Mark IV headset becoming uncomfortable quickly due to the use of dry electrodes combined with limited adjustability. This complaint on user comfort for the Ultracortex Mark IV headset is recurring with other authors including the one from this master thesis. Because of this, they opted for wet EEG electrodes attached to a very flexible and far more comfortable Electro-Cap. Peterson et al. (2020) opting for wet electrodes is slightly odd, considering the user experience of having to apply a conductive gel before each use is not great and therefore not favourable for general real-life BCI applications.

Secondly, for the data gathering of their system Peterson et al. (2020) used a *common office room*, rather than a lab-like environment. Except for a 3D printed holder for the OpenBCI board, there was no specialized shielding in place to protect the electrodes or OpenBCI boards

from unwanted interference. They argue this makes the data more realistic, and whilst true, it is important to note there is still far less stochasticity than there would be in real life. The office room and hardware were identical for each of the participants in the data collection stage. The room was free of external stimulus as it was separated from the experiment organisers and the participant was *left alone* for the entire trial. All participants were right-handed and thus the dominant hand used for the MI task was always the right hand. Adding to all of this, the dominant hand of the participant was placed inside a box to not allow them to see their hand. It could be argued these factors still make the data more lab-like than home-like. Still, Peterson et al. (2020) found that the non-shielded regular office caused significant noise. In one trial the electromagnetic noise amplitude (50 Hz or 60 Hz depending on location) was four times higher than the meaningful EEG data. Other types of noise were also present, proving their environment was indeed far less ideal than a lab environment, which indeed makes it more realistic.

Third, during the data collection, a EMG system was also in place. This EMG system was used to filter out samples of the collected MI data where the movement was also physically performed, meaning it wasn't true MI data. Whilst this forms an interesting approach to data filtering when collecting training samples, the extra needed hardware that is only used during training is probably a tough sell in a commercial application. However, it does automate the data filtering process used, which is more realistic for widespread use than the expert-based data filtering in most literature.

Fourth, they used the KVIQ-10 questionnaire by Malouin et al. (2007) to determine how *good* a participant would be in MI, a task that is proven to be harder for some individuals. Interestingly, Peterson et al. (2020) found no statistically significant correlation between the KVIQ-10 score of participants and the found classification accuracy. Thus, whilst they did find significant inter-participant differences in accuracy, determining beforehand if a participant will have pleasant accuracy results beforehand through the KVIQ-10 questionnaire by Malouin et al. (2007) wasn't reliable for Peterson et al. (2020). This is an issue, as knowing this information beforehand can give a potential buyer a better indication if the system would be fit for them or not.

Finally, even though the data acquisition happened under guidance in the work of Peterson et al. (2020), there were still issues with the recordings. Out of the 12 participants, there were multiple moments where connection loss with the OpenBCI main board occurred, one participant where a mechanical defect rendered the data useless and one where there were EMG detected movements of the hand for more than half of the MI tasks rendering the trial of that patient useless as well. Whilst the latter is an issue independent of the hardware used, the other two are probably bigger issues on the cheaper hardware than the more reliable medical-grade hardware. Part of the reason medical-grade hardware is about 5 to 10 times as expensive for a similar experiment is due to the medical-grade hardware requiring certifications, which also guarantee some form of quality and reliability.

To conclude, the work by Peterson et al. (2020) discusses the creation of a complete low-cost consumer-grade BCI system. This system consists of the OpenBCI measuring equipment where the dry electrodes on the 3D printed Ultracortex Mark IV are replaced with electrodes in a more comfortable Electro-Cap. The effective classification of the system is a binary motor imagery (MI) classification on whether or not the participant imagines a grasping movement of the hand or not. Peterson et al. (2020) achieved an average accuracy between 70% and 85%, being higher as the approach used becomes more complex. It is important to note that the evaluated models are on a patient-per-patient basis. This means that each patient has their own uniquely trained model and that data from the same patient is used in the evaluation process. Whilst the binary nature of the system makes it hard to find viable real-life applications, the performance reached is almost identical to those of medical-grade systems and follows from a less lab-like environment

than is typically the case. The system proposed by Peterson et al. (2020) is of less importance in their work, rather the steps and pitfalls highlighted are of value.

### **Multiclass classification methods for motor imagery EEG data**

The above discussed paper by Peterson et al. (2020) provides great insight on the steps required to develop an EEG-based consumer-grade BCI which uses MI related signals. Since Peterson et al. (2020) uses a binary classification model, there are only two possible outputs of the classifier, which is too limited for most applications. However, the lack of training samples combined with noisy and often high-dimensional data of EEG makes multi-class classification considerably harder than binary using any BCI system. Adding to this, the consumer-grade hardware suffers even more from noise and MI data is notorious for being noise-prone in EEG data. This makes the choice for a binary classifier by Peterson et al. (2020) understandable. Peterson et al. (2020) also opted for an approach where a model is trained per user. This is often done in the field as it makes for far higher accuracy for that specific user. This does however mean that each new user should also undergo a training data collection procedure to have a custom-trained model as well before being able to use the system. This is less than ideal for commercial BCI systems and can have a physiological burden in medical applications.

To combat these points, many multi-class classification pipelines have been proposed in literature that work well with MI related EEG data (Abdeltawab & Ahmad, 2020; Z. Chen et al., 2021; Hou, Zhou, et al., 2020; Kai Keng Ang et al., 2008; Lawhern et al., 2016; Mane et al., 2021; Mussi et al., 2019; Olivas-Padilla & Chacon-Murguia, 2019; Schirrmeister et al., 2017). These pipelines generally work on both consumer-grade and medical-grade systems, although consumer-grade systems can often benefit more from specific noise-reduction steps in the pipeline. Some of the proposed pipelines also focus on generalisation to allow a general model which has usable performance for completely new users. Whilst such models have poorer performance overall, they can be used as an initial model to allow the user to explore the possibilities of the BCI model without having to undergo the often tedious training data collection process.

A complete in-depth review of all of the different approaches that can be taken to classify EEG data falls outside the scope of this research paper. Guerrero et al. (2021) compared logistic regression (LR), artificial neural network (ANN), support vector machines (SVM) and convolutional neural network (CNN) for a binary classification task of either being epileptic EEG data or not. Whilst this is again a binary classification that is more tailored towards computer-aided diagnosis (CADx), the techniques used in the experiments are also often used in the multi-class classification of EEG data for other BCI purposes. Guerrero et al. (2021) found that artificial neural networks performed best for their classification task. In general, ANNs and other DL models such as CNNs have proven to be successful at EEG data related tasks.

Because of this, many of the current state-of-the-art models for EEG classification rely on DL models. Especially classification pipelines that include CNNs have proven to be successful for EEG classification. From the previously referenced pipelines, Hou, Zhou, et al. (2020), Lawhern et al. (2016), Mane et al. (2021), Mussi et al. (2019), Olivas-Padilla and Chacon-Murguia (2019), and Schirrmeister et al. (2017) make use of CNNs. Although CNNs are only part of a good EEG classification pipelines. Kai Keng Ang et al. (2008) and Mane et al. (2021) used CNNs after a Filter-Bank approach, which has satisfactory accuracy for EEG classification and MI-related EEG data in specific. Some complex pipelines that rely on CNN can be too demanding for real-time classification, something that is needed for an online BCI system. Pipelines such as the one by Lawhern et al. (2016) have been developed to use CNNs in such a way that real-time classification is possible.

Other types of pipelines incorporate techniques based on spatial patterns, with common



spatial patterns being a very popular method used for EEG data. The pipelines by Abdeltawab and Ahmad (2020) and Olivas-Padilla and Chacon-Murguia (2019) use such spatial pattern techniques, as well as the previously discussed work by Peterson et al. (2020). a CSP approach is often combined with a CNN for classification but can also be combined with more classical ML techniques such as a SVM model or a linear discriminant analysis (LDA) (Abdeltawab & Ahmad, 2020).

As discussed, many approaches to EEG data classification exist and not all of them can be listed here. Some more noteworthy pipelines include the one by Z. Chen et al. (2021). Their approach is interesting in the fact that it visualizes EEG data so that an image processing method can be used for classification. Whilst this yields okay results, it doesn't improve upon the state-of-the-art. However, other novel approaches such as the one by Hou, Zhou, et al. (2020) which incorporates the technique of scout EEG source imaging (ESI) have shown to be as good as or even better than state-of-the-art in specific experiments.

### Connecting the classification model to physical devices

A complete BCI system could be thought of as a combination of three different components: a data collection process, the data processing step and the effective performing of actions by the system. The previously discussed motivating works focus mainly on how to collect EEG data, especially for MI tasks, and how to process this data to classify it. Whilst these two steps are already very challenging, a component to go from the classification labels to effective actions should also be in place to form a complete BCI system. For this last component, the labels provided by the classifier can be seen as an incoming input stream. Whilst it is intuitive to link certain labels with specific actions, for example, a left movement is linked to an imagined left-hand squeeze and a right movement is linked to an imagined right-hand squeeze, this isn't strictly required.

One of the challenges with BCI systems is the limited classification labels that can reliably be extracted. It is also the case that it is easier to distinguish between imagined left-hand movement and imagined right-hand movement than it is to differentiate between imagined left-hand thumb movement and imagined left-hand index finger movement. Because of this, less intuitive controls that are easier to classify and offer higher efficiency should often be considered. To illustrate this, imagine the movement of a robotic arm to pick up objects using MI related EEG data through a classifier that can distinguish left-hand squeeze, right-hand squeeze and an idle state. Intuitively, one might want to obtain complete control over the robotic arm but the limited inputs render a direct mapping between the classification label and all possible movements of the arm impossible. A menu for movement options could be made such as shown in Figure 1.8a. The imagined right-hand squeeze could be used to scroll through the menu options whilst the left-hand squeeze could be used to select that option. Stepping away from the idea of wanting to control every movement of the robotic arm can improve the efficiency of the system. With grasp detection algorithms such as the one proposed by Asif et al. (2018), the detection of objects of interest for the robot arm to interact with can be detected through computer-vision algorithms. Using such algorithms, another way of controlling the robotic arm could be using the imagined right-hand squeeze to switch between detected objects and using the imagined left-hand squeeze to pick them up, as depicted in Figure 1.8b. Highlighting the item detected by the arm could for example be done by the arm hovering over that specific item. Since the task of the robotic arm was picking up items, both systems would succeed, but the latter would be more efficient. Whilst this is a very naive example, it should illustrate that special thought should be put into this last component of the BCI system to maximize accuracy and efficiency.

As discussed, the field of BCI is highly interdisciplinary. The EEG data collection part relies




Figure 1.8: Contrast between a complete control design of a robotic grasping arm that is less efficient to operate and one that is more efficient to operate but has less control. Considering the task of grasping items, both systems would succeed but the more efficient one would likely be more pleasant to use.

heavily on neuroscience for the working of the brain to know what signals can be extracted and where they originate from. To extract those signals, engineers should develop hardware capable of measuring the tiny electrical current that is EEG. Whilst those engineers also focus on noise reduction, computer scientists are also needed for preprocessing to further reduce this noise. The data processing is mainly a computer scientist task, although knowledge from neuroscience can be very helpful in this step as well as insight on the flaws of the hardware. The final component, where effective actions are performed, can relate to a wide variety of sciences once again. For example, the robotic arm proposed before requires computer vision knowledge for grasp detection, engineering knowledge to make the arm and general computer science knowledge to create an intuitive link between classification labels and action controls. Because of this, it is often the case that research focuses mainly on improving one of these three components rather than the whole system. Take for example the BCI system proposed by Herath and de Mel (2021). It has a significantly sophisticated robotic hand that functions almost completely as a human hand does. The effective hardware used for the complete BCI system, including the processing unit, are also well detailed and shows thorough knowledge as it proposes a very affordable custom system. However, the classification algorithms used and the user interface proposed could benefit from future extensions to make the complete BCI system even better.

This is by no means a criticism to Herath and de Mel (2021) but demonstrates the interdisciplinary nature of BCI systems and how researchers that are specialized in one of these disciplines will outperform certain aspects of a BCI system while leaving room for improvement in other aspects. Likewise, many papers on the data processing and classification algorithms for EEG data are from computer scientists. Those papers often don't even include the final component where effective actions are taken, or the proposed system is limited to simulations because the development of robotics falls outside the scope of their discipline. This interdisciplinary nature of the BCI field is part of what makes it so fascinating yet also sophisticated. Commercial institutions that can hire many of the required professions will likely accelerate the creation of true complete BCI systems with state-of-the-art in each component of the system.

### 1.4.2 The potential of an AutoML variant for BCI pipelines

TODO

### 1.4.3 A lack of standardized testing

Evaluating and comparing different BCI systems is not easy. This is in part due to a BCI system consisting of different components. Thus, evaluating a BCI system purely on the tasks it can achieve and with which accuracy doesn't tell all that much. Evaluating the system for the above discussed three different components can improve on this, but a lack of standardized testing also makes this rather challenging.

Take for example the comparison and evaluation of the performance of the measuring equipment of a BCI system. Many BCI systems use the noisy EEG modality, but others make use of EMG and other modalities which can be more or less prone to noise. Some make use of easy to use dry electrodes, whilst others make use of wet electrodes that require considerable preparation. One could use the signal-to-noise ration (SNR) to compare and evaluate the EEG measuring device. However, this doesn't take into account affordability and user experience. A higher SNR but with more predictable noise is also preferred over a lower SNR with completely stochastic noise. Comparing the classification accuracy for specific tasks might seem like a better option then, but how do you make a fair classification pipeline for all different headsets. It should become visible that there is no easy solution to evaluate the (e.g. EEG) measuring equipment of a BCI system.

Similar issues arise for the data processing component. For starters, whilst relatively complete datasets such as the one by Kaya et al. (2018) exist, there is no real *reference dataset* for BCI systems. In computer vision, for example, popular datasets such as MNIST (Li Deng, 2012) and ImageNet (Deng et al., 2009) are often used to train and evaluate image classification algorithms. But EEG classification algorithms are optimised to the input data. For example, some work better with noisy data whilst others are optimized for MI specific classification tasks and so on. Do you train the models on a single patient and test them on the same patient, or do you test for generalisability? Do you allow it to run on very capable hardware or limited but very affordable hardware? Again, it should be apparent that there is no straightforward way of evaluating the data processing step.

This thesis only aims to highlight the issue that arises when trying to evaluate and compare different BCI systems. These problems such as a standardized testing suite are open problems in the field and one that could greatly improve the field's work once a solution is proposed and accepted by the community. For now, focusing on reproducibility and specifying the data used and potential ways it makes tasks easier or harder is the best most researchers can do. It becomes apparent that a 90% accuracy for a model that can be used on any user without re-training is far more impressive than one custom made for a specific user. Likewise, 80% accuracy on non-lab environment data with high stochasticity is far more impressive than the same accuracy on the best measuring hardware in the most controlled environment.

### 1.4.4 Challenges from the highly interdisciplinary nature of BCI systems

TODO

### 1.4.5 Difficulties with repeatability and reproducibility of experiments

TODO

#### 1.4.6 Complex data variability and user-training

### 1.5 Ethical challenges for BCIs

SECTION TO BE COMPLETED IN A LATER STAGE, LOREM IPSUM PLACED FOR APPROXIMATED LENGTH.

TODO

#### 1.5.1 The return of the Luddites

TODO

#### 1.5.2 Advertisements based on your thoughts

TODO

#### 1.5.3 Hacking BCI systems

TODO

#### 1.5.4 Changing peoples personal identities

TODO

#### 1.5.5 Painfully confronting users with their brain

TODO

#### 1.5.6 E-waste inside your skull

TODO

### 1.6 Chapter conclusions and proposing a three-signal system for basic controls

**NOTE: this will be edited once the thesis is "finished"**

As discussed, this master thesis focuses on providing a great foundation for the knowledge required for working in the BCI field as a computer scientist. The exhaustive literature review from this chapter should provide a great general introduction to the field and current state-of-the-art as well as challenges and promises of the field. As touched upon in Section 1.4.1 and further discussed in Chapter 3 many different pipelines and approaches exists for the data processing component of a BCI system. Whilst some of the libraries available which make use of deep learning (DL) allow for raw EEG input, the author of this paper believes it to be of importance to know the nature of this EEG data to some extend. For this, an introduction to biosignals and how they can be measured is given in the next chapter. This deeper understanding of the data source ultimately leads to better design decisions in the data processing component.

In part 2 of this thesis, a focus is put on how BCI systems can be development. To accomplish this, an offline BCI that classifies EEG data is discussed in Chapter 4. Chapter 5 extends on this offline BCI pipeline by proposing a three-signal system for live control. Part 3 of this thesis

consists of Chapter 6 and 7 which aim to evaluate this three-signal system, taking into account the lack of generalized evaluation strategies as discussed earlier in Section 1.4.3



## Chapter 2

# Origin and acquisition of biomedical signals

### 2.1 Introduction to this chapter

TODO

### 2.2 Origins of biosignals

TODO

#### 2.2.1 Electrical biosignals

TODO

#### 2.2.2 Non-electrical biosignals

TODO

### 2.3 Biosignals from the brain

TODO

#### 2.3.1 Anatomy of the brain

TODO

#### 2.3.2 Brain waves

TODO

### 2.3.3 Measurable brain-activity

Motor imagery (MI) is the process in which a person generates brain-activity in the motor cortex merely by imagining motor movements. MI-based brain-computer interfaces (BCIs) are interesting because they don't require any external stimulus nor effective motor movements

TODO

### 2.3.4 Generalisation issues of brain activity

TODO

## 2.4 Measuring brain-signals

Many comparisons between different types of measuring equipment, often with greatly differing costs, have already been made (David Hairston et al., 2014; McCrimmon et al., 2017; Nijboer et al., 2015; Pathirana et al., 2018; Ratti et al., 2017). The main consensus is that the cheaper consumer-grade equipment has the potential to reach similar performance of a conventional, often medical-grade, BCI system. These results are promising but due to the controlled nature of the experiments, they might not reflect real-life applications accurately. As discussed before, the user experience of a BCI system is as important if not more important than the raw performance of the system.

### 2.4.1 Measuring modalities

TODO

Research by Berger (1929) is the first in describing the measurement of brain waves from the human skull in a non-invasive manner. Because of this, the German neuroscientist and psychiatrist Hans Berger is often seen as the inventor of electroencephalography (EEG). Whilst he was one of the first to use the term *elektrenkephalogramm*, it was Richard Caton who first described the findings of brain waves in general. He found this phenomena in animal brains as early as 1875 (Haas, 2003). Since then, EEG methodology and equipment has matured and evolved a lot.

### 2.4.2 Motivation for using non-invasive EEG

TODO

### 2.4.3 Standards for EEG measuring systems

TODO

### 2.4.4 Comparison of available EEG measuring equipment

TODO

### 2.4.5 Common EEG artefacts

TODO




## 2.5 Chapter conclusions

TODO



(a) Experimental analog EEG recording equipment used by Berger (1929).  
Figure by by Holmes (2014).



(b) Medical-grade analog EEG recording equipment estimated to be from the 1950's.  
Figure by Devotor<sup>1</sup>.



(c) Early portable analog EEG recording equipment from the late 1950's.  
Figure by Sam Brusco<sup>2</sup>.



(d) William Grey Walter and medical-grade analog EEG recording equipment, 1964.  
Figure by Burden Neurological Institute<sup>3</sup>.

Figure 2.1: Early analog EEG equipment.

<sup>1</sup><https://www.charismaticplanet.com/the-electroencephalogram-1924/>

<sup>2</sup><https://www.medicaldesignandoutsourcing.com/medtech-memoirs-the-electroencephalograph-eeg/>

<sup>3</sup><http://dx.doi.org/10.15180/181003/019>

## Chapter 3

# Processing brain-signals and taking actions from their interpretation

### 3.1 Introduction to this chapter

TODO

### 3.2 A general EEG-based BCI pipeline

#### 3.2.1 Data acquisition

TODO

#### 3.2.2 Preprocessing

TODO

#### 3.2.3 Windowing

TODO

#### 3.2.4 Feature engineering

TODO

#### 3.2.5 Classification model

TODO

#### 3.2.6 Performing an action

TODO

### **3.3 Alternative BCI pipelines**

#### **3.3.1 Using existing open-source data for training and testing**

TODO

#### **3.3.2 Neglecting preprocessing and feature engineering**

TODO

#### **3.3.3 No action performing step**

TODO

#### **3.3.4 Offline vs online BCI systems**

TODO

### **3.4 The role of machine learning and deep learning**

#### **3.4.1 Difference between traditional machine learning and deep learning**

TODO

#### **3.4.2 Supervised, Semi-Supervised, Unsupervised, and Self-Supervised Learning**

TODO

#### **3.4.3 Common regular machine learning classifiers**

TODO

#### **3.4.4 Common deep learning classifiers**

TODO

### **3.5 Common issues when processing brain-signals and how to avoid them**

TODO

#### **3.5.1 Biased data**

TODO

### **3.5.2 Incorrect or ambiguous evaluation**

TODO

### **3.5.3 No explainability or interpretability**

TODO

Deep learning (DL) often requires significant processing power and time to train, impacting the affordability of brain-computer interface (BCI) research. This is especially true when working with many electroencephalography (EEG) sensors and features, and thus a high dimensional setting. DL is often also used in a black-box principle. This means that the trained system lacks explainability and interpretability. Recent governmental reports have suggested that laws will be coming in place to require these properties (European Commission. Joint Research Centre., 2020; The Royal Society, 2019).

### **3.5.4 Overfitting**

TODO

## **3.6 Chapter conclusions**

TODO



## Part II

**Development an EEG-based  
brain-computer interface that  
classifies motor imagery tasks**





## Chapter 4

# EEG-based offline classification system for motor imagery tasks

TODO



## Chapter 5

# Moving from an offline classification system towards an online BCI system

TODO



## Part III

# Reflection on the results of this thesis



## Chapter 6

# Using the system and verifying the results

TODO





## Chapter 7

# Self-reflection and conclusions

TODO







# List of abbreviations and acronyms

## Symbols

$\mu V$  microvolts.

## A

**AI** artificial intelligence.

**ALS** amyotrophic lateral sclerosis.

**ANN** artificial neural network.

**AR** augmented reality.

## B

**BBCI** Berlin Brain-Computer Interface research program.

**BCI** brain-computer interface.

**biosignal** biomedical signal.

**biosignal control** biological signal control.

**BMI** brain-machine interface.

**BPNN** back propagation neural network.

## C

**CADe** computer-aided detection.

**CADx** computer-aided diagnosis.

**CNN** convolutional neural network.

**CPU** central processing unit.

**CSF** cerebrospinal fluid.

**CSP** common spatial pattern.

**D**

**DEEG** digital electroencephalography.

**DL** deep learning.

**DMD** duchenne muscular dystrophy.

**DSP** digital signal processor.

**E**

**ECG** electrocardiography.

**ECoG** electrocorticography.

**EEG** electroencephalography.

**EMG** electromyography.

**ERP** event-related potential.

**ESI** scout EEG source imaging.

**F**

**FBCSP** filter bank common spatial pattern.

**FDA** Food and Drug Administration.

**FN** false negative.

**FP** false positive.

**FPGA** field programmable gate arrays.

**G**

**GAN** generative adversarial network.

**GDPR** general data protection regulation.

**GPU** graphics processing unit.

**H**

**HMI** human-machine interface.

**HU** Hounsfield unit.

**L**

**latex** Is a markup language specially suited for scientific documents.

**LDA** linear discriminant analysis.

**LR** logistic regression.

**M**

**MEG** magnetoencephalography.

**MI** motor imagery.

**ML** machine learning.

**N**

**NLP** natural language processing.

**NN** neural network.

**P**

**PFBCSP** penalized frequency band common spatial pattern.

**POC** proof of concept.

**PTFBCSP** penalized time-frequency band common spatial pattern.

**R**

**RF** random forest.

**RL** reinforcement learning.

**RNN** recurrent neural network.

**S**

**SNR** signal-to-noise ration.

**SSVEP** steady-state visual evoked potential.

**SVM** support vector machines.

**T**

**TL** Transfer learning.

**topomap** topographic map.

**U**

**UI** user interface.

**UX** user experience.

**V**

**VR** virtual reality.





# References

- Abdeltawab, A., & Ahmad, A. (2020). Classification of Motor Imagery EEG Signals Using Machine Learning. *2020 IEEE 10th International Conference on System Engineering and Technology (ICSET)*, 196–201. <https://doi.org/10.1109/ICSET51301.2020.9265364>
- Abdulkader, S. N., Atia, A., & Mostafa, M.-S. M. (2015). Brain computer interfacing: Applications and challenges. *Egyptian Informatics Journal*, 16(2), 213–230. <https://doi.org/10.1016/j.eij.2015.06.002>
- Adama, S. V., & Bogdan, M. (2021). Stroke Rehabilitation and Parkinson’s Disease Tremor Reduction Using BCIs Combined With FES: In I. R. Management Association (Ed.), *Research Anthology on Rehabilitation Practices and Therapy* (pp. 679–697). IGI Global. <https://doi.org/10.4018/978-1-7998-3432-8.ch033>
- Aflalo, T., Kellis, S., Klaes, C., Lee, B., Shi, Y., Pejisa, K., Shanfield, K., Hayes-Jackson, S., Aisen, M., Heck, C., Liu, C., & Andersen, R. A. (2015). Decoding motor imagery from the posterior parietal cortex of a tetraplegic human. *Science*, 348(6237), 906–910. <https://doi.org/10.1126/science.aaa5417>
- Alcaide, R., Agarwal, N., Candassamy, J., Cavanagh, S., Lim, M., Meschede-Krasa, B., McIntyre, J., Ruiz-Blondet, M. V., Siebert, B., Stanley, D., Valeriani, D., & Yousefi, A. (2021). *EEG-Based Focus Estimation Using Neurable’s Enten Headphones and Analytics Platform* (preprint). Neuroscience. <https://doi.org/10.1101/2021.06.21.448991>
- AL-Quraishi, M., Elamvazuthi, I., Daud, S., Parasuraman, S., & Borboni, A. (2018). EEG-Based Control for Upper and Lower Limb Exoskeletons and Prostheses: A Systematic Review. *Sensors*, 18(10), 3342. <https://doi.org/10.3390/s18103342>
- Asif, U., Tang, J., & Harrer, S. (2018). GraspNet: An Efficient Convolutional Neural Network for Real-time Grasp Detection for Low-powered Devices. *Proceedings of the Twenty-Seventh International Joint Conference on Artificial Intelligence*, 4875–4882. <https://doi.org/10.24963/ijcai.2018/677>
- Badrakalimuthu, V. R., Swamiraju, R., & de Waal, H. (2011). EEG in psychiatric practice: To do or not to do? *Advances in Psychiatric Treatment*, 17(2), 114–121. <https://doi.org/10.1192/apt.bp.109.006916>
- Baldi, P. (2018). Deep Learning in Biomedical Data Science. *Annual Review of Biomedical Data Science*, 1(1), 181–205. <https://doi.org/10.1146/annurev-biodatasci-080917-013343>
- Berger, H. (1929). Über das Elektrenkephalogramm des Menschen. *Archiv für Psychiatrie und Nervenkrankheiten*, 87(1), 527–570. <https://doi.org/10.1007/BF01797193>
- Bernal, G. (2021). Developing Galea: An open source tool at the intersection of VR and neuroscience. *MIT Media Lab*. <https://medium.com/mit-media-lab/developing-galea-an-open-source-tool-at-the-intersection-of-vr-and-neuroscience-61ef60359b96>
- Binnendijk, A., Marler, T., & Bartels, E. (2020). *Brain-Computer Interfaces: U.S. Military Applications and Implications, An Initial Assessment*. RAND Corporation. <https://doi.org/10.7249/RR2996>

- Blankertz, B., Muller, K.-R., Curio, G., Vaughan, T., Schalk, G., Wolpaw, J., Schlogl, A., Neuper, C., Pfurtscheller, G., Hinterberger, T., Schroder, M., & Birbaumer, N. (2004). The BCI competition 2003: Progress and perspectives in detection and discrimination of EEG single trials. *IEEE Transactions on Biomedical Engineering*, 51(6), 1044–1051. <https://doi.org/10.1109/TBME.2004.826692>
- Blankertz, B., Muller, K.-R., Krusienski, D., Schalk, G., Wolpaw, J., Schlogl, A., Pfurtscheller, G., Millan, J., Schroder, M., & Birbaumer, N. (2006). The BCI competition III: Validating alternative approaches to actual BCI problems. *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, 14(2), 153–159. <https://doi.org/10.1109/TNSRE.2006.875642>
- Blaus, B. (2014). Medical gallery of Blausen Medical 2014. *WikiJournal of Medicine*, 1(2). <https://doi.org/10.15347/wjm/2014.010>
- Bontinck, L. (2021). *Bci master thesis @ vub 2021 - 2022* [GitHub commit: 7e32955...]. Retrieved September 28, 2021, from <https://github.com/pikawika/bci-master-thesis>
- Boser, B. E., Guyon, I. M., & Vapnik, V. N. (1992). A training algorithm for optimal margin classifiers. *Proceedings of the fifth annual workshop on Computational learning theory - COLT '92*, 144–152. <https://doi.org/10.1145/130385.130401>
- Brown, T. B., Mann, B., Ryder, N., Subbiah, M., Kaplan, J., Dhariwal, P., Neelakantan, A., Shyam, P., Sastry, G., Askell, A., Agarwal, S., Herbert-Voss, A., Krueger, G., Henighan, T., Child, R., Ramesh, A., Ziegler, D. M., Wu, J., Winter, C., ... Amodei, D. (2020). Language Models are Few-Shot Learners [Publisher: arXiv Version Number: 4]. <https://doi.org/10.48550/ARXIV.2005.14165>
- Burle, B., Spieser, L., Roger, C., Casini, L., Hasbroucq, T., & Vidal, F. (2015). Spatial and temporal resolutions of EEG: Is it really black and white? A scalp current density view. *International Journal of Psychophysiology*, 97(3), 210–220. <https://doi.org/10.1016/j.ijpsycho.2015.05.004>
- Capati., F. A., Bechelli., R. P., & Castro., M. C. F. Hybrid ssvep/p300 bci keyboard - controlled by visual evoked potential. In: *Proceedings of the 9th international joint conference on biomedical engineering systems and technologies - biosignals, (biostec 2016)*. INSTICC. SciTePress, 2016, 214–218. ISBN: 978-989-758-170-0. <https://doi.org/10.5220/0005705202140218>
- Cardoso, V. F., Delisle-Rodriguez, D., Romero-Laiseca, M. A., Loterio, F. A., Gurve, D., Floriano, A., Krishnan, S., Frizera-Neto, A., & Filho, T. F. B. (2022). BCI based on pedal end-effector triggered through pedaling imagery to promote excitability over the feet motor area. *Research on Biomedical Engineering*. <https://doi.org/10.1007/s42600-021-00196-7>
- Cardoso, V. F., Delisle-Rodriguez, D., Romero-Laiseca, M. A., Loterio, F. A., Gurve, D., Floriano, A., Valadão, C., Silva, L., Krishnan, S., Frizera-Neto, A., & Freire Bastos-Filho, T. (2021). Effect of a Brain–Computer Interface Based on Pedaling Motor Imagery on Cortical Excitability and Connectivity. *Sensors*, 21(6), 2020. <https://doi.org/10.3390/s21062020>
- Center for Security and Emerging Technology, Konaev, M., Chahal, H., Fedasiuk, R., Huang, T., & Rahkovsky, I. (2020). *U.S. Military Investments in Autonomy and AI: A Budgetary Assessment* (tech. rep.). Center for Security and Emerging Technology. <https://doi.org/10.51593/20200069>
- Chen, S. (2022). Mind-reading device to detect porn could speed China’s policing of illicit content, say researchers. Retrieved July 13, 2022, from <https://www.scmp.com/news/china/science/article/3182087/mind-reading-device-detect-porn-could-speed-chinas-policing>

- Chen, Z., Wang, Y., & Song, Z. (2021). Classification of Motor Imagery Electroencephalography Signals Based on Image Processing Method. *Sensors*, 21(14), 4646. <https://doi.org/10.3390/s21144646>
- Chervonenkis, A. Y. (2013). Early History of Support Vector Machines. In B. Schölkopf, Z. Luo, & V. Vovk (Eds.), *Empirical Inference* (pp. 13–20). Springer Berlin Heidelberg. [https://doi.org/10.1007/978-3-642-41136-6\\_3](https://doi.org/10.1007/978-3-642-41136-6_3)
- Chin-Teng Lin, Shu-Fang Tsai, & Li-Wei Ko. (2013). EEG-Based Learning System for Online Motion Sickness Level Estimation in a Dynamic Vehicle Environment. *IEEE Transactions on Neural Networks and Learning Systems*, 24(10), 1689–1700. <https://doi.org/10.1109/TNNLS.2013.2275003>
- Chollet, F. et al. (2015). Keras.
- Cong Wang, Bin Xia, Jie Li, Wenlu Yang, Dianyun, Alejandro Cardona Velez, & Yang, H. (2011). Motor imagery BCI-based robot arm system. *2011 Seventh International Conference on Natural Computation*, 181–184. <https://doi.org/10.1109/ICNC.2011.6021923>
- Cruz-Garza, J. G., Brantley, J. A., Nakagome, S., Kontson, K., Megjhani, M., Robleto, D., & Contreras-Vidal, J. L. (2017). Deployment of Mobile EEG Technology in an Art Museum Setting: Evaluation of Signal Quality and Usability. *Frontiers in Human Neuroscience*, 11, 527. <https://doi.org/10.3389/fnhum.2017.00527>
- Cuthbertson, A. (2021). Valve is working on brain-computer interface gaming, president reveals. *Independent*. <https://www.independent.co.uk/life-style/gadgets-and-tech/valve-brain-computer-interface-video-game-b1792225.html>
- Dadia, T., & Greenbaum, D. (2019). Neuralink: The Ethical ‘Rithmetic of Reading and Writing to the Brain. *AJOB Neuroscience*, 10(4), 187–189. <https://doi.org/10.1080/21507740.2019.1665129>
- Daly, J. J., & Wolpaw, J. R. (2008). Brain–computer interfaces in neurological rehabilitation. *The Lancet Neurology*, 7(11), 1032–1043. [https://doi.org/10.1016/S1474-4422\(08\)70223-0](https://doi.org/10.1016/S1474-4422(08)70223-0)
- da Silva Souto, C., Lüddemann, H., Lipski, S., Dietz, M., & Kollmeier, B. (2016). Influence of attention on speech-rhythm evoked potentials: First steps towards an auditory brain-computer interface driven by speech. *Biomedical Physics & Engineering Express*, 2(6), 065009. <https://doi.org/10.1088/2057-1976/2/6/065009>
- David Hairston, W., Whitaker, K. W., Ries, A. J., Vettel, J. M., Courtney Bradford, J., Kerick, S. E., & McDowell, K. (2014). Usability of four commercially-oriented EEG systems. *Journal of Neural Engineering*, 11(4), 046018. <https://doi.org/10.1088/1741-2560/11/4/046018>
- De Smet, R. (2021). *Vub latex huisstijl* [GitHub commit: 2de903e5...]. Retrieved September 28, 2021, from <https://gitlab.com/rubdos/texlive-vub>
- De Wulf, W. (2022). *Transfer learning in brain-computer interfaces: Language-pretrained transformers for classifying electroencephalography* (Doctoral dissertation). VUB. Brussels. Retrieved July 16, 2022, from <https://github.com/wulfdewolf/lpt-for-eeg>
- Deiss, O., Biswal, S., Jin, J., Sun, H., Westover, M. B., & Sun, J. (2018). HAMLET: Interpretable Human And Machine co-LEarning Technique [arXiv: 1803.09702]. *arXiv:1803.09702 [cs, stat]*. Retrieved December 6, 2021, from <http://arxiv.org/abs/1803.09702>
- Delorme, A., & Makeig, S. (2004). EEGLAB: An open source toolbox for analysis of single-trial EEG dynamics including independent component analysis. *Journal of Neuroscience Methods*, 134(1), 9–21. <https://doi.org/10.1016/j.jneumeth.2003.10.009>
- Deng, J., Dong, W., Socher, R., Li, L.-J., Kai Li, & Li Fei-Fei. (2009). ImageNet: A large-scale hierarchical image database. *2009 IEEE Conference on Computer Vision and Pattern Recognition*, 248–255. <https://doi.org/10.1109/CVPR.2009.5206848>

- Dillen, A., Steckelmacher, D., Efthymiadis, K., Langlois, K., De Beir, A., Marusic, U., Vanderborght, B., Nowé, A., Meeusen, R., Ghaffari, F., Romain, O., & De Pauw, K. (2022). Deep learning for biosignal control: Insights from basic to real-time methods with recommendations. *Journal of Neural Engineering*, 19(1), 011003. <https://doi.org/10.1088/1741-2552/ac4f9a>
- European Commission. Joint Research Centre. (2020). *Robustness and explainability of Artificial Intelligence: From technical to policy solutions*. Publications Office. Retrieved March 1, 2022, from <https://data.europa.eu/doi/10.2760/57493>
- Facebook. (2017). F8 2017: AI, Building 8 and More Technology Updates From Day Two. Retrieved April 11, 2022, from <https://about.fb.com/news/2017/04/f8-2017-day-2/>
- Facebook. (2021). BCI milestone: New research from UCSF with support from Facebook shows the potential of brain-computer interfaces for restoring speech communication. <https://tech.fb.com/bci-milestone-new-research-from-ucsf-with-support-from-facebook-shows-the-potential-of-brain-computer-interfaces-for-restoring-speech-communication/>
- Fan, X., Bi, L., & Wang, Z. (2012). Detecting emergency situations by monitoring drivers' states from EEG. *2012 ICME International Conference on Complex Medical Engineering (CME)*, 245–248. <https://doi.org/10.1109/ICCME.2012.6275717>
- Ferree, T., Clay, M., & Tucker, D. (2001). The spatial resolution of scalp EEG. *Neurocomputing*, 38–40, 1209–1216. [https://doi.org/10.1016/S0925-2312\(01\)00568-9](https://doi.org/10.1016/S0925-2312(01)00568-9)
- Fleischer, C. (2007). Controlling Exoskeletons with EMG signals and a Biomechanical Body Model [Publisher: Technische Universität Berlin]. <https://doi.org/10.14279/DEPOSITONCE-1674>
- Floridi, L. (2020). AI and Its New Winter: From Myths to Realities. *Philosophy & Technology*, 33(1), 1–3. <https://doi.org/10.1007/s13347-020-00396-6>
- Frey, J. (2016). Comparison of an open-hardware electroencephalography amplifier with medical grade device in brain-computer interface applications [Publisher: arXiv Version Number: 1]. <https://doi.org/10.48550/ARXIV.1606.02438>
- Fukushima, K. (1980). Neocognitron: A self-organizing neural network model for a mechanism of pattern recognition unaffected by shift in position. *Biological Cybernetics*, 36(4), 193–202. <https://doi.org/10.1007/BF00344251>
- Fuller, M. (2019). Big data and the Facebook scandal: Issues and responses. *Theology*, 122(1), 14–21. <https://doi.org/10.1177/0040571X18805908>
- Gramfort, A. (2013). MEG and EEG data analysis with MNE-Python. *Frontiers in Neuroscience*, 7. <https://doi.org/10.3389/fnins.2013.00267>
- Guerrero, M. C., Parada, J. S., & Espitia, H. E. (2021). EEG signal analysis using classification techniques: Logistic regression, artificial neural networks, support vector machines, and convolutional neural networks. *Heliyon*, 7(6), e07258. <https://doi.org/10.1016/j.heliyon.2021.e07258>
- Guy, V., Soriani, M.-H., Bruno, M., Papadopoulou, T., Desnuelle, C., & Clerc, M. (2018). Brain computer interface with the P300 speller: Usability for disabled people with amyotrophic lateral sclerosis. *Annals of Physical and Rehabilitation Medicine*, 61(1), 5–11. <https://doi.org/10.1016/j.rehab.2017.09.004>
- Haas, L. F. (2003). Hans Berger (1873-1941), Richard Caton (1842-1926), and electroencephalography. *Journal of Neurology, Neurosurgery & Psychiatry*, 74(1), 9–9. <https://doi.org/10.1136/jnnp.74.1.9>
- Herath, H. M. K. M. B., & de Mel, W. (2021). Controlling an Anatomical Robot Hand Using the Brain-Computer Interface Based on Motor Imagery (A. Piccinno, Ed.). *Advances in Human-Computer Interaction*, 2021, 1–15. <https://doi.org/10.1155/2021/5515759>

- Hinrichs, H., Scholz, M., Baum, A. K., Kam, J. W. Y., Knight, R. T., & Heinze, H.-J. (2020). Comparison between a wireless dry electrode EEG system with a conventional wired wet electrode EEG system for clinical applications. *Scientific Reports*, 10(1), 5218. <https://doi.org/10/gm3k2p>
- Hodson, R. (2019). The brain. *Nature*, 571(7766), S1–S1. <https://doi.org/10.1038/d41586-019-02206-2>
- Holmes, K. (2014). Hans Berger and the E.E.G. *Thresholds*, 42, 88–97. [https://doi.org/10.1162/thld\\_a\\_00080](https://doi.org/10.1162/thld_a_00080)
- Hou, Y., Jia, S., Lun, X., Zhang, S., Chen, T., Wang, F., & Lv, J. (2020). GCNs-Net: A Graph Convolutional Neural Network Approach for Decoding Time-resolved EEG Motor Imagery Signals [Publisher: arXiv Version Number: 3]. <https://doi.org/10.48550/ARXIV.2006.08924>
- Hou, Y., Jia, S., Lun, X., Zhang, S., Chen, T., Wang, F., & Lv, J. (2022). Deep Feature Mining via the Attention-Based Bidirectional Long Short Term Memory Graph Convolutional Neural Network for Human Motor Imagery Recognition. *Frontiers in Bioengineering and Biotechnology*, 9, 706229. <https://doi.org/10.3389/fbioe.2021.706229>
- Hou, Y., Zhou, L., Jia, S., & Lun, X. (2020). A novel approach of decoding EEG four-class motor imagery tasks via scout ESI and CNN. *Journal of Neural Engineering*, 17(1), 016048. <https://doi.org/10.1088/1741-2552/ab4af6>
- Hu, M. (2020). Cambridge Analytica’s black box. *Big Data & Society*, 7(2), 205395172093809. <https://doi.org/10.1177/2053951720938091>
- Humayun, M. S., Dorn, J. D., da Cruz, L., Dagnelie, G., Sahel, J.-A., Stanga, P. E., Cideciyan, A. V., Duncan, J. L., Elliott, D., Filley, E., Ho, A. C., Santos, A., Safran, A. B., Arditi, A., Del Priore, L. V., & Greenberg, R. J. (2012). Interim Results from the International Trial of Second Sight’s Visual Prosthesis. *Ophthalmology*, 119(4), 779–788. <https://doi.org/10.1016/j.ophttha.2011.09.028>
- Hunter Christie, E. (2022). Defence cooperation in artificial intelligence: Bridging the transatlantic gap for a stronger Europe. *European View*, 21(1), 13–21. <https://doi.org/10.1177/17816858221089372>
- Hussein, D., Ibrahim, D., Alrumiah, S., Alhajjaj, L., & Alshobaili, J. (2020). A review on brain-computer interface (bci) spellers: P300 speller. *Bioscience Biotechnology Research Communications*, 13, 1191–1199. <https://doi.org/10.21786/bbrc/13.3/31>
- Ifft, P. J., Shokur, S., Li, Z., Lebedev, M. A., & Nicolelis, M. A. L. (2013). A Brain-Machine Interface Enables Bimanual Arm Movements in Monkeys. *Science Translational Medicine*, 5(210). <https://doi.org/10.1126/scitranslmed.3006159>
- Jawad, A. J. (2020). Engineering Ethics of Neuralink Brain Computer Interfaces Devices. *Annals of Bioethics & Clinical Applications*, 4(1). <https://doi.org/10.23880/abca-16000160>
- Jia, S., Hou, Y., Shi, Y., & Li, Y. (2020). Attention-based Graph ResNet for Motor Intent Detection from Raw EEG signals [Publisher: arXiv Version Number: 1]. <https://doi.org/10.48550/ARXIV.2007.13484>
- Kai Keng Ang, Zhang Yang Chin, Haihong Zhang, & Cuntai Guan. (2008). Filter Bank Common Spatial Pattern (FBCSP) in Brain-Computer Interface. *2008 IEEE International Joint Conference on Neural Networks (IEEE World Congress on Computational Intelligence)*, 2390–2397. <https://doi.org/10.1109/IJCNN.2008.4634130>
- Kalika, D., Collins, L., Caves, K., & Throckmorton, C. (2017). Fusion of P300 and eye-tracker data for spelling using BCI2000. *Journal of Neural Engineering*, 14(5), 056010. <https://doi.org/10.1088/1741-2552/aa776b>

- Kapgate, D. (2022). Efficient Quadcopter Flight Control Using Hybrid SSVEP + P300 Visual Brain Computer Interface. *International Journal of Human-Computer Interaction*, 38(1), 42–52. <https://doi.org/10.1080/10447318.2021.1921482>
- Kawala-Sterniuk, A., Browarska, N., Al-Bakri, A., Pelc, M., Zygarlicki, J., Sidikova, M., Martinek, R., & Gorzelanczyk, E. J. (2021). Summary of over Fifty Years with Brain-Computer Interfaces—A Review. *Brain Sciences*, 11(1), 43. <https://doi.org/10.3390/brainsci11010043>
- Kaya, M., Binli, M. K., Ozbay, E., Yanar, H., & Mishchenko, Y. (2018). A large electroencephalographic motor imagery dataset for electroencephalographic brain computer interfaces. *Scientific Data*, 5(1), 180211. <https://doi.org/10.1038/sdata.2018.211>
- Kindermans, P.-J., Verschore, H., Verstraeten, D., & Schrauwen, B. (2012). A P300 BCI for the masses: Prior information enables instant unsupervised spelling. <http://lib.ugent.be/catalog/pug01:3008919>
- Koles, Z. J., Lazar, M. S., & Zhou, S. Z. (1990). Spatial patterns underlying population differences in the background EEG. *Brain Topography*, 2(4), 275–284. <https://doi.org/10.1007/BF01129656>
- Kübler, A. (2020). The history of BCI: From a vision for the future to real support for personhood in people with locked-in syndrome. *Neuroethics*, 13(2), 163–180. <https://doi.org/10.1007/s12152-019-09409-4>
- Kwon, M., Han, S., Kim, K., & Jun, S. C. (2019). Super-Resolution for Improving EEG Spatial Resolution using Deep Convolutional Neural Network—Feasibility Study. *Sensors*, 19(23), 5317. <https://doi.org/10.3390/s19235317>
- Lawhern, V. J., Solon, A. J., Waytowich, N. R., Gordon, S. M., Hung, C. P., & Lance, B. J. (2016). EEGNet: A Compact Convolutional Network for EEG-based Brain-Computer Interfaces [Publisher: arXiv Version Number: 4]. <https://doi.org/10.48550/ARXIV.1611.08024>
- Lebedev, M. A. (2014). How to read neuron-dropping curves? *Frontiers in Systems Neuroscience*, 8. <https://doi.org/10.3389/fnsys.2014.00102>
- Lee, D. (2020). Comparison of Reinforcement Learning Activation Functions to Improve the Performance of the Racing Game Learning Agent. *Journal of Information Processing Systems*, 16(5), 1074–1082. <https://doi.org/10.3745/JIPS.02.0141>
- Li Deng. (2012). The MNIST Database of Handwritten Digit Images for Machine Learning Research [Best of the Web]. *IEEE Signal Processing Magazine*, 29(6), 141–142. <https://doi.org/10.1109/MSP.2012.2211477>
- Lin, C.-T., Chen, Y.-C., Huang, T.-Y., Chiu, T.-T., Ko, L.-W., Liang, S.-F., Hsieh, H.-Y., Hsu, S.-H., & Duann, J.-R. (2008). Development of Wireless Brain Computer Interface With Embedded Multitask Scheduling and its Application on Real-Time Driver's Drowsiness Detection and Warning. *IEEE Transactions on Biomedical Engineering*, 55(5), 1582–1591. <https://doi.org/10.1109/TBME.2008.918566>
- Liu, X., Makeyev, O., & Besio, W. (2020). Improved Spatial Resolution of Electroencephalogram Using Tripolar Concentric Ring Electrode Sensors. *Journal of Sensors*, 2020, 1–9. <https://doi.org/10.1155/2020/6269394>
- Malin, B. A., Emam, K. E., & O'Keefe, C. M. (2013). Biomedical data privacy: Problems, perspectives, and recent advances. *Journal of the American Medical Informatics Association*, 20(1), 2–6. <https://doi.org/10.1136/amiajnl-2012-001509>
- Malouin, F., Richards, C. L., Jackson, P. L., Lafleur, M. F., Durand, A., & Doyon, J. (2007). The Kinesthetic and Visual Imagery Questionnaire (KVIQ) for Assessing Motor Imagery in Persons with Physical Disabilities: A Reliability and Construct Validity Study. *Journal of Neurologic Physical Therapy*, 31(1), 20–29. <https://doi.org/10.1097/01.NPT.0000260567.24122.64>

- Mane, R., Chew, E., Chua, K., Ang, K. K., Robinson, N., Vinod, A. P., Lee, S.-W., & Guan, C. (2021). FBCNet: A Multi-view Convolutional Neural Network for Brain-Computer Interface [arXiv: 2104.01233]. *arXiv:2104.01233 [cs, eess]*. Retrieved September 28, 2021, from <http://arxiv.org/abs/2104.01233>
- Martín Abadi, Ashish Agarwal, Paul Barham, Eugene Brevdo, Zhifeng Chen, Craig Citro, Greg S. Corrado, Andy Davis, Jeffrey Dean, Matthieu Devin, Sanjay Ghemawat, Ian Goodfellow, Andrew Harp, Geoffrey Irving, Michael Isard, Jia, Y., Rafal Jozefowicz, Lukasz Kaiser, Manjunath Kudlur, . . . Xiaoqiang Zheng. (2015). TensorFlow: Large-scale machine learning on heterogeneous systems [Software available from tensorflow.org]. <https://www.tensorflow.org/>
- Mathewson, K. E., Harrison, T. J. L., & Kizuk, S. A. D. (2017). High and dry? Comparing active dry EEG electrodes to active and passive wet electrodes: Active dry vs. active & passive wet EEG electrodes. *Psychophysiology*, 54(1), 74–82. <https://doi.org/10.1111/psyp.12536>
- McCrimmon, C. M., Fu, J. L., Wang, M., Lopes, L. S., Wang, P. T., Karimi-Bidhendi, A., Liu, C. Y., Heydari, P., Nenadic, Z., & Do, A. H. (2017). Performance Assessment of a Custom, Portable, and Low-Cost Brain-Computer Interface Platform. *IEEE Transactions on Biomedical Engineering*, 64(10), 2313–2320. <https://doi.org/10.1109/TBME.2017.2667579>
- McFarland, D. J., Daly, J., Boulay, C., & Parvaz, M. A. (2017). Therapeutic applications of BCI technologies. *Brain-Computer Interfaces*, 4(1-2), 37–52. <https://doi.org/10.1080/2326263X.2017.1307625>
- Michel, C. M., & Brunet, D. (2019). EEG Source Imaging: A Practical Review of the Analysis Steps. *Frontiers in Neurology*, 10, 325. <https://doi.org/10.3389/fneur.2019.00325>
- Moses, D. A., Metzger, S. L., Liu, J. R., Anumanchipalli, G. K., Makin, J. G., Sun, P. F., Chartier, J., Dougherty, M. E., Liu, P. M., Abrams, G. M., Tu-Chan, A., Ganguly, K., & Chang, E. F. (2021). Neuroprosthesis for Decoding Speech in a Paralyzed Person with Anarthria. *New England Journal of Medicine*, 385(3), 217–227. <https://doi.org/10.1056/NEJMoa2027540>
- Musk, E., & Neuralink. (2019). *An integrated brain-machine interface platform with thousands of channels* (preprint). Neuroscience. <https://doi.org/10.1101/703801>
- Mussi, M., Spindola, M., & Marques, P. (2019). Classification of EEG Motor Imagery signal processed in time-frequency with Convolutional Neural Network.
- Nam, C. S., Nijholt, A., & Lotte, F. (Eds.). (2018). *Brain-Computer Interfaces Handbook: Technological and Theoretical Advances* (1st ed.). CRC Press. <https://doi.org/10.1201/9781351231954>
- Nicolas-Alonso, L. F., & Gomez-Gil, J. (2012). Brain Computer Interfaces, a Review. *Sensors*, 12(2), 1211–1279. <https://doi.org/10.3390/s120201211>
- Nijboer, F., van de Laar, B., Gerritsen, S., Nijholt, A., & Poel, M. (2015). Usability of Three Electroencephalogram Headsets for Brain-Computer Interfaces: A Within Subject Comparison. *Interacting with Computers*, 27(5), 500–511. <https://doi.org/10.1093/iwc/iwv023>
- Niketeghad, S., & Pouratian, N. (2019). Brain Machine Interfaces for Vision Restoration: The Current State of Cortical Visual Prosthetics. *Neurotherapeutics*, 16(1), 134–143. <https://doi.org/10.1007/s13311-018-0660-1>
- Nunez, P. L., & Cutillo, B. A. (Eds.). (1995). *Neocortical dynamics and human EEG rhythms*. Oxford University Press.
- Olivas-Padilla, B. E., & Chacon-Murguia, M. I. (2019). Classification of multiple motor imagery using deep convolutional neural networks and spatial filters. *Applied Soft Computing*, 75, 461–472. <https://doi.org/10.1016/j.asoc.2018.11.031>

- Panoulas, K. J., Hadjileontiadis, L. J., & Panas, S. M. (2010). Brain-Computer Interface (BCI): Types, Processing Perspectives and Applications [Series Title: Smart Innovation, Systems and Technologies]. In R. J. Howlett, L. C. Jain, G. A. Tsihrintzis, & L. C. Jain (Eds.), *Multimedia Services in Intelligent Environments* (pp. 299–321). Springer Berlin Heidelberg. [https://doi.org/10.1007/978-3-642-13396-1\\_14](https://doi.org/10.1007/978-3-642-13396-1_14)
- Pathirana, S., Asirvatham, D., & Johar, G. (2018). A Critical Evaluation on Low-Cost Consumer-Grade Electroencephalographic Devices. *2018 2nd International Conference on BioSignal Analysis, Processing and Systems (ICBAPS)*, 160–165. <https://doi.org/10.1109/ICBAPS.2018.8527413>
- Pedregosa, F., Varoquaux, G., Gramfort, A., Michel, V., Thirion, B., Grisel, O., Blondel, M., Müller, A., Nothman, J., Louppe, G., Prettenhofer, P., Weiss, R., Dubourg, V., Vanderplas, J., Passos, A., Cournapeau, D., Brucher, M., Perrot, M., & Duchesnay, É. (2018). Scikit-learn: Machine Learning in Python [arXiv: 1201.0490]. *arXiv:1201.0490 [cs]*. Retrieved March 14, 2022, from <http://arxiv.org/abs/1201.0490>
- Peterson, V., Galván, C., Hernández, H., & Spies, R. (2020). A feasibility study of a complete low-cost consumer-grade brain-computer interface system. *Heliyon*, 6(3), e03425. <https://doi.org/10.1016/j.heliyon.2020.e03425>
- Pfurtscheller, G., Neuper, C., Flotzinger, D., & Pergenzer, M. (1997). EEG-based discrimination between imagination of right and left hand movement. *Electroencephalography and Clinical Neurophysiology*, 103(6), 642–651. [https://doi.org/10.1016/S0013-4694\(97\)00080-1](https://doi.org/10.1016/S0013-4694(97)00080-1)
- Poorvitha, H. R., Aishwarya, G., Kavya, K. B., & Nayana, R. (2020). Brain Computer Interface for Sleep Apnea Detection. *IJERT*, 8(15). <https://www.ijert.org/brain-computer-interface-for-sleep-apnea-detection>
- Ptito, M., Bleau, M., Djerourou, I., Paré, S., Schneider, F. C., & Chebat, D.-R. (2021). Brain-Machine Interfaces to Assist the Blind. *Frontiers in Human Neuroscience*, 15, 638887. <https://doi.org/10.3389/fnhum.2021.638887>
- Ramesh, A., Pavlov, M., Goh, G., Gray, S., Voss, C., Radford, A., Chen, M., & Sutskever, I. (2021). Zero-Shot Text-to-Image Generation [Publisher: arXiv Version Number: 2]. <https://doi.org/10.48550/ARXIV.2102.12092>
- Rao, P. (2020). Funding for Brain-Computer Interface Ventures. Retrieved February 23, 2022, from <https://www.from-the-interface.com/BCI-venture-funding/>
- Rasheed, S. (2021). A Review of the Role of Machine Learning Techniques towards Brain-Computer Interface Applications. *Machine Learning and Knowledge Extraction*, 3(4), 835–862. <https://doi.org/10.3390/make3040042>
- Rashid, U., Niazi, I., Signal, N., & Taylor, D. (2018). An EEG Experimental Study Evaluating the Performance of Texas Instruments ADS1299. *Sensors*, 18(11), 3721. <https://doi.org/10.3390/s18113721>
- Ratti, E., Waninger, S., Berka, C., Ruffini, G., & Verma, A. (2017). Comparison of Medical and Consumer Wireless EEG Systems for Use in Clinical Trials. *Frontiers in Human Neuroscience*, 11, 398. <https://doi.org/10.3389/fnhum.2017.00398>
- Renton, A. I., Mattingley, J. B., & Painter, D. R. (2019). Optimising non-invasive brain-computer interface systems for free communication between naïve human participants. *Scientific Reports*, 9(1), 18705. <https://doi.org/10.1038/s41598-019-55166-y>
- Rezende, I. N. (2020). Facial recognition in police hands: Assessing the ‘Clearview case’ from a European perspective. *New Journal of European Criminal Law*, 11(3), 375–389. <https://doi.org/10.1177/2032284420948161>
- Saha, S., Mamun, K. A., Ahmed, K., Mostafa, R., Naik, G. R., Darvishi, S., Khandoker, A. H., & Baumert, M. (2021). Progress in Brain Computer Interface: Challenges and Oppor-



- tunities. *Frontiers in Systems Neuroscience*, 15, 578875. <https://doi.org/10.3389/fnsys.2021.578875>
- Schaller, R. (1997). Moore's law: Past, present and future. *IEEE Spectrum*, 34(6), 52–59. <https://doi.org/10.1109/6.591665>
- Schirrneister, R. T., Springenberg, J. T., Fiederer, L. D. J., Glasstetter, M., Eggersperger, K., Tangermann, M., Hutter, F., Burgard, W., & Ball, T. (2017). Deep learning with convolutional neural networks for EEG decoding and visualization: Convolutional Neural Networks in EEG Analysis. *Human Brain Mapping*, 38(11), 5391–5420. <https://doi.org/10.1002/hbm.23730>
- Seol, H. Y., Park, S., Ji, Y. S., Hong, S. H., & Moon, I. J. (2020). Impact of hearing aid noise reduction algorithms on the speech-evoked auditory brainstem response. *Scientific Reports*, 10(1), 10773. <https://doi.org/10.1038/s41598-020-66970-2>
- Shih, J. J., Krusienski, D. J., & Wolpaw, J. R. (2012). Brain-Computer Interfaces in Medicine. *Mayo Clinic Proceedings*, 87(3), 268–279. <https://doi.org/10.1016/j.mayocp.2011.12.008>
- Shyu, K.-K., Lee, P.-L., Lee, M.-H., Lin, M.-H., Lai, R.-J., & Chiu, Y.-J. (2010). Development of a Low-Cost FPGA-Based SSVEP BCI Multimedia Control System. *IEEE Transactions on Biomedical Circuits and Systems*, 4(2), 125–132. <https://doi.org/10.1109/TBCAS.2010.2042595>
- Silver, D., Huang, A., Maddison, C. J., Guez, A., Sifre, L., van den Driessche, G., Schrittwieser, J., Antonoglou, I., Panneershelvam, V., Lanctot, M., Dieleman, S., Grewe, D., Nham, J., Kalchbrenner, N., Sutskever, I., Lillicrap, T., Leach, M., Kavukcuoglu, K., Graepel, T., & Hassabis, D. (2016). Mastering the game of go with deep neural networks and tree search. *Nature*, 529(7587), 484–489. <https://doi.org/10.1038/nature16961>
- Sonam, Y. S. (2018). A Review Paper on Brain Computer Interface. *IJERT*, 3(10). <https://www.ijert.org/a-review-paper-on-brain-computer-interface>
- Sonoda, S., & Murata, N. (2017). Neural network with unbounded activation functions is universal approximator. *Applied and Computational Harmonic Analysis*, 43(2), 233–268. <https://doi.org/10.1016/j.acha.2015.12.005>
- Srinivasan, R. (1999). Methods to improve the spatial resolution of EEG. *International journal of bioelectromagnetism*, 1(1), 102–111.
- Srinivasan, R., Nunez, P. L., Tucker, D. M., Silberstein, R. B., & Cadusch, P. J. (1996). Spatial sampling and filtering of EEG with spline Laplacians to estimate cortical potentials. *Brain Topography*, 8(4), 355–366. <https://doi.org/10.1007/BF01186911>
- Stockman, C. (2020). Can a Technology Teach Meditation? Experiencing the EEG Headband InteraXon Muse as a Meditation Guide. *International Journal of Emerging Technologies in Learning (iJET)*, 15(08), 83. <https://doi.org/10.3991/ijet.v15i08.12415>
- Strickland, E., & Harris, M. (2022). Their bionic eyes are now obsolete and unsupported. <https://spectrum.ieee.org/bionic-eye-obsolete>
- Sudarsan, S., & Sekaran, E. C. (2012). Design and Development of EMG Controlled Prosthetics Limb. *Procedia Engineering*, 38, 3547–3551. <https://doi.org/10.1016/j.proeng.2012.06.409>
- Tangermann, M., Müller, K.-R., Aertsen, A., Birbaumer, N., Braun, C., Brunner, C., Leeb, R., Mehring, C., Miller, K. J., Müller-Putz, G. R., Nolte, G., Pfurtscheller, G., Preissl, H., Schalk, G., Schlögl, A., Vidaurre, C., Waldert, S., & Blankertz, B. (2012). Review of the BCI Competition IV. *Frontiers in Neuroscience*, 6. <https://doi.org/10.3389/fnins.2012.00055>
- Taran, S., & Bajaj, V. (2020). Sleep Apnea Detection Using Artificial Bee Colony Optimize Hermite Basis Functions for EEG Signals. *IEEE Transactions on Instrumentation and Measurement*, 69(2), 608–616. <https://doi.org/10.1109/TIM.2019.2902809>

- The Royal Society. (2019). *Explainable AI: The basics*. Publications Office. <https://royalsociety.org/-/media/policy/projects/explainable-ai/AI-and-interpretability-policy-briefing.pdf>
- Tosi, J., Taffoni, F., Santacatterina, M., Sannino, R., & Formica, D. (2017). Performance Evaluation of Bluetooth Low Energy: A Systematic Review. *Sensors*, 17(12), 2898. <https://doi.org/10.3390/s17122898>
- Tseghai, G. B., Malengier, B., Fante, K. A., & Langenhove, L. V. (2021). The Status of Textile-Based Dry EEG Electrodes. *Autex Research Journal*, 21(1), 63–70. <https://doi.org/10.2478/aut-2019-0071>
- Utsumi, K., Takano, K., Okahara, Y., Komori, T., Onodera, O., & Kansaku, K. (2018). Operation of a P300-based brain-computer interface in patients with Duchenne muscular dystrophy. *Scientific Reports*, 8(1), 1753. <https://doi.org/10.1038/s41598-018-20125-6>
- V. Utkin, L. (2019). An imprecise extension of SVM-based machine learning models. *Neurocomputing*, 331, 18–32. <https://doi.org/10.1016/j.neucom.2018.11.053>
- van Ginneken, B. (2017). Fifty years of computer analysis in chest imaging: Rule-based, machine learning, deep learning. *Radiological Physics and Technology*, 10(1), 23–32. <https://doi.org/10.1007/s12194-017-0394-5>
- Vidal, J. J. (1973). Toward Direct Brain-Computer Communication. *Annual Review of Biophysics and Bioengineering*, 2(1), 157–180. <https://doi.org/10.1146/annurev.bb.02.060173.001105>
- Vlahou, A., Hallinan, D., Apweiler, R., Argiles, A., Beige, J., Benigni, A., Bischoff, R., Black, P. C., Boehm, F., Céraline, J., Chrousos, G. P., Delles, C., Evenepoel, P., Fridolin, I., Glorieux, G., van Gool, A. J., Heidegger, I., Ioannidis, J. P., Jankowski, J., ... Vanholder, R. (2021). Data Sharing Under the General Data Protection Regulation: Time to Harmonize Law and Research Ethics? *Hypertension*, 77(4), 1029–1035. <https://doi.org/10.1161/HYPERTENSIONAHA.120.16340>
- Wang, Y.-T., Wang, Y., & Jung, T.-P. (2011). A cell-phone-based brain-computer interface for communication in daily life. *Journal of Neural Engineering*, 8(2), 025018. <https://doi.org/10.1088/1741-2560/8/2/025018>
- Warren, B., & Randolph, A. B. (2019). Facebrain: A p300 BCI to facebook [Series Title: Lecture Notes in Information Systems and Organisation]. In F. D. Davis, R. Riedl, J. vom Brocke, P.-M. Léger, & A. B. Randolph (Eds.), *Information systems and neuroscience* (pp. 119–124). Springer International Publishing. [https://doi.org/10.1007/978-3-030-01087-4\\_14](https://doi.org/10.1007/978-3-030-01087-4_14)
- Waytowich, N. R., Lawhern, V., Garcia, J. O., Cummings, J., Faller, J., Sajda, P., & Vettel, J. M. (2018). Compact Convolutional Neural Networks for Classification of Asynchronous Steady-state Visual Evoked Potentials [Publisher: arXiv Version Number: 2]. <https://doi.org/10.48550/ARXIV.1803.04566>
- Wolpaw, J., & Wolpaw, E. W. (2012). *Brain-Computer Interfaces Principles and Practice*. Oxford University Press. <https://doi.org/10.1093/acprof:oso/9780195388855.001.0001>
- Wolpaw, J., Birbaumer, N., Heetderks, W., McFarland, D., Peckham, P., Schalk, G., Donchin, E., Quatrano, L., Robinson, C., & Vaughan, T. (2000). Brain-computer interface technology: A review of the first international meeting [Conference Name: IEEE Transactions on Rehabilitation Engineering]. *IEEE Transactions on Rehabilitation Engineering*, 8(2), 164–173. <https://doi.org/10/dz4238>
- Won, K., Kwon, M., Jang, S., Ahn, M., & Jun, S. C. (2019). P300 Speller Performance Predictor Based on RSVP Multi-feature. *Frontiers in Human Neuroscience*, 13, 261. <https://doi.org/10.3389/fnhum.2019.00261>
- Wong, L. L. N., Chen, Y., Wang, Q., & Kuehnel, V. (2018). Efficacy of a Hearing Aid Noise Reduction Function. *Trends in Hearing*, 22, 233121651878283. <https://doi.org/10.1177/2331216518782839>

- World Health Organization. (2018). *Global status report on road safety 2018*. World Health Organization. Retrieved March 15, 2022, from <https://apps.who.int/iris/handle/10665/276462>
- Xu, B., Li, W., Liu, D., Zhang, K., Miao, M., Xu, G., & Song, A. (2022). Continuous Hybrid BCI Control for Robotic Arm Using Noninvasive Electroencephalogram, Computer Vision, and Eye Tracking. *Mathematics*, 10(4), 618. <https://doi.org/10.3390/math10040618>
- Xu, G., Shen, X., Chen, S., Zong, Y., Zhang, C., Yue, H., Liu, M., Chen, F., & Che, W. (2019). A Deep Transfer Convolutional Neural Network Framework for EEG Signal Classification [Conference Name: IEEE Access]. *IEEE Access*, 7, 112767–112776. <https://doi.org/10.1109/ACCESS.2019.2930958>
- Xu, J., Mitra, S., Van Hoof, C., Yazicioglu, R. F., & Makinwa, K. A. A. (2017). Active Electrodes for Wearable EEG Acquisition: Review and Electronics Design Methodology [Conference Name: IEEE Reviews in Biomedical Engineering]. *IEEE Reviews in Biomedical Engineering*, 10, 187–198. <https://doi.org/10.1109/RBME.2017.2656388>
- Yao, D., Qin, Y., Hu, S., Dong, L., Bringas Vega, M. L., & Valdés Sosa, P. A. (2019). Which Reference Should We Use for EEG and ERP practice? *Brain Topography*, 32(4), 530–549. <https://doi.org/10.1007/s10548-019-00707-x>
- Zhuang, J., & Yin, G. (2017). Motion control of a four-wheel-independent-drive electric vehicle by motor imagery EEG based BCI system. *2017 36th Chinese Control Conference (CCC)*, 5449–5454. <https://doi.org/10.23919/ChiCC.2017.8028220>