



**DALMIA CEMENT (B) LIMITED – CEMENT PLANT  
ARIYALUR  
INSTRUMENTATION DEPARTMENT**



Issue No. 01	Rev. No: 00	Effective Date: 01.11.2010	WI-15 (SP-8; AM-7)
Issued By: M.R		Approved By: HOD-INSTRUMENTATION	
CALIBRATION PROCEDURE OF PRESSURE CALIBRATOR			

**CALIBRATION PROCEDURE:**

- 1.1 Head of department gets the Pressure Calibrator calibrated once in a year by a suitable sub-contractor for calibration, who maintains calibration standards traceable to the national standards. Please refer to the List of approved sub-contractors for calibration,
- 2.0 Maintain calibration record in the Instrument calibration list.