



**DALMIA CEMENT (B) LIMITED – CEMENT PLANT
ARIYALUR
INSTRUMENTATION DEPARTMENT**

Issue No. 02	Rev. No: 00	Effective Date: 01.11.2019	WI-14 (SP-8; AM-7)
Issued By: M.R		Approved By: HOD-INSTRUMENTATION	
CALIBRATION PROCEDURE OF DIGITAL MULTIMETER			

CALIBRATION PROCEDURE:

- 1.1 Head of department gets the Multimeters calibrated once in a year by a suitable Sub-Contractor for calibration, who maintains calibration standards traceable to the national standards.

Please refer to the List of approved sub-contractors for calibration.

- 2.0 Maintain calibration record in the Instrument calibration list.