



**DALMIA CEMENT (B) LIMITED – CEMENT PLANT
ARIYALUR
INSTRUMENTATION MANUAL**



Issue No. 01	Rev. No: 02	Effective Date: 20.09.2017	INST/SOP/038
Issued By: M.R		Approved By: HOD - INSTRUMENTATION	
SOP for Offline PSA System Start and shutdown procedure			

Startup Procedure

- Turn-On PC On Button
- Wait until Login window appeared
- Select the User Name : DalmiaCements
- Double click the Mastersizer 2000
- The message will display Current User Name
- Please ensure the user name Dalmia Cement or Bmi_Cims and Click OK.
- Once Master Sizer opened do the following for set system in remote mode
 - Tools->Options->Select Remote Mode
- Double Click Double Agent Icon at Desktop
- Check Double agent Connection at right side of windows task bar. It shall be display as connected when mouse tracking to Double agent icon at windows task bar.
- If it shows Connecting / Not connected, Check the network connection.

Shutdown Procedure:

- Tracking the mouse to Double agent Icon at windows task bar right side
- Right click the mouse and select Exit
- Do the following for Close the Master sizer 2000
 - File->Exit
- Wait until complete close of Malvern Windows.
- Select Windows Start button and select Shutdown for shutdown the PC
 - Start->Shutdown