

INDIAN INSTITUTE OF TECHNOLOGY BOMBAY MATERIALS MANAGEMENT DIVISION Powai, Mumbai - 400076

<u>Technical Specifications of TLC Mass Spectrometer</u> <u>RFx No. 6100001150 (Reference No. 1000026692)</u>

| Sr. No. | Technical Specifications for TLC Mass Spectrometer |
|------------|---|
| 1 | Mass detector |
| | a. Reaction Monitoring for compound identification and conformity via Direct Injection for crude reaction mixtures. |
| | b. Should be upgradeable and interfaced with to any make UPLC or HPLC systems. |
| | c. Standalone system. |
| | d. Should be able to extract TLC spots automatically to find out the molecular weight. |
| 2 | Specifications: |
| | Ion Source: ESI (Electron Spray Ionization) |
| 3 | API: Atmospheric pressure interface & Ionization interface with orthogonal ion sampling from |
| | heated capillary-allows for small single Turbo pumps |
| 4 | Analyzer: Single Quadrupole analyzer |
| 5 | Detector: Electron multiplier with high-energy conversion Dynode. |
| 6 | Flow rate range: 10μL/ min to 1mL/ min or more |
| 7 | Polarity: Positive & negative ion switching in single analysis |
| 8 | Polarity switching speed:50 ms. |
| 9 | m/z Range: Up to m/z 2000 or better. |
| 10 | Acquisition rate: 10,000 m/z units / sec or higher. |
| 11 | Resolution :0.5 to 0.7 m/z units (FWHM) at 1000 m/z units /sec over entire acquisition range |
| 12 | Sensitivity :10 pg Reserpine (FIA -5μ L sample injection volume at 100 μ L/min) 100:1 S/N (RMS) with SIM of m/z609.28 |
| 13 | Accuracy: +0.1 m/z units over entire acquisition range. |
| 14 | Stability: +0.1 m/z units for over 12-hour period. |
| 15 | Dynamic range: 4.5 orders of magnitude. |
| 16 | Pump for make-up flow (HPLC Pump) & Suitable Vacuum pumps should be supplied. |
| 17 | Necessary accessories for smooth functioning of the system must be supplied. |
| 18 | Suitable computer and Nitrogen generator should be supplied along with the system. |
| 19 | 1 year warranty from the date of successful installation. |