

Table 3. Bilateral ZUPT stride-length accuracy by condition. n = participants; N = matched stride pairs; bias = mean IMU-minus-GaitRite error; LoA = Bland–Altman 95% limits of agreement.

Condition	n	N pairs	Bias (m)	RMSE (m)	MAE (m)	95% LoA (m)	ICC(2,1)
Single-task	32	2,544	+0.007	0.040	0.027	[-0.069, +0.084]	0.973
Serial 2s	30	2,618	+0.001	0.037	0.024	[-0.072, +0.073]	0.977
Serial 3s	32	2,770	+0.000	0.039	0.024	[-0.077, +0.077]	0.975
Serial 7s	32	2,540	+0.003	0.039	0.025	[-0.073, +0.078]	0.971
Pooled	32	10,472	+0.003	0.039	0.025	[-0.073, +0.078]	0.976