Healthy nerve-trunks feel it very little, but to it neuralgia is most amenable, as is also the function of the optic nerve.

Ordinary primary current batteries, however, to be effective must be bulky and expensive; and judging from the practical success of the system herein detailed, almost needless; cures being rapidly obtained in curable cases by simpler and more portable, as well as less costly apparatus; a point on which superiority may justly be claimed for the system itself.

Acute lesions or acute morbid processes, even of the brain and spinal cord, are no contra-indications for the proper use of electricity as a curative. In this, of course, the author is aware that she is contradicted by most practitioners, and yet can but reiterate the statement from the abundant teachings of a long and varied experience.

Except for purposes of rapid cauterization (comparatively but rarely needed, even for the separation of pedunculated tumors, as will be seen under that head, in this volume), the combination batteries of the author are quite adequate to all the exigencies of Electro-Therapeutics. These batteries consist of zinc and carbon plates (immersed in a solution), with a system of helices, whereby the tension of the resulting currents is modified, as every factor in every current must modify the current itself; and the primary current is also combined with the secondary in actual administration in greatly different degrees.

Hence, according to the posts (lettered from A to