

PET Imaging

Positron annihilation

511 keV photons in opposite directions

Coincidence detection

Simultaneous detection localizes source

Radiotracers (FDG, etc.)

FDG shows glucose metabolism

SUV calculations

Standardized uptake value quantification

PET/CT integration

Functional and anatomical fusion

