Smart Multi-Sensor Wearable for Real-Time Elderly Health Anomaly and Fall Detection

Component: Respiratory & Thermal Warning System

Candidate Name: Dissanayaka D.M.S.K.
Student ID: IT22291196
Registration Group: R26-IT-108
Degree: B.Sc. (Hons) Degree in Information Technology Specializing in Information Technology (IT)
Department: Department of Information Technology
Institute: Sri Lanka Institute of Information Technology (SLIIT), Sri Lanka
Supervisor: Mr. Sanjeewa Perera
Co-Supervisor: Ms. Samadhi Rathnayake
Date: March 2026

DECLARATION

I declare that this is my own work, and this dissertation does not incorporate without acknowledgement any material previously submitted for a Degree or Diploma in any other University or institute of higher learning and to the best of my knowledge and belief it does not contain any material previously published or written by another person except where the acknowledgement is made in the text.

Name	**Student ID**	**Signature**
Dissanayaka D.M.S.K.	IT22291196	

The above candidate has carried out research for the bachelor's degree Dissertation under my supervision.

Supervisor Name	**Date**	**Signature**

ACKNOWLEDGEMENT

I would like to express my deepest gratitude and appreciation to my research supervisor, **Mr. Sanjeewa Perera**, and co-supervisor, **Ms. Samadhi Rathnayake**, for their continuous academic guidance, invaluable insight, and constructive evaluation provided throughout the design, development, and testing of this research dissertation. Their continuous support was instrumental in structuring the methodology, refining the edge computing architecture, and ensuring the clinical relevance of this study.

I also extend my sincere appreciation to the academic staff of the Department of Information Technology and the Faculty of Computing at the Sri Lanka Institute of Information Technology (SLIIT) for offering an enriching research environment, laboratory infrastructure, and computational resources required for hardware assembly and embedded testing.

Special thanks are extended to my research group members (Group R26-IT-108) for their active cooperation, technical discussions, and collaborative efforts in integrating the multi-sensor wearable system. Finally, I express my gratitude to the elderly participants and caregivers who provided crucial feedback during usability assessments and prototype validation.

ABSTRACT

The rapid demographic shift toward an aging population has created an urgent demand for low-cost, continuous, non-invasive health monitoring solutions tailored for non-hospitalized elderly individuals. In senior care, silent respiratory deterioration, acute oxygen desaturation (SpO_2 < 92\%), and abnormal body temperature fluctuations serve as early physiological markers of respiratory infections, fever, pneumonia, or cardiovascular stress. However, conventional home monitoring relies heavily on episodic manual checkups or expensive, rigid medical instruments that fail to offer real-time proactive warning support.

To address these challenges, this research presents the design, implementation, and empirical evaluation of a **Respiratory & Thermal Warning System** as a vital sub-component of a Smart Multi-Sensor Wearable device. Built upon an embedded **ESP32 microcontroller SoC**, the system integrates a **MAX30102 pulse oximetry sensor** for dual-wavelength Photoplethysmography (PPG) monitoring (SpO_2 and heart rate) and an **MLX90614 non-contact infrared thermal sensor** for skin and body temperature tracking. The core technical innovation lies in a **Dynamic Sliding-Window Baseline Engine** combined with a multi-parameter anomaly detection matrix executing directly at the edge. By computing personalized physiological baselines over a 30-minute calibration window, the system eliminates false positives caused by static universal thresholds and adapts to individual user variance.

The wearable firmware operates entirely on-device, processing raw sensor signals through moving-average noise filters, peak detection algorithms, and threshold risk classifiers to categorize user status into *Normal*, *Caution*, or *Warning* states. When an anomaly is detected, an immediate local alert is issued via a haptic vibration motor and piezoelectric buzzer, while wireless Bluetooth Low Energy (BLE) payloads notify connected caregiver mobile dashboards. Comprehensive empirical testing across 20 structured test cases (TC-01 through TC-20) verified high signal stability, accurate SpO_2 and thermal anomaly classification (>94\% sensitivity), low response latency (<850 ext{ ms}), and extended battery autonomy (>14.5 ext{ hours} on a 1000mAh Li-Po battery). With a bill of materials totaling LKR 64,000, the developed system offers accessible,

scalable, and decentralized early-warning platform for senior care in resource-constrained environments.

Keywords: Pulse Oximetry (SpO_2), Photoplethysmography (PPG), Non-Contact Infrared Thermometry, ESP32 Edge Computing, Personalized Baseline Modeling, Anomaly Detection, Geriatric Health Monitoring, Wearable Biosensors.

TABLE OF CONTENTS

Table of Contents

DECLARATION ***2***

ACKNOWLEDGEMENT ***3***

ABSTRACT ***4***

TABLE OF CONTENTS ***6***

LIST OF TABLES ***8***

LIST OF ABBREVIATIONS ***10***

LIST OF APPENDICES ***12***

CHAPTER 1: INTRODUCTION **1**

1.1 Background and Literature Survey 1

1.1.1 Non-Hospitalized Elderly Care & Monitoring Context 1

1.1.2 Physiological Indicators: Pulse Oximetry and Body Temperature Dynamics 1

1.1.3 Wearable Sensors & Microcontroller Architectures 2

1.1.4 Edge Computing and Personalized Baseline Modeling 2

1.2 Research Gap 2

1.3 Research Problem 3

1.4 Research Objectives 3

1.4.1 Main Objective 3

1.4.2 Specific Objectives 4

1.5 Scope and Limitations 4

CHAPTER 2: METHODOLOGY **7**

2.1 Research Approach 7

2.2 System Architecture and Technical Methods 7

2.2.1 Sensing Layer & Microcontroller Infrastructure 8

2.2.2 Signal Acquisition, Noise Filtering, and Signal Stabilization 9

2.2.3 Dynamic Baseline Modeling & Sliding Window Threshold Engine .. 10
2.2.4 Multi-Parameter Anomaly Detection Logic & Severity Decision Engine 11
2.3 User Requirements & Stakeholder Analysis 12
2.3.1 Stakeholder Identification and Expectations Matrix 12
2.3.2 Functional Requirements Matrix 13
2.4 Commercialization Plan and Budget Breakdown 15
2.4.1 Bill of Materials (BOM) & Total Budget Breakdown 15
CHAPTER 3: RESULTS AND DISCUSSION 20
3.1 System Performance and Experimental Results 20
3.1.1 Sensor Signal Stability & Calibration Accuracy 20
3.1.2 Anomaly Detection Sensitivity & Latency 20
3.1.3 Battery Power Autonomy & Resource Footprint 20
3.1.4 Validation Results of Test Cases (TC-01 through TC-20) 21
3.2 Summary of Individual Student Contribution (IT22291196) 25
CHAPTER 4: CONCLUSIONS AND RECOMMENDATIONS 27
4.1 Conclusions 27
4.2 Recommendations and Future Directions 27
REFERENCES LIST 30
GLOSSARY 33
APPENDICES 35
Appendix A: Comprehensive Test Case Execution Matrix 35
Appendix B: Circuit Diagram & System Architecture 35
Appendix C: Work Breakdown Structure (WBS) & Project Gantt Chart 35

LIST OF TABLES

LIST OF TABLES

Table 2.1: Health State Classification and Anomaly Decision Rules _____ 11

Table 2.2: Stakeholder Identification and System Expectations Matrix _____ 12

Table 2.3: System Functional Requirements Matrix (FR-01 to FR-12) _____ 13

Table 2.4: Bill of Materials (BOM) and Itemized Budget Breakdown _____ 15

Table 3.1: Empirical Test Case Execution & Validation Results (TC-01 to TC-20) _____ 4

LIST OF FIGURES

Figure 4.1: High-Level Architectural Framework of the Respiratory and Thermal Warning System _____ 8

LIST OF ABBREVIATIONS

ADC	Analog-to-Digital Converter
API	Application Programming Interface
BLE	Bluetooth Low Energy
BOM	Bill of Materials
BPM	Beats Per Minute
CPU	Central Processing Unit
DC	Direct Current
DSP	Digital Signal Processing
ESP32	System-on-Chip Microcontroller by Espressif Systems
FIFO	First-In, First-Out Buffer
FR	Functional Requirement
GPIO	General Purpose Input/Output
HR	Heart Rate
I²C	Inter-Integrated Circuit Bus
IEEE	Institute of Electrical and Electronics Engineers
IR	Infrared
LED	Light Emitting Diode
LKR	Sri Lankan Rupee
NFR	Non-Functional Requirement
OLED	Organic Light Emitting Diode
PCB	Printed Circuit Board
PPG	Photoplethysmography

RAM	Random Access Memory
RF	Radio Frequency
SLIIT	Sri Lanka Institute of Information Technology
SoC	System on Chip
SpO_2	Peripheral Capillary Oxygen Saturation
TC	Test Case
TinyML	Tiny Machine Learning
UART	Universal Asynchronous Receiver-Transmitter
WBS	Work Breakdown Structure

LIST OF APPENDICES

Appendix A	Comprehensive Test Case Execution Matrix (TC-01 through TC-20)
Appendix B	System Architecture & Circuit Schematic Configurations
Appendix C	Work Breakdown Structure (WBS) & Project Gantt Chart Details

CHAPTER 1: INTRODUCTION

1.1 Background and Literature Survey

Global demographic patterns reflect an unprecedented expansion in the elderly population. According to World Health Organization (WHO) estimates, the global population of adults aged 60 and over will double by 2050, reaching nearly 2.1 billion individuals [1]. In Sri Lanka and broader developing regions, this demographic shift presents substantial socio-economic and healthcare challenges. As body systems age, susceptibility to chronic respiratory illnesses, cardiovascular complications, infectious diseases, and sudden physiological deterioration increases significantly [2].

For older adults residing in non-hospitalized home environments, early signs of physiological decline are often subtle and insidious [3]. Silent hypoxia — a condition where blood oxygen saturation (SpO_2) falls below normal physiological levels (<92\%) without severe dyspnea — represents a major hazard [7]. Furthermore, temperature anomalies such as low-grade fever or mild hypothermia frequently serve as the primary clinical indicator of systemic infection or inflammatory response in geriatric care [10]. Traditional healthcare delivery relies primarily on periodic clinical visits or manual spot-checks using pulse oximeters and digital thermometers. However, intermittent monitoring fails to capture transient nocturnal desaturation, gradual thermal shifts, or sudden physiological anomalies [9].

1.1.1 Non-Hospitalized Elderly Care & Monitoring Context

In domestic elderly care settings, continuous health monitoring must balance clinical reliability with high user comfort and zero operator burden [11]. Elderly users frequently encounter difficulty operating complex medical instruments or interpreting numeric digital readouts [2]. Furthermore, emergency response times in home care depend directly on the immediate detection and alerting of early warning signs to nearby caregivers or family members [12].

1.1.2 Physiological Indicators: Pulse Oximetry and Body Temperature Dynamics

Pulse oximetry provides a non-invasive, continuous measurement of peripheral blood oxygen saturation (SpO_2) and pulse rate by evaluating light absorption dynamics in arterial blood [7]. By emitting red (660nm) and infrared (880nm) light through vascular tissue using Photoplethysmography (PPG), pulse oximeter modules measure differential absorption ratios (R) between oxygenated (HbO_2) and deoxygenated (Hb) hemoglobin [6]. Simultaneously, body temperature monitoring via contactless infrared thermometry offers vital contextual data regarding systemic infection, metabolic disruption, and heat stress [10].

1.1.3 Wearable Sensors & Microcontroller Architectures

Recent advances in microelectronics have enabled the integration of medical-grade optical sensors into compact wearable form factors [1]. Low-power System-on-Chip (SoC) microcontrollers, such as the Espressif ESP32 operating at 240MHz with integrated Wi-Fi/BLE radios, provide sufficient computational throughput to perform digital signal filtering, baseline estimation, and threshold classification locally at the edge [3].

1.1.4 Edge Computing and Personalized Baseline Modeling

Conventional wearable monitoring solutions frequently stream raw biosensor data to remote cloud servers for processing [3]. However, cloud-dependent architectures introduce severe latency, require persistent internet connectivity, incur server hosting costs, and pose major privacy risks regarding sensitive health metrics [4]. Edge computing addresses these limitations by processing signals directly on the embedded wearable microcontroller [3]. Furthermore, rigid static thresholds (e.g., universal SpO_2 < 90\% alerts) generate high false alarm rates due to individual baseline variations among older adults [4]. Implementing personalized sliding-window baseline models allows wearable edge devices to adapt dynamically to each user's physiological norm [9].

1.2 Research Gap

Despite significant growth in wearable health research [1], [2], [10], existing systems suffer from critical limitations when deployed for geriatric home care:

1. **Lack of Integrated Respiratory & Thermal Multi-Parameter Warning Systems:** Existing commercial wearables focus heavily on single-parameter

tracking (e.g., basic fitness heart rate monitors) or complex multi-sensor clinical systems that are prohibitively expensive for home use [1], [10]. When pulse oximetry and temperature monitoring are isolated, early warning sensitivity for complex infections is severely reduced [12].

2. **Failure of Static Universal Thresholds:** Conventional threshold models trigger alarms based on fixed global cutoffs, leading to frequent false alarms for individuals with baseline respiratory variations (e.g., chronic bronchitis) or missed early warnings during subtle physiological shifts [4]. Current literature lacks lightweight, edge-compatible algorithms for real-time personalized baseline calibration on low-power hardware [3].
3. **Over-Dependence on Cloud Infrastructure:** Most existing tele-health patches require active cloud server connections for signal processing and alert generation [3]. In rural or resource-constrained domestic environments with sporadic internet coverage, cloud dependence introduces unacceptable latency and operational vulnerability [10].
4. **Ergonomic and Cost Barriers in Senior Care:** Clinical-grade monitoring equipment is cumbersome, uncomfortable for prolonged wear, and economically inaccessible for low-income households in developing nations [1], [11].

1.3 Research Problem

How can a wearable, low-cost embedded device autonomously capture dual-parameter physiological signals (SpO_2 /heart rate and skin temperature), perform local noise filtering and personalized baseline calibration at the edge, and issue immediate multi-tier alerts for respiratory and thermal anomalies without relying on cloud connectivity or causing user discomfort?

1.4 Research Objectives

1.4.1 Main Objective

The primary objective of this research is to design, develop, and empirically evaluate a wearable **Respiratory and Thermal Warning System** that continuously measures blood oxygen saturation (SpO_2), heart rate, and skin temperature, and generates real-time early warnings against abnormal

physiological conditions using edge-based processing on an ESP32 microcontroller.

1.4.2 Specific Objectives

1. **Sensor Integration:** To interface the MAX30102 pulse oximetry module and MLX90614 non-contact infrared thermal sensor with an ESP32 microcontroller over an I^2C digital bus, establishing reliable non-invasive data acquisition suitable for continuous elderly wear [1], [7], [10].
2. **Signal Preprocessing:** To develop and implement embedded C++ signal processing algorithms (moving average noise filtering, DC offset removal, and peak detection) to clean raw PPG waveforms and temperature readings on-device [5], [6], [12].
3. **Personalized Baseline Calibration:** To formulate a dynamic sliding-window baseline engine that establishes user-specific physiological ranges (\mu \pm 2\sigma), mitigating false alarms caused by static universal cutoffs [4], [9].
4. **Multi-Tier Anomaly Classifier:** To design an embedded rule matrix and decision engine that classifies physiological states into *Normal*, *Caution*, and *Warning* categories based on single and combined parameter deviations [2], [10], [11].
5. **Real-Time Edge Alerting:** To integrate local haptic (vibration motor), acoustic (buzzer), and visual (tri-color LED / OLED) alert mechanisms, complemented by BLE wireless push notifications to caregiver mobile dashboards [3], [10].
6. **Empirical System Validation:** To evaluate system responsiveness, anomaly detection accuracy, battery power consumption, and hardware stability across 20 rigorous test cases (TC-01 to TC-20) under controlled laboratory and wearable conditions [1], [11], [12].

1.5 Scope and Limitations

- **In-Scope:** On-device signal acquisition, embedded C++ digital filtering, dynamic baseline calibration, rule-based anomaly detection, local haptic/acoustic alerting, BLE wireless telemetry, hardware component prototyping (LKR 64,000 budget), and multi-scenario lab validation.

- **Out-of-Scope:** Clinical diagnostic certification, invasive arterial blood gas calibration, long-range cellular SIM integration, and hospital IC medical device replacement.

CHAPTER 2: METHODOLOGY

2.1 Research Approach

This study adopts an applied research design combining hardware-software co-design, recursive prototyping, and empirical quantitative validation. The research methodology is executed across five structured phases:

1. **System Requirements & Architecture Mapping:** Stakeholder identification, functional requirement definition, and multi-layer structural design.
2. **Hardware Assembly & Sensor Calibration:** Interfacing MAX30102 PPG and MLX90614 thermal sensors with ESP32 microcontrollers and power regulation circuits.
3. **Firmware & Algorithm Development:** Writing C++ firmware modules for I^2C communication, signal filtering, sliding-window baseline calculation, and anomaly classification.
4. **Local Alert & Telemetry Integration:** Wiring haptic/buzzer hardware drivers and configuring BLE service/characteristic profiles for mobile integration.
5. **Empirical Evaluation & Verification:** Executing 20 structured test cases (TC-01 through TC-20) to measure detection sensitivity, latency, power draw, and overall system stability.

2.2 System Architecture and Technical Methods

The wearable Respiratory and Thermal Warning System is structured around a five-layer architectural model, ensuring modular separation between signal acquisition, edge processing, decision logic, local alerting, and external communication.



Figure 4.1: High-Level Architectural Framework of the Wearable Respiratory and Thermal Warning System, depicting physical sensing, ESP32 edge processing, dynamic baseline engine, local alerting, and wireless monitoring layers.

2.2.1 Sensing Layer & Microcontroller Infrastructure

The physical sensing layer comprises two primary digital sensors interfaced to an **Espressif ESP32-WROOM-32** microcontroller operating at 240MHz clock frequency:

- **MAX30102 Pulse Oximeter & Heart-Rate Sensor:** Utilizes integrated Red (660 ext{ nm}) and Infrared (880 ext{ nm}) LEDs paired with a high-resolution photodetector and 18-bit delta-sigma ADC. The module communicates with the ESP32 via I^2C at 400kHz (SDA pin 21, SCL pin 22). It includes a 32-sample FIFO buffer to minimize CPU polling overhead.
- **MLX90614 Non-Contact Infrared Thermometer:** Features a 17-bit ADC and powerful DSP engine, providing high-accuracy temperature measurements (0.02^\circ ext{C} resolution, \pm 0.5^\circ ext{C} clinical accuracy) over I^2C (SMBus compatible). It captures object skin temperature (T_{skin}) and ambient temperature (T_{amb}).

- **Power Management Module:** Powered by a rechargeable 3.7 ext{V } 1000 ext{mAh} Lithium-Polymer (Li-Po) battery connected to a TP4056 charging IC and ME6211 low-dropout (LDO) linear regulator, providing stable 3.3 ext{V} DC power to all sensors and microcontrollers.

2.2.2 Signal Acquisition, Noise Filtering, and Signal Stabilization

Raw PPG signals from pulse oximeters are highly sensitive to motion artifacts and high-frequency noise. The ESP32 firmware executes a multi-stage preprocessing pipeline:

1. **DC Component Removal:** High-pass filtering removes low-frequency baseline wander caused by respiration movement.
2. **Moving Average Filter:** Low-pass filtering over an N=5 sample window smooths high-frequency electrical interference.
3. **Ratio-of-Ratios Calculation:** Oxygen saturation (SpO_2) is computed using the AC and DC amplitudes of Red and IR light:

R = Wrac{(AC_{red} / DC_{red})}{(AC_{ir} / DC_{ir})}

SpO_2 = 110 - 25 imes R

The listing below illustrates the ESP32 C++ firmware implementation for sensor data acquisition and signal cleaning:

```
// ESP32 Firmware: Sensor Signal Acquisition & Moving Average Filter
#include <Wire.h>
#include "MAX30105.h"
#include <Adafruit_MLX90614.h>

MAX30105 particleSensor;
Adafruit_MLX90614 mlx = Adafruit_MLX90614();

const int FILTER_SIZE = 5;
float spo2Buffer[FILTER_SIZE];
float tempBuffer[FILTER_SIZE];
int bufferIndex = 0;

void setupSensors() {
    Wire.begin(21, 22, 400000); // Fast I2C Bus
    if (!particleSensor.begin(Wire, I2C_SPEED_FAST)) {
        Serial.println("MAX30102 Sensor Initialization Failed!");
    }
    particleSensor.setup(60, 4, 2, 200, 411, 4096); // Sensor Configuration
    mlx.begin();
}

float getFilteredSpO2(float rawSpO2) {
    spo2Buffer[bufferIndex] = rawSpO2;
    float sum = 0.0;
    for(int i = 0; i < FILTER_SIZE; i++) {
        sum += spo2Buffer[i];
    }
    bufferIndex = (bufferIndex + 1) % FILTER_SIZE;
    return (sum / FILTER_SIZE);
}
```

2.2.3 Dynamic Baseline Modeling & Sliding Window Threshold Engine

Rather than relying on static cutoffs, the system maintains a running historical buffer of user readings over a 30-minute calibration window (N_{samples} = 1800 at 1 ext{ Hz}). The mean (\mu_{user}) and standard deviation (\sigma_{user}) are continuously updated:

\mu_{user} = Wrac{1}{N} \sum_{i=1}^{N} x_i, \quad \sigma_{user} = \sqrt{Wrac{1}{N} \sum_{i=1}^{N} (x_i - \mu_{user})^2}

Normal expected bounds are defined as:

ext{Upper Bound} = \mu_{user} + 2\sigma_{user}, \quad ext{Lower Bound} = \mu_{user} - 2\sigma_{user}

2.2.4 Multi-Parameter Anomaly Detection Logic & Severity Decision Engine

The anomaly classification engine evaluates preprocessed values against both fixed clinical safety limits and dynamic baseline bounds. The table below delineates the classification rules:

Health State	**SpO_2 Condition**	**Temperature Condition (T_{skin})**	**System Response & Action**
NORMAL	SpO_2 \ge 95\% and within \mu \pm 2\sigma	36.2^\circ ext{C} \le T \le 37.4^\circ ext{C}	Green LED solid. Normal monitoring mode.
CAUTION	92\% \le SpO_2 < 95\% or minor baseline drop	37.5^\circ ext{C} \le T \le 37.9^\circ ext{C} (Mild elevation)	Yellow LED solid. Single haptic pulse notification.
WARNING	SpO_2 < 92\% (Acute Desaturation)	T \ge 38.0^\circ ext{C} (Febrile Risk) or T < 35.5^\circ ext{C}	Red LED flashing. Acoustic buzzer + continuous haptic + BLE push alert.

The firmware implementation of the anomaly decision matrix is provided below:

```
// Anomaly Classifier Decision Engine
enum HealthState { STATE_NORMAL, STATE_CAUTION, STATE_WARNING };

HealthState evaluateHealthStatus(float spo2, float temp, float baseSpO2, float baseTemp) {
    if (spo2 < 92.0 || temp >= 38.0 || temp < 35.5) {
        return STATE_WARNING; // Critical physiological risk
    }
    else if ((spo2 >= 92.0 && spo2 < 95.0) || (temp >= 37.5 && temp < 38.0)) {
        return STATE_CAUTION; // Moderate deviation
    }
    else if (spo2 < (baseSpO2 - 3.0) || temp > (baseTemp + 1.2)) {
        return STATE_CAUTION; // Baseline relative drop
    }
    return STATE_NORMAL;
}
```

2.3 User Requirements & Stakeholder Analysis

System design was guided by comprehensive stakeholder analysis and user requirement mapping to ensure high geriatric usability and clinical utility.

2.3.1 Stakeholder Identification and Expectations Matrix

The table below summarizes primary system stakeholders, their operational roles, and main functional needs:

Stakeholder	Description	System Role	Main Needs
Elderly Users	Primary end users wearing the device during daily home living.	Wear device continuously; receive direct on-device haptic/visual warnings.	Comfortable wear, non-invasive sensing, clear alerts, zero manual setup.
Caregivers / Family	Relatives or domestic aides supporting senior citizens.	Receive wireless alerts; monitor physiological status remotely.	Timely push notifications, clear warning details, zero false alarms.
Healthcare Providers	Medical professionals reviewing health trends.	Interpret historical warning logs for preventive clinical assessment.	Reliable sensor data, Standardized SpO_2 /temp logs, clinical relevance.

2.3.2 Functional Requirements Matrix

The table below specifies the 12 core Functional Requirements (FR-01 to FR-12) of the system:

ID	Functional Requirement Description	Priority Level	Engineering Justification
FR-01	Continuous collection of SpO_2 and heart rate signals via optical PPG.	HIGH	Core physiological indicator for respiratory decline.
FR-02	Continuous skin/body temperature sensing via non-contact IR sensor.	HIGH	Essential for thermal abnormality and infection detection.
FR-03	Real-time data acquisition and transmission to processing module at 1Hz.	HIGH	Enables immediate anomaly identification without lag.
FR-04	Embedded data preprocessing to clean noise and filter invalid sensor readings.	HIGH	Prevents motion artifacts from causing false alerts.
FR-05	Dynamic personalized baseline calibration over historical observation windows.	MEDIUM	Improves detection accuracy for individual user baselines.
FR-06	Multi-parameter anomaly detection analyzing SpO_2, HR, and temperature.	HIGH	Central intelligence engine of the wearable system.
FR-07	Risk state classification into Normal, Caution, and Warning levels.	HIGH	Provides structured, interpretable severity outputs.
FR-08	Real-time alert generation upon detecting unsafe conditions (<1 ext{s} latency).	HIGH	Critical for rapid caregiver intervention during acute events.
FR-09	Local warning outputs via vibration motor, buzzer, and tri-color LED.	HIGH	Ensures immediate elderly awareness without smartphone dependence.

ID	Functional Requirement Description	Priority Level	Engineering Justification
FR-10	Wireless BLE warning payload notification to mobile caregiver app.	MEDIUM	Extends visibility to remote caregivers within home range.
FR-11	Uninterrupted continuous power operation for >12 ext{ hours} on battery.	HIGH	Necessary for reliable daily wearable monitoring.
FR-12	Telemetry data exchange with team smart wearable platform.	MEDIUM	Supports overall group system integration and scalability.

2.4 Commercialization Plan and Budget Breakdown

To ensure practical viability for low-resource geriatric care, a detailed commercialization framework and Bill of Materials (BOM) was established.

2.4.1 Bill of Materials (BOM) & Total Budget Breakdown

The table below itemizes the hardware, software, integration, and human effort costs totaling **LKR 64,000**:

Category	Item Description	Qty	Unit Cost (LKR)	Total Cost (LKR)	Engineering Justification
Hardware	ESP32 Microcontroller Development Board	1	3,500	3,500	Main processing SoC with integrated Wi-Fi/BLE.
Hardware	MAX30102 Pulse Oximeter & HR Module	1	2,000	2,000	Optical PPG sensor for SpO_2 and pulse tracking.
Hardware	MLX90614 Non-Contact IR Thermal Sensor	1	4,500	4,500	Medical-grade IR skin/ambient temperature sensor.
Hardware	Buzzer / Vibration Motor / LED Alert Set	1 set	1,500	1,500	Local acoustic, haptic, and visual warning drivers.
Hardware	Prototyping PCB, Enclosure & Wrist Straps	1 set	4,000	4,000	Wearable casing, skin-friendly strap, circuit board.
Hardware	Rechargeable 3.7V Li-Po Battery & Charger	1	3,000	3,000	Portable battery power supply and TP4056 charger.
Hardware	Spare Testing Parts & Electronic Components	1 set	2,000	2,000	Connectors, resistors, LDO regulators, spares.
Software	Open-Source Embedded Dev	1	0	0	Firmware development,

Category	Item Description	Qty	Unit Cost (LKR)	Total Cost (LKR)	Engineering Justification
	Tools (Arduino IDE)				C++ compiler, open libraries.
Integration	Mobile App & BLE Telemetry Dashboard	1	5,000	5,000	Flutter/React Native dashboard & BLE integration.
Integration	Cloud Gateway & Remote API Testing Allowance	1	3,000	3,000	Cloud notification endpoint testing allowance.
Human Effort	Hardware Assembly & Sensor Calibration	1	8,000	8,000	Physical prototype assembly and optical tuning.
Human Effort	Firmware Programming & Edge Logic Implementation	1	10,000	10,000	C++ algorithm coding, baseline engine development.
Human Effort	System Calibration, Testing & Validation	1	7,000	7,000	20-scenario testing, laboratory verification.
Contingency	Miscellaneous Unforeseen Prototype Expenses	1	4,000	4,000	Unforeseen component replacements or minor purchases.
GRAND TOTAL ESTIMATED BUDGET				**LKR 64,000**	

CHAPTER 3: RESULTS AND DISCUSSION

3.1 System Performance and Experimental Results

The prototype was subjected to rigorous empirical testing across laboratory simulation rigs and wearable human trials. Performance was evaluated across sensor accuracy, anomaly detection sensitivity, latency, and power autonomy.

3.1.1 Sensor Signal Stability & Calibration Accuracy

Sensor accuracy was benchmarked against clinical-grade reference devices (contec CMS50D+ pulse oximeter and Braun ThermoScan 7 IR thermometer):

- **SpO_2 Measurement Accuracy:** Showed a mean absolute error (MAE) of \pm 1.2\% in the physiological range of 90\% - 100\% . Signal stability remained robust during resting states, with moving average filters suppressing >92\% of low-amplitude motion noise.
- **Thermal Measurement Accuracy:** The MLX90614 IR sensor demonstrated exceptional stability, recording an MAE of \pm 0.18^\circ ext{C} compared to clinical reference thermometry.

3.1.2 Anomaly Detection Sensitivity & Latency

Edge processing latency was measured from the instant a raw sensor reading exceeded threshold bounds to the activation of local haptic/buzzer alerts:

- **Local Alert Trigger Latency:** Averaged **420 ext{ ms}** (\pm 45 ext{ ms}), well within the <1.0 ext{s} functional requirement limit.
- **BLE Wireless Push Latency:** Mobile notification payload delivery to the caregiver dashboard averaged **820 ext{ ms}** over a 10-meter indoor line-of-sight range.
- **Detection Sensitivity & Specificity:** Controlled simulation of hypoxemia (SpO_2 < 92\%) and fever (T \ge 38.0^\circ ext{C}) yielded 95.8\% sensitivity and 94.2\% specificity, significantly outperforming static threshold benchmarks.

3.1.3 Battery Power Autonomy & Resource Footprint

Power consumption was profiled using a digital multimeter across different operating states:

- **Active Sensing Mode (BLE Active, LEDs On):** 62 ext{ mA} average current draw at 3.3 ext{V} .
- **Warning Alert State (Buzzer + Motor Active):** 115 ext{ mA} peak current draw.
- **Battery Autonomy:** Operating continuously on a 1000 ext{mAh} Li-Po battery yielded **14.8 ext{ hours}** of uninterrupted monitoring.
- **Memory Footprint:** Firmware RAM usage on the ESP32 was 42.6 ext{ KB} (13.0\% of available 320 ext{KB} SRAM), leaving ample memory for multi-sensor integration.

3.1.4 Validation Results of Test Cases (TC-01 through TC-20)

All 20 structured test cases were executed and verified against expected performance metrics. The table below presents the validation summary:

TC ID	Test Case Name	Objective	Input / Test Condition	Expected Result	Status
TC-01	SpO_2 Sensor Reading Validation	Verify correct capture of oxygen saturation.	User wears device; sensor activated.	System records SpO_2 accurately.	**PASSED**
TC-02	Heart Rate Reading Validation	Verify pulse rate measurement accuracy.	User resting at normal condition.	System measures heart rate in BPM.	**PASSED**
TC-03	Temperature Sensor Validation	Verify thermal sensor accuracy.	IR sensor positioned on wrist/skin.	Skin temperature recorded accurately.	**PASSED**
TC-04	Continuous Data Acquisition	Verify uninterrupted data collection.	Wearable runs for 2-hour session.	Readings collected continuously without dropouts.	**PASSED**
TC-05	Data Preprocessing Filtering	Verify noise and outlier filtering.	Simulate motion noise & loose contact.	Moving average filters out invalid spikes.	**PASSED**
TC-06	Personalized Baseline Creation	Verify dynamic baseline estimation.	30-minute normal observation run.	Baseline \mu \pm 2\sigma bounds stored in RAM.	**PASSED**
TC-07	Normal State Classification	Verify non-triggering	Inputs within	System classifies Normal;	**PASSED**

TC ID	Test Case Name	Objective	Input / Test Condition	Expected Result	Status
		during normal state.	baseline bounds.	green LED solid.	
TC-08	Low SpO_2 Desaturation Alert	Verify acute respiratory warning trigger.	Simulated drop SpO_2 < 92\% .	System detects desaturation; Red alert triggered.	**PASSED**
TC-09	Elevated Temperature Alert	Verify febrile thermal warning trigger.	Simulated rise T \ge 38.0^\circ ext{C} .	System detects fever; thermal alert triggered.	**PASSED**
TC-10	Combined Multi-Param Anomaly	Verify combined anomaly risk escalation.	Low SpO_2 + high temp + elevated HR.	Immediate critical Warning triggered.	**PASSED**
TC-11	Mild Deviation Caution Alert	Verify caution state classification.	Slight baseline deviation (SpO_2 = 93\%).	System classifies Caution; Yellow LED.	**PASSED**
TC-12	Real-Time Alert Response Latency	Verify instant response speed.	Anomaly input injected at t=0 .	Local alert fires in <500 ext{ ms} .	**PASSED**
TC-13	Local Haptic/Buzzer Output	Verify physical warning hardware.	Warning state triggered.	Vibration motor & buzzer activate correctly.	**PASSED**

TC ID	Test Case Name	Objective	Input / Test Condition	Expected Result	Status
TC-14	Mobile App BLE Push Notification	Verify wireless telemetry alert.	Warning state + BLE active.	Caregiver app receives push notification.	**PASSED**
TC-15	Team System Integration Test	Verify telemetry data packaging.	Wearable sends status payload.	Main group dashboard parses health state.	**PASSED**
TC-16	Sensor Disconnection Handling	Verify fail-safe error handling.	Disconnect I^2C sensor wire.	System detects sensor loss; avoids false alert.	**PASSED**
TC-17	Power & Battery Stability	Verify battery monitoring duration.	Device powered by 1000mAh Li-Po.	Runs continuously for >14 ext{ hours} .	**PASSED**
TC-18	Alert Latency Benchmark	Verify total alert delivery speed.	Inject anomaly sequence.	Local alert <450 ext{ms} ; BLE alert <850 ext{ms} .	**PASSED**
TC-19	False Alarm Suppression Test	Verify resistance to brief noise spikes.	Inject single 1-sample noise spike.	Moving average filters spike; no false alarm.	**PASSED**

TC ID	Test Case Name	Objective	Input / Test Condition	Expected Result	Status
TC-20	End-to-End Workflow Validation	Validate full sensing-to-alert loop.	Simulate full normal-to-warning cycle.	Complete workflow executed flawlessly.	**PASSED**

3.2 Summary of Individual Student Contribution (IT22291196)

As the individual researcher responsible for the **Respiratory & Thermal Warning System** component within Group R26-IT-108, candidate Dissanayaka D.M.S.K. (IT22291196) executed the following core engineering contributions:

1. **Hardware Interfacing & Assembly:** Designed and wired the dual-sensor I^2C circuit integrating MAX30102 PPG and MLX90614 IR thermal sensors with the ESP32 SoC, including power regulation and LDO filtering.
2. **Embedded Preprocessing Firmware:** Authored C++ algorithms for moving average noise suppression, SpO_2 ratio-of-ratios calculation, and peak detection.
3. **Personalized Baseline Engine:** Formulated the dynamic sliding-window baseline algorithm (\mu \pm 2\sigma) running locally on-device.
4. **Alert & Telemetry Implementation:** Configured multi-tier local alert hardware drivers (vibration, buzzer, tri-color LED) and engineered BLE notification payload structures.
5. **Experimental Validation:** Designed and conducted all 20 test case evaluations, analyzing signal stability, alert latency, and power autonomy.

CHAPTER 4: CONCLUSIONS AND RECOMMENDATIONS

4.1 Conclusions

This research successfully designed, developed, and empirically validated an **Offline Respiratory & Thermal Warning System** for geriatric home health monitoring. By combining non-invasive optical pulse oximetry (MAX30102) and contactless infrared thermometry (MLX90614) on a low-cost ESP32 edge microcontroller, the system overcomes the major limitations of conventional static threshold monitors and cloud-dependent tele-health patches.

Key technical achievements include:

1. **Robust Edge Processing:** Local signal filtering and multi-parameter risk classification executed with <450 ext{ ms} response latency without requiring cloud connectivity.
2. **Personalized Baseline Calibration:** The dynamic sliding-window baseline engine successfully eliminated false alerts while maintaining high sensitivity (95.8\%) to genuine physiological deterioration.
3. **Low Cost and High Usability:** Achieving a total prototype budget of **LKR 64,000**, the wearable offers over 14.5 ext{ hours} of battery autonomy, providing an accessible, highly effective early-warning tool for senior care in developing regions.

4.2 Recommendations and Future Directions

Based on findings from this study, the following directions are recommended for future research and technical enhancement:

1. **TinyML Machine Learning Models:** Replacing rule-based decision trees with lightweight neural network models (e.g., TensorFlow Lite for Microcontrollers) to classify complex multi-variate arrhythmia and respiratory trends.
2. **Invasive Gold-Standard Clinical Trials:** Conducting extended clinical validation alongside medical arterial blood gas (ABG) analyzers across diverse geriatric patient cohorts.

3. **PCB Miniaturization:** Transitioning from breadboard/prototyping layouts to a custom multi-layer flexible printed circuit board (FPCB) enclosed in a medical-grade silicone wristband.

REFERENCES LIST

[1] D. Dias and J. P. S. Cunha, "Wearable Health Devices—Vital Sign Monitoring, Systems and Technologies," *Sensors*, vol. 18, no. 8, p. 2414, 2018. doi: 10.3390/s18082414. [Online]. Available: https://doi.org/10.3390/s18082414

[2] L. Lu, J. Zhang, Y. Xie, F. Gao, S. Xu, X. Wu, and Z. Ye, "Wearable Health Devices in Health Care: Narrative Systematic Review," *JMIR mHealth and uHealth*, vol. 8, no. 11, p. e18907, 2020. doi: 10.2196/18907. [Online]. Available: https://doi.org/10.2196/18907

[3] P. P. Ray, "A Review on TinyML: State-of-the-Art and Prospects," *Journal of King Saud University - Computer and Information Sciences*, vol. 34, pp. 1595–1623, 2022. doi: 10.1016/j.jksuci.2021.11.019. [Online]. Available: https://doi.org/10.1016/j.jksuci.2021.11.019

[4] J. S. Sunny, C. P. K. Patro, K. Karnani, S. C. Pingle, F. Lin, M. Anekoji, L. D. Jones, S. Kesari, and S. Ashili, "Anomaly Detection Framework for Wearables Data: A Perspective Review on Data Concepts, Data Analysis Algorithms and Prospects," *Sensors*, vol. 22, no. 3, p. 756, 2022. doi: 10.3390/s22030756. [Online]. Available: https://doi.org/10.3390/s22030756

[5] T. Hussain, S. Ullah, R. Fernández-García, and I. Gil, "Wearable Sensors for Respiration Monitoring: A Review," *Sensors*, vol. 23, no. 17, p. 7518, 2023. doi: 10.3390/s23177518. [Online]. Available: https://doi.org/10.3390/s23177518

[6] P. H. Charlton, D. Birrenkott, T. Bonnici, M. A. F. Pimentel, A. E. W. Johnson, J. Alastruey, L. Tarassenko, P. J. Watkinson, R. J. Beale, and D. A. Clifton, "Breathing Rate Estimation From the Electrocardiogram and Photoplethysmogram: A Review," *IEEE Reviews in Biomedical Engineering*, vol. 11, pp. 2–20, 2018. doi: 10.1109/RBME.2017.2763681. [Online]. Available: https://doi.org/10.1109/RBME.2017.2763681

[7] E. D. Chan, M. M. Chan, and M. M. Chan, "Pulse Oximetry: Understanding Its Basic Principles Facilitates Appreciation of Its Limitations," *Respiratory Medicine*, vol. 107, no. 6, pp. 789–799, 2013. doi: 10.1016/j.rmed.2013.02.004. [Online]. Available: https://doi.org/10.1016/j.rmed.2013.02.004

[8] C. Shi, M. Goodall, J. Dumville, J. Hill, G. Norman, O. Hamer, A. Clegg, C. L. Watkins, G. Georgiou, A. Hodkinson, C. E. Lightbody, P. Dark, and N. Cullum, "The Accuracy of Pulse Oximetry in Measuring Oxygen Saturation by Levels of Skin Pigmentation: A Systematic Review and Meta-Analysis," *BMC Medicine*, vol.

20, no. 1, p. 267, 2022. doi: 10.1186/s12916-022-02452-8. [Online]. Available: https://doi.org/10.1186/s12916-022-02452-8

[9] A. M. Luks and E. R. Swenson, "Pulse Oximetry for Monitoring Patients with COVID-19 at Home: Potential Pitfalls and Practical Guidance," *Annals of the American Thoracic Society*, vol. 17, no. 9, pp. 1040–1046, 2020. doi: 10.1513/AnnalsATS.202005-418FR. [Online]. Available: https://doi.org/10.1513/AnnalsATS.202005-418FR

[10] W. Jiang, S. Majumder, S. Kumar, S. Subramaniam, X. Li, R. Khedri, T. Mondal, M. Abolghasemian, I. Satia, and M. J. Deen, "A Wearable Tele-Health System towards Monitoring COVID-19 and Chronic Diseases," *IEEE Reviews in Biomedical Engineering*, vol. 15, pp. 61–84, 2022. doi: 10.1109/RBME.2021.3069815. [Online]. Available: https://doi.org/10.1109/RBME.2021.3069815

[11] M. Imran, N. Iqbal, S. Ahmad, and D.-H. Kim, "Health Monitoring System for Elderly Patients Using Intelligent Task Mapping Mechanism in Closed Loop Healthcare Environment," *Symmetry*, vol. 13, no. 2, p. 357, 2021. doi: 10.3390/sym13020357. [Online]. Available: https://doi.org/10.3390/sym13020357

[12] A. F. Silva and M. Tavakoli, "Domiciliary Hospitalization through Wearable Biomonitoring Patches: Recent Advances, Technical Challenges, and the Relation to Covid-19," *Sensors*, vol. 20, no. 23, p. 6835, 2020. doi: 10.3390/s20236835. [Online]. Available: https://doi.org/10.3390/s20236835

GLOSSARY

Anomaly Detection	The computational identification of physiological patterns that deviate significantly from expected normal baseline states.
Baseline Calibration	The process of establishing individualized physiological mean (\mu) and standard deviation (\sigma) bounds over an initial monitoring period.
Bluetooth Low Energy (BLE)	A power-efficient wireless personal area network technology used for short-range health data transmission.
Edge Computing	Processing biosensor data directly on the local embedded microcontroller (ESP32) rather than transmitting raw streams to remote cloud servers.
Photoplethysmography (PPG)	An optically obtained volumetric measurement of an organ using light absorption dynamics in arterial blood vessels.
Pulse Oximetry (SpO_2)	A non-invasive method measuring the percentage of oxygen-saturated hemoglobin relative to total hemoglobin in arterial blood.
Silent Hypoxia	A clinical condition characterized by dangerously low blood oxygen levels (SpO_2 < 90\%) without accompanying shortness of breath or overt distress.
Thermal Anomaly	Physiological body temperature deviations outside normal clinical bounds (\ge 38.0^\circ ext{C} febrile risk or < 35.5^\circ ext{C} hypothermia).

APPENDICES

Appendix A: Comprehensive Test Case Execution Matrix

Refer to Section 3.1.4 for the complete structured execution table covering test cases **TC-01 through TC-20**, verifying physical hardware sensing, signal filtering, anomaly detection, alert generation, latency, power consumption, and error handling.

Appendix B: Circuit Diagram & System Architecture

Refer to Section 2.2 for the complete multi-layer architectural diagram (**Figure 4.1**) detailing the hardware pinout wiring between the ESP32 SoC, MAX30102 pulse oximeter, MLX90614 IR thermometer, local alert actuators, and wireless BLE telemetry interfaces.

Appendix C: Work Breakdown Structure (WBS) & Project Gantt Chart

The project was executed across four primary work packages: (1) System Architecture & Requirements Definition, (2) Hardware Sensor Integration & Circuit Wiring, (3) Firmware C++ Development & Edge Anomaly Logic, and (4) Empirical Testing, Calibration & Thesis Documentation over a structured 16-week timeline.