

Honeywell Specialty Chemicals Seelze GmbH D-30918 Seelze

Telefon: +49 5137 999-353

28.09.2022 / COAH_01003610

Analyzed for:
Honeywell PMT Advanced Materials E&C
Offering Management
115 Tabor Road, Morris Plains
NJ 07950, USA

QC Release Date: 28.Sep.2022

Production date: 26.Sep.2022

Expiry Date: 31.Aug.2027

Article/Product: 37817

Lot: M2690

HYDRANAL-Methanol Rapid

The quality certificate is valid for the time of delivery

HYDRANAL-Methanol Rapid is a medium for accelerated volumetric one-component Karl Fischer titration to be used with HYDRANAL-Composite.

Parameter	Specification	QC results
Suitability (for accelerated KF-Titration)*	test	test passed
Assay of water*	max. 0.01 %	0.007 %
Number of samples (n)	minimum 2	4

* volumetric KF titration according to Test Method 01 HYDRANAL version 13. Nov. 2017 (including ISO 760 version 1978 and ASTM E 203 version 2016).

The KF suitability test was performed on the first bottle from the filling line.

The assay of water is the highest value from a series of measurements on (n) samples.

- The expiry date is based on the current knowledge and holds only for proper storage conditions in the originally closed flasks/ packages.
- We herewith confirm that the delivery is effected according to the technical delivery conditions agreed.
- Particular properties of the products or the suitability for a particular area of application are not assured.
- We guarantee a proper quality within our General Conditions of Sales.

Thomas Wendt

Honeywell Specialty Chemicals Seelze GmbH
Thomas Wendt

Supervisor HYDRANAL Center of Excellence
Wunstorfer Str. 40, D-30926 Seelze



Quality Control according to DIN EN ISO/IEC 17025