



Certificate of Analysis

1.88005.1000 CombiTitrant 5 one-component reagent for volumetric Karl Fischer titration 1
ml \pm ca. 5 mg H₂O Aquastar®
Batch HX29878305

	Spec. Values		Batch Values	
Efficiency	≥ 5	mg/ml	5.37	mg/ml

The concentration (efficiency) of Karl Fischer titrants decreases over time. Therefore the titer, being a corrective factor that compensates for this effect as well as variations of equipment, personal handling and laboratory conditions, must be determined directly after opening a new bottle and should be verified at regular intervals.

Date of release (DD.MM.YYYY) 08.07.2022
Minimum shelf life (DD.MM.YYYY) 31.07.2025

Ayfer Yildirim
Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.