



# Certificate of Analysis

1.00029.4000 Acetonitrile hypergrade for LC-MS LiChrosolv®  
Batch I1277629

Batch Values		
Purity (GC)	≥ 99.97	%
Identity (IR)	conforms	
Evaporation residue	≤ 1.0	mg/l
Water	≤ 0.01	%
Color	≤ 10	Hazen
Acidity	≤ 0.0001	meq/g
Alkalinity	≤ 0.0002	meq/g
Al (Aluminium)	≤ 10	ppb
Ca (Calcium)	≤ 10	ppb
Fe (Iron)	≤ 10	ppb
K (Potassium)	≤ 5	ppb
Mg (Magnesium)	≤ 10	ppb
Na (Sodium)	≤ 50	ppb
Any further metal (ICP-MS)	≤ 5	ppb
Gradient grade (at 210 nm)	< 0.1	mAU
Gradient grade (at 254 nm)	< 0.1	mAU
Fluorescence (as quinine at 254 nm)	0.1	ppb
Fluorescence (as quinine at 365 nm)	0.1	ppb
Transmission (at 191 nm)	≥ 25	%
Transmission (at 195 nm)	≥ 85	%
Transmission (at 200 nm)	≥ 96	%
Transmission (at 215 nm)	≥ 98	%
Transmission (from 230 nm)	≥ 99	%
Suitable for LC-MS; Intensity of background mass peak based on reserpine (APCI/ESI positive)	≤ 2	ppb
Suitable for LC-MS; Intensity of background mass peak based on reserpine (APCI/ESI negative)	≤ 20	ppb

Filtered by 0.2 µm filter

Date of release (DD.MM.YYYY) 22.03.2023  
Minimum shelf life (DD.MM.YYYY) 31.03.2026

Jeannette David  
Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.