Supelco_®

Certificate of Analysis - Certified Reference Material

Certipur® Buffer solution pH 7.00 (20°C)

Certified Reference Material for pH measurement

Product no.: 1.09439.1000 **Lot no.:** HC29800539

Description of CRM: Certipur® Buffer solution pH 7.00 (20°C)

Certified Reference Material for pH measurement

Expiry date: 2025/08/31

Storage: +15°C to +25°C tightly closed in the original container

Composition: di-sodium hydrogen phosphate / potassium dihydrogen phosphate

| Certified value | Associated uncertainty, <i>U=k·u</i> (<i>k</i> =2) | |
|-----------------|--|--|
| pH value 7.00 | ±0.02 (20°C) | |

Metrological traceability: The pH value of this certified buffer solution is directly traceable to primary certi-

fied reference materials characterised by PTB and verified by SRMs from NIST.

NIST 189c, 188, 185i, 186 Ig, 186 IIg, 187f

PTB OX-530/22, TA-442/19, PHT-467/20, PHO-490/20, BO-468/20 PTB: Physikalisch Technische Bundesanstalt, Braunschweig, Germany NIST: National Institute of Standards and Technology, Gaithersburg, USA.

Measurement method: pH value is measured with a combined glass electrode after 5-point calibration

according to DIN 19268 with reference buffer solutions according to DIN 19266,

IUPAC, NIST, Ph.Eur. and USP.

Accreditation: Merck KGaA, Darmstadt, Germany is accredited by the German accreditation

authority DAkkS as registered reference material producer D-RM-15185-01-00 in

accordance with ISO 17034.

Certificate issue date: 2022/08/29

DAKKS

Deutsche
Akkreditierungsstelle
D-RM-15185-01-00

CRM released by Approving Officer or delegate of Quality Control

Dipl.-Ing. Ayfer Yildirim Responsible Laboratory Manager

A. Yildirim

Merck KGaA, 64271 Darmstadt, Germany, Tel. +49(0)6151 72-2440 EMD Millipore Corporation, 400 Summit Drive, Burlington MA 01803, USA, Tel. +1-978-715-4321 Sigma-Aldrich Canada Co. or Millipore (Canada) Ltd. 2149 Winston Park, Dr. Oakville, Ontario, L6H 6J8, Phone: +1 800-565-1400



Intended use: This reference material is intended for use as a calibration standard for pH instru-

ments or pH electrodes or as a control sample for measuring the pH value.

Instructions for handling

and correct use:

Health and safety

information: **Preparation:** The pH value is strongly dependent on the temperature. It is therefore necessary

to keep the temperature constant within the measurement.

Please refer to the Safety Data Sheet for detailed information about the nature of any hazard and appropriate precautions to be taken.

This reference material is prepared gravimetrically from di-sodium hydrogen

phosphate, potassium dihydrogen phosphate and high purity water.

Associated uncertainty:

The expanded uncertainty U_{CRM} is calculated as $U_{CRM} = k \cdot u_{CRM}$, where k=2 is the coverage factor for a 95% coverage probability and u_{CRM} is the combined standard uncertainty in accordance to ISO 17034.

The combined uncertainty u_{CRM} is derived from combination of the squared uncertainty contributions:

$$\mathbf{u}_{CRM} = \sqrt{\mathbf{u}^2 Characterisation} + \mathbf{u}^2 Homogeneity + \mathbf{u}^2 Stability$$

is the uncertainty in accordance with DIN EN ISO/IEC 17025 which includes the Ucharacterisation:

contributions of the primary reference material and the measuring system. The characterisation measurements have been conducted by our DAkkS accredited

calibration laboratory (D-K-15185-01).

is the between-bottle variation in accordance with ISO 17034. The assessment of Uhomogeneity:

homogeneity is performed by analysis of a representative number of systematically

chosen sample units.

is the uncertainty obtained from short-term and long-term stability in accordance Ustability:

with ISO 17034. The stability studies are the basis for the quantification of the

expiry date of this reference material for the unopened bottle.

Informative values:

| Temperature dependence ¹ : | Temperature [°C] | ΔpH |
|---------------------------------------|------------------|--------|
| | 0 | + 0.13 |
| | 5 | + 0.07 |
| | 10 | + 0.05 |
| | 15 | + 0.02 |
| | 20 | ± 0 |
| | 25 | - 0.02 |
| | 30 | - 0.02 |
| | 35 | - 0.04 |
| | 40 | - 0.05 |
| | 50 | - 0.05 |

¹Temperature deviation data provided for reference only. Values are not batch-specific and should not be considered certified values.

For more detailed information please read the certification report on our website.

Certificate of analysis revision history:

| Certificate version | Date | Reason for version |
|---------------------|------------|--------------------|
| 01 | 2022/08/29 | Initial version |

The vibrant M, Supelco, Certipur and Sigma-Aldrich are trademarks of Merck KGaA, Darmstadt, Germany or its affiliates.

Detailed information on trademarks is available via publicly accessible resources.

© 2022 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved.

The life science business of Merck KGaA, Darmstadt, Germany operates as MilliporeSigma in the U.S. and Canada.

