



Certificate of Analysis

1.08101.2500 Tetrahydrofuran for liquid chromatography LiChrosolv®
Batch 11272101

Batch Values		
Purity (GC)	≥ 99.9	%
Identity (IR)	conforms	
Evaporation residue	≤ 1.0	mg/l
Water	≤ 0.02	%
Acidity	≤ 0.0002	meq/g
Alkalinity	≤ 0.0002	meq/g
Peroxides	≤ 0.02	%
Transmission (at 218 nm)	≥ 30	%
Transmission (at 230 nm)	≥ 35	%
Transmission (at 250 nm)	≥ 65	%
Transmission (at 280 nm)	≥ 95	%

Filtred by 0.2µm filter

Date of release (DD.MM.YYYY) 13.02.2023
Minimum shelf life (DD.MM.YYYY) 28.02.2026

Jeannette David
Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.