

Package #135

Integrate acoustic emission sensors along the infusion tubing that analyze pressure wave signatures characteristic of occlusion events

Technology Maturity: ENGINEERING_DEFINITION

Transfer Posture: Sponsored Validation

Problem: Documented recurring problem 'downstream occlusion detection failure' for infusion pumps: 100 MAUDE adverse-event records retrieved live at benchmark time (2007–2032). Event narratives: 'custody in retrieval log'. Literature context: 10 EuropePMC records on infusion-pump occlusion alarms. MAUDE LIMITS STAMPED INLINE: incidence UNKNOWN, causation UNVERIFIED, voluntary reports (Art. XXI.5).

Mechanism: Industrial fluid dynamics monitoring system using pressure wave pattern recognition

Contents:

- 01 - Executive Technology Brief
- 02 - Engineering Technology-Transfer Dossier
- 03 - Buyer Decision Card
- 04 - Evidence Summary
- 05 - Transfer Manifest

Status: ENGINEERING_DEFINITION. Not physically validated.