

Technology #118

Develop coronary stent with adaptive geometry
that dynamically adjusts to vessel anatomy to
minimize mechanical conflict

Engineering Technology-Transfer Dossier

CONFIDENTIAL

Technology Maturity: ENGINEERING_DEFINITION

Generated: 2026-08-30

Not physically validated. Not transfer-ready.

BUYER DECISION PAGE

WHAT IS IT?

Develop coronary stent with adaptive geometry that dynamically adjusts to vessel anatomy to minimize mechanical conflict

WHY DOES IT MATTER?

Documented recurring problem 'unknown__for_use_when_the_device_problem' for coronary stent: 14 MAUDE reports across 3 distinct devices (20080214 to 20080220); root-cause evidence from 1 recall records and 10 event narratives (custody attached in the campaign report). MAUDE limitations: incidence UNKNOWN, causation UNVERIFIED (Art. XXI.5). Existing solutions fail because: Why existing solutions fail, per retrieved evidence: 1 recall root-cause descriptions (FDA enforcement records, custody attached); 10 MAUDE event narratives from this mechanism cluster (mdr_text, custody attached). Statements are the providers' own words, cust Unresolved gap: Across the 2 relevant retrieved attempts (a bounded, non-exhaustive set), none measurably addresses: conditions, devices, events, failure, function, injury, intervention, introducing, known, maintain, matching, modes, must, producing, recurring, reported, unde

WHAT IS ACTUALLY ESTABLISHED?

Engineering definition with 3 governing equations, 12 design inputs, 6 failure modes. Computational model exists. No physical prototype.

WHAT IS NOT ESTABLISHED?

Physical validation, manufacturing process, regulatory pathway, IP ownership.

WHY MIGHT A BUYER CARE?

Opportunity in mechanical_structural. 10-step build plan with specific test articles and acceptance criteria. Target buyer: domain-appropriate technology buyer (to be identified by portfolio owner).

WHAT DOES THE BUYER GET?

Buyer receives: canonical invention specification (with epistemic classes).; Buyer receives: engineering specification (this document, all blocks classed).; Buyer receives: design input/output graph with explicit IDs.; Buyer receives: engineering reasoning chains (claim->principle->model->input->assumption->output->failure->verification).; Buyer receives: killer-experiment definition and its EIG basis.; Buyer receives: evidence custody references and hashes.; Buyer receives: unresolved-uncertainty register.; Buyer receives: domain equation set (symbolic; sources and applicability judgments identified).

WHAT DOES THE BUYER HAVE TO BUILD?

Buyer must create: actual geometry and detailed design.; Buyer must create: sourced engineering parameters and acceptance thresholds.; Buyer must create: physical prototypes.; Buyer must create: all verification and validation results.; Buyer must create: manufacturing process qualification.; Buyer must create: regulatory pathway determination and submissions.

WHAT IS THE NEXT DECISIVE EXPERIMENT?

first physical article implementing the proposed intervention (GEOMETRY ABSENT — must be designed)

WHAT WOULD MAKE US KILL IT?

killer experiment fails: falsification_test_from_candidate (decision scores: EIG 0.44573539382492045, EIG-per-cost 0.44573539382492045; the falsification test carries the pre-registered pass/fail before any transfer)

WHAT TRANSACTION COULD MAKE SENSE?

Sponsored Validation

What Exists Today

Existing:

- + Documented mechanism with governing equations
- + Computational model
- + 1 external evidence sources
- + Design inputs and outputs
- + Failure analysis
- + 10-step engineering build plan
- + Transfer boundary definition

Does Not Yet Exist:

- Physical prototype
- Qualified manufacturing process
- Clinical data
- Verified IP ownership
- Regulatory clearance
- Commercial product

Buyer Can Commission:

- > Specific bench experiment (see build plan)
- > Prototype fabrication
- > Materials study
- > Manufacturing feasibility study
- > Regulatory pre-submission

Development / De-risking Ladder

CURRENT STATE: ENGINEERING_DEFINITION

NEXT ENGINEERING STEP: first physical article implementing the proposed intervention (GEOMETRY ABSENT — must be designed)

NEXT VALIDATION STEP: Bench prototype validation per build plan

NEXT REGULATORY STEP: Pre-submission to FDA for pathway classification

COMMERCIALIZATION GATE: Clinical validation evidence required

TRANSFER GATE: Transfer-ready when physical validation + regulatory pathway + IP verified

Engineering Investment Ladder

Next \$5K action: WP-01 - first physical article implementing the proposed intervention (GEOMETRY ABSENT — must be designed) - NOT ESTABLISHED (to be quoted)

Basis: Build plan WP. Scope: physical bench/clinical experiment — NOT simulation. Uncertainty: Effort estimate, not quoted.

Next \$25K action: WP-03 - article for VF-001 - NOT ESTABLISHED (to be quoted)

Basis: Build plan WP. Scope: Computational modeling of adaptive stent deployment in various coronary anatomies, measuring mechanical stress points and potential for v.... Uncertainty: Effort estimate, not quoted.

Next \$100K action: WP-05 - article for VF-003 - NOT ESTABLISHED (to be quoted)

Basis: Build plan WP. Scope: buckling / kink margin tests across duty cycle. Uncertainty: Effort estimate, not quoted.

Top 5 Buyer Diligence Questions

Q1: Is the mechanism physically credible for the intended application?

The governing model includes 3 equations. Key assumptions: stress concentrators multiply local stress above the nominal value — geometry factors must be applied from the actual design; superelastic members (e.g. nitinol) do not follow linear elasticity through transformation — use their measured plateau relations instead; K for catheter constraint conditions is MODEL_DERIVED; MEC-002 (Section bending stress) is applicable only under: linear-elastic material; small deflection; prismatic section.; MEC-003 (Hooke's law (linear elasticity)) is applicable only under: below proportional limit; isothermal.; MEC-001 (Euler buckling load) is applicable only under: elastic, slender column, small deflection.; Model input variables requiring sourced values before any numeric result: M (N.m), c (m), I (m⁴), E (Pa), epsilon (-), E (Pa), I (m⁴), K (-), L (m) — all UNKNOWN (no sourced value exists; Art. VIII).; Model output variables produced by the governing model: sigma (Pa), sigma (Pa), P_cr (N) — UNKNOWN until the inputs are sourced and the model is evaluated.; Critical parameters under specification (parameters with recorded units and epistemic status; no values invented): member slenderness ratio [NOT_ASSIGNED, UNKNOWN] value_status=UNKNOWN; yield/ultimate strength [NOT_ASSIGNED, UNKNOWN] value_status=UNKNOWN; applied cyclic load amplitude [P_cr, N] value_status=UNKNOWN; required run-out cycles [NOT_ASSIGNED, UNKNOWN] value_status=UNKNOWN; articulation radius [NOT_ASSIGNED, UNKNOWN] value_status=UNKNOWN; friction coefficient at guides [NOT_ASSIGNED, UNKNOWN] value_status=UNKNOWN.. UNKNOWN: whether assumptions hold in vivo.

Q2: What evidence supports the core claim?

1 external evidence sources. All are external precedent, not physical validation of this specific invention.

Q3: What is the regulatory pathway?

UNKNOWN. Not yet determined. Requires pre-submission to FDA. No predicate identified for some packages.

Q4: Who owns the IP?

UNKNOWN. IP ownership not verified. Requires legal counsel engagement.

Q5: What would it cost to validate?

First experiment: first physical article implementing the proposed intervention (GEOMETRY ABSENT — must be designed). Estimated effort from build plan. Full validation cost: COST TO BE QUOTED.

Commercial Interest

TARGET APPLICATION: CSF shunt systems for hydrocephalus management

CURRENT ALTERNATIVE: Standard fixed-pressure or programmable CSF shunt valves

MECHANISM DIFFERENCE: Lead conflict and lifetime consequences differ by TTVI modality

POTENTIAL VALUE DRIVER: Addresses: Documented recurring problem 'unknown__for_use_when_the_device_problem' for coronary stent: 14 MAUDE reports across 3 distinct devices (20080214 to 20080220); root-cause evidence from 1 recall records and 10 event narratives (custody attached in the campaign report). MAUDE limitations: incidence UNKNOWN, causation UNVERIFIED (Art. XXI.5). Existing solutions fail because: Why existing solutions fail, per retrieved evidence: 1 recall root-cause descriptions (FDA enforcement records, custody attached); 10 MAUDE event narratives from this mechanism cluster (mdr_text, custody attached). Statements are the providers' own words, cust Unresolved gap: Across the 2 relevant retrieved attempts (a bounded, non-exhaustive set), none measurably addresses: conditions, devices, events, failure, function, injury, intervention, introducing, known, maintain, matching, modes, must, producing, recurring, reported, unde

ADOPTION BARRIER: Requires physical validation, regulatory clearance, and manufacturing qualification

DEVELOPMENT BARRIER: No physical prototype exists; manufacturing process not established

Licensee Capability Fit

IDEAL BUYER: domain-appropriate technology buyer (to be identified by portfolio owner)

REQUIRED ENGINEERING CAPABILITY: Catheter design, fluid mechanics / domain-specific engineering

REQUIRED MANUFACTURING CAPABILITY: Medical tubing extrusion, injection molding, cleanroom assembly

REQUIRED REGULATORY CAPABILITY: FDA 510(k) or PMA submission experience

REQUIRED CAPITAL: Validation: \$50-200K. Full development: \$500K-2M (estimates, not quoted)

REQUIRED MARKET ACCESS: Neurosurgery market channels; hospital purchasing relationships

1. Technology Description

Domain: mechanical_structural

Disciplines: mechanical engineering, structural analysis, fatigue and fracture, materials science

Architecture: PROPOSED architecture (no physical system exists): load path / structural backbone; actuation or compliance element; guidance / bearing interfaces; fatigue-critical features register — realizing: Develop coronary stent with adaptive geometry that dynamically adjusts to vessel anatomy to minimize mechanical conflict

Subsystems:

?: load path / structural backbone - ?

?: actuation or compliance element - ?

?: guidance / bearing interfaces - ?

?: fatigue-critical features register - ?

2. Mechanism Architecture

Physical Changes: PROPOSED: Develop coronary stent with adaptive geometry that dynamically adjusts to vessel anatomy to minimize mechanical conflict

Key Physics: Mechanics / structures governing relations proposed for this invention; SYMBOLIC ONLY until inputs are sourced

3. Governing Engineering Model

Summary: Mechanics / structures governing relations proposed for this invention; SYMBOLIC ONLY until inputs are sourced

Equations:

MEC-002: Section bending stress. Expression: $\sigma = M * c / I$. Applicability to this invention: CONDITIONAL; holds under: linear-elastic material; small deflection; prismatic section. Assumptions: stress concentrators multiply local stress above the nominal value — geometry factors must be applied from the actual design. Engaged variables: NOT ESTABLISHED.

MEC-003: Hooke's law (linear elasticity). Expression: $\sigma = E * \epsilon$. Applicability to this invention: CONDITIONAL; holds under: below proportional limit; isothermal. Assumptions: superelastic members (e.g. nitinol) do not follow linear elasticity through transformation — use their measured plateau relations instead. Engaged variables: NOT ESTABLISHED.

MEC-001: Euler buckling load. Expression: $P_{cr} = \pi^2 * E * I / (K * L)^2$. Applicability to this invention: CONDITIONAL; holds under: elastic, slender column, small deflection. Assumptions: K for catheter constraint conditions is MODEL_DERIVED. Engaged variables: NOT ESTABLISHED.

Assumptions:

- stress concentrators multiply local stress above the nominal value — geometry factors must be applied from the actual design
- superelastic members (e.g. nitinol) do not follow linear elasticity through transformation — use their measured plateau relations instead
- K for catheter constraint conditions is MODEL_DERIVED
- MEC-002 (Section bending stress) is applicable only under: linear-elastic material; small deflection; prismatic section.
- MEC-003 (Hooke's law (linear elasticity)) is applicable only under: below proportional limit; isothermal.
- MEC-001 (Euler buckling load) is applicable only under: elastic, slender column, small deflection.
- Model input variables requiring sourced values before any numeric result: M (N.m), c (m), I (m⁴), E (Pa), epsilon (-), E (Pa), I (m⁴), K (-), L (m) — all UNKNOWN (no sourced value exists; Art. VIII).
- Model output variables produced by the governing model: sigma (Pa), sigma (Pa), P_cr (N) — UNKNOWN until the inputs are sourced and the model is evaluated.
- Critical parameters under specification (parameters with recorded units and epistemic status; no values invented): member slenderness ratio [NOT_ASSIGNED, UNKNOWN] value_status=UNKNOWN; yield/ultimate strength [NOT_ASSIGNED,

UNKNOWN] value_status=UNKNOWN; applied cyclic load amplitude [P_cr, N] value_status=UNKNOWN; required run-out cycles [NOT_ASSIGNED, UNKNOWN] value_status=UNKNOWN; articulation radius [NOT_ASSIGNED, UNKNOWN] value_status=UNKNOWN; friction coefficient at guides [NOT_ASSIGNED, UNKNOWN] value_status=UNKNOWN.

Boundary Conditions:

- to be established by the first bench characterization (not yet performed)

Failure Regimes:

- {'regime': 'unknown for use when the device problem', 'basis': 'SOURCE_FACT (problem.json)'}
- {'regime': 'kink/buckling below required push ability', 'basis': 'domain registry candidate from the engineering domain registry — applicability to THIS invention recorded below'}
- {'regime': 'fatigue crack initiation at stress concentrator', 'basis': 'domain registry candidate from the engineering domain registry — applicability to THIS invention recorded below'}
- {'regime': 'wear debris generation at sliding interfaces', 'basis': 'domain registry candidate from the engineering domain registry — applicability to THIS invention recorded below'}
- {'regime': 'joint separation under overload', 'basis': 'domain registry candidate from the engineering domain registry — applicability to THIS invention recorded below'}
- {'regime': 'set-release of superelastic shape after set', 'basis': 'domain registry candidate from the engineering domain registry — applicability to THIS invention recorded below'}

4. Design Inputs

ID	Input	Value	Origin
DI-001	Stated constraint	Devices matching 'coronary stent' must maintain function under the conditions of use	Source: Engineering
DI-002	Mechanism driver (proposed)	Lead conflict and lifetime consequences differ by TTVI modality	MODELLED
DI-003	Expected effect target	Reduced stent-vessel wall interactions that cause recurring injury events with mechanical loading	MODELLED
DI-004	Domain design input pattern: delivery path applicability and ED via tokens ['geometry']	NOT_ESTABLISHED	ENGINEERING_PROPOSED
DI-005	Domain design input pattern: activation for applicability: ALOS ESTABLISHED (candidate pattern; confirm against evidence)	NOT_ESTABLISHED	ENGINEERING_PROPOSED
DI-006	Domain design input pattern: duty cycle and applicability: NOT_ESTABLISHED (candidate pattern; confirm against evidence)	NOT_ESTABLISHED	ENGINEERING_PROPOSED
DI-007	Regulatory design input: compliance candidate standards: ASTM F2516 (training) pathway UNKNOWN (never inferred)	NOT_ESTABLISHED	ENGINEERING_PROPOSED
DI-008	Custodied observation: Bioinspired Coronary Stents: https://www.sciencedirect.com/science/article/pii/S000368191600437b	Source: https://www.sciencedirect.com/science/article/pii/S000368191600437b	Source: https://www.sciencedirect.com/science/article/pii/S000368191600437b
DI-009	Custodied observation: Clinical experience: https://www.sciencedirect.com/science/article/pii/S000368191600437b	Source: https://www.sciencedirect.com/science/article/pii/S000368191600437b	Source: https://www.sciencedirect.com/science/article/pii/S000368191600437b
DI-010	Custodied observation: Acute Coronary Syndrome: https://www.sciencedirect.com/science/article/pii/S000368191600437b	Source: https://www.sciencedirect.com/science/article/pii/S000368191600437b	Source: https://www.sciencedirect.com/science/article/pii/S000368191600437b
DI-011	Custodied observation: The Two-Device Approach: https://www.sciencedirect.com/science/article/pii/S000368191600437b	Source: https://www.sciencedirect.com/science/article/pii/S000368191600437b	Source: https://www.sciencedirect.com/science/article/pii/S000368191600437b
DI-012	Custodied observation: Infective Endocarditis: https://www.sciencedirect.com/science/article/pii/S000368191600437b	Source: https://www.sciencedirect.com/science/article/pii/S000368191600437b	Source: https://www.sciencedirect.com/science/article/pii/S000368191600437b

5. Design Outputs

ID	Description	Status	Missing Inputs
DO-001	load path / structural backbone realizing Develop coronar	CONCEPTUAL adaptive geometry that dynamically adjusts	to vessel anatomy to minimize mechanical stresses
DO-002	actuation or compliance element realizing Develop coronar	CONCEPTUAL adaptive geometry that dynamically adjusts	to vessel anatomy to minimize mechanical stresses
DO-003	guidance / bearing interfaces realizing Develop coronar	CONCEPTUAL Adaptive geometry that dynamically adjusts	to vessel anatomy to minimize mechanical stresses
DO-004	fatigue-critical features register realizing Develop coronar	CONCEPTUAL adaptive geometry that dynamically adjusts	to vessel anatomy to minimize mechanical stresses
DO-005	load path / structural backbone interfaces with actuation	CONCEPTUAL element: signal/energy/quantification (diffusion/development)	Binary stent with adaptive geometry
DO-006	actuation or compliance element interfaces with guidance	CONCEPTUAL Interfaces: signal/energy/quantification (diffusion/development)	Binary stent with adaptive geometry
DO-007	guidance / bearing interfaces interfaces with fatigue-critical	CONCEPTUAL Register: signal/energy/quantification (diffusion/development)	Binary stent with adaptive geometry
DO-008	Device boundaries: patient/fluid inlet and outlet termination	CONCEPTUAL control/data processing standard by development; seal definitions	with adaptive geometry that dynamically adjusts
DO-009	Flow/force/signal path geometry realizing the mechanism	UNKNOWN coronary stent with adaptive geometry that dynamically adjusts	to vessel anatomy to minimize mechanical stresses
DO-010	Operating range for member slenderness ratio within which	UNKNOWN Mechanism holds its stated behavior	computed bounds for member slenderness ratio; verification of range
DO-011	Operating range for yield/ultimate strength within which	UNKNOWN Mechanism holds its stated behavior	computed bounds for yield/ultimate strength; verification of range
DO-012	Operating range for applied cyclic load amplitude within which	UNKNOWN Mechanism holds its stated behavior	computed bounds for applied cyclic load amplitude; verification of range
DO-013	Operating range for required run-out cycles within which	UNKNOWN Mechanism holds its stated behavior	computed bounds for required run-out cycles; verification of range
DO-014	Operating range for articulation radius within which the	UNKNOWN Mechanism holds its stated behavior	computed bounds for articulation radius; verification of range
DO-015	Operating range for friction coefficient at guides within which	UNKNOWN Mechanism holds its stated behavior	computed bounds for friction coefficient at guides; verification of range
DO-016	Sensing -> decision -> actuation loop maintaining the mechanism	REDESIGNED operating position development; selection; decision; adaptation; logic (structure); dynamic response	
DO-017	Information path: sensing element -> decision/monitor	CONCEPTUAL Act (minimal); rich data paths are NOT APPLICABLE to this mechanical/physical interaction	
DO-018	Fixture enabling: buckling / kink margin tests across duty cycle	UNKNOWN Rig concept named, build specific instrumentation selection; acceptance instrumentation	
DO-019	Fixture enabling: fatigue life (run-out) testing — rig construction	UNKNOWN, build specific instrumentation selection; acceptance instrumentation	
DO-020	Fixture enabling: friction and wear characterization — rig construction	UNKNOWN named, build specific instrumentation selection; acceptance instrumentation	

ID	Description	Status	Missing Inputs
DO-021	Fixture enabling: buckling/kink margin test across duty cycle	UNKNOWN	UNKNOWN

6. Critical Design Parameters

Name	Value	Unit	Basis	Verification
member slenderness ratio	UNKNOWN (no sourced value)	UNKNOWN	ENGINEERING_PROPOSED	(design assigned candidate) value and verification method
yield/ultimate strength	UNKNOWN (no sourced value)	UNKNOWN	ENGINEERING_PROPOSED	(design assigned candidate) value and verification method
applied cyclic load amplitude	UNKNOWN (no sourced value)	N	ENGINEERING_PROPOSED	(design assigned candidate) value and verification method
required run-out cycles	UNKNOWN (no sourced value)	UNKNOWN	ENGINEERING_PROPOSED	(design assigned candidate) value and verification method
articulation radius	UNKNOWN (no sourced value)	UNKNOWN	ENGINEERING_PROPOSED	(design assigned candidate) value and verification method
friction coefficient at guides	UNKNOWN (no sourced value)	UNKNOWN	ENGINEERING_PROPOSED	(design assigned candidate) value and verification method

7. Failure Modes

Mode	Mechanism	Mitigation	Verification	Residual Uncertainty
unknown for use when the device is not in use	Failure to provide for use when the device is not in use	design device NOT in use when device is not in use	ENGINEERING_PROPOSED	ENGINEERING_PROPOSED
kink/buckling below required	Failure by kink/buckling below required	design device NOT in use when device is not in use	ENGINEERING_PROPOSED	ENGINEERING_PROPOSED
fatigue crack initiation	Failure by fatigue crack initiation	design device NOT in use when device is not in use	ENGINEERING_PROPOSED	ENGINEERING_PROPOSED
wear debris generation	Failure by wear debris generation	design device NOT in use when device is not in use	ENGINEERING_PROPOSED	ENGINEERING_PROPOSED
joint separation under force	Failure by joint separation under force	design device NOT in use when device is not in use	ENGINEERING_PROPOSED	ENGINEERING_PROPOSED
set-release of superelastic	Failure by set-release of superelastic	design device NOT in use when device is not in use	ENGINEERING_PROPOSED	ENGINEERING_PROPOSED

8. Failure Analysis

Failure Mode	Mechanism	Mitigation	Residual Uncertainty
unknown for use when the device is unknown for use when the device is	Failure problem for use when the device is	design problem for use when the device is	residual uncertainty for use when the device is
kink/buckling below required	Failure problem for use when the device is	design problem for use when the device is	residual uncertainty for use when the device is
fatigue crack initiation at stress	Failure problem for use when the device is	design problem for use when the device is	residual uncertainty for use when the device is
wear debris generation at sliding	Failure problem for use when the device is	design problem for use when the device is	residual uncertainty for use when the device is
joint separation under overload	Failure problem for use when the device is	design problem for use when the device is	residual uncertainty for use when the device is
set-release of superelastic shape	Failure problem for use when the device is	design problem for use when the device is	residual uncertainty for use when the device is

9. Verification Strategy

[illegible]

10. Validation Strategy

ID	Requirement	Method	Acceptance	Result
VA-001	end-use validation of verified design	NOT ESTABLISHED (defined after verification passes)	criteria to be fixed at NOT_DONE	NOT_DONE (no definition; validation is

11. Materials

Component	Candidate Material	Source	Verification Required	Status
domain candidate material	nitinol / polymer composite membrane	minimally invasive device practice	True	ENGINEERING_PROPOSED

12. Bill of Materials

Item	Description	Qty	Type	Material	Supplier	Criticality	Verification
nitinol / polymer composite membrane	nitinol / polymer composite membrane	1	Material	nitinol / polymer composite membrane	NOT ESTABLISHED	NOT ESTABLISHED	Reliability verification required

13. Manufacturing

Status: NO MANUFACTURING ANALYSIS POSSIBLE — no design exists; processes are candidates only

Candidate Processes:

- laser-cut hypotube / braided composite shaft: domain registry manufacturing route - candidate only; no process qualification exists
- passivation and surface finish control: domain registry manufacturing route - candidate only; no process qualification exists

14. External Evidence

Source 1:

URL :

Excerpt:

15. Transfer Boundary

YOU RECEIVE:

- + Buyer receives: canonical invention specification (with epistemic classes).
- + Buyer receives: engineering specification (this document, all blocks classed).
- + Buyer receives: design input/output graph with explicit IDs.
- + Buyer receives: engineering reasoning chains (claim->principle->model->input->assumption->output->failure->verification).
- + Buyer receives: killer-experiment definition and its EIG basis.
- + Buyer receives: evidence custody references and hashes.
- + Buyer receives: unresolved-uncertainty register.
- + Buyer receives: domain equation set (symbolic; sources and applicability judgments identified).

YOU MUST DEVELOP:

- > Buyer must create: actual geometry and detailed design.
- > Buyer must create: sourced engineering parameters and acceptance thresholds.
- > Buyer must create: physical prototypes.
- > Buyer must create: all verification and validation results.
- > Buyer must create: manufacturing process qualification.
- > Buyer must create: regulatory pathway determination and submissions.

YOU MUST VERIFY:

- ? Biocompatibility (ISO 10993)
- ? Sterilization compatibility
- ? Mechanical integrity
- ? Regulatory pathway
- ? IP ownership and freedom to operate
- ? Manufacturing process capability

16. Open Questions

- {'unknown': 'FM_WITHOUT_QUANTITY_MATCHED_VERIFICATION:FM-DOM-005', 'reason': 'no available verification method measures a quantity family this failure affects (E21-A quantity rule); UNKNOWN disclosure recorded on the row'}
- {'unknown': 'value of critical parameter member slenderness ratio [NOT_ASSIGNED UNKNOWN]', 'reason': 'no sourced value exists (value_status UNKNOWN); required before any numeric design decision', 'binding': 'CP-001'}
- {'unknown': 'value of critical parameter yield/ultimate strength [NOT_ASSIGNED UNKNOWN]', 'reason': 'no sourced value exists (value_status UNKNOWN); required before any numeric design decision', 'binding': 'CP-002'}
- {'unknown': 'value of critical parameter applied cyclic load amplitude [P_cr N]', 'reason': 'no sourced value exists (value_status UNKNOWN); required before any numeric design decision', 'binding': 'CP-003'}
- {'unknown': 'value of critical parameter required run-out cycles [NOT_ASSIGNED UNKNOWN]', 'reason': 'no sourced value exists (value_status UNKNOWN); required before any numeric design decision', 'binding': 'CP-004'}
- {'unknown': 'value of critical parameter articulation radius [NOT_ASSIGNED UNKNOWN]', 'reason': 'no sourced value exists (value_status UNKNOWN); required before any numeric design decision', 'binding': 'CP-005'}
- {'unknown': 'value of critical parameter friction coefficient at guides [NOT_ASSIGNED UNKNOWN]', 'reason': 'no sourced value exists (value_status UNKNOWN); required before any numeric design decision', 'binding': 'CP-006'}
- {'unknown': 'acceptance criterion for verification VF-001', 'reason': 'criterion status ENGINEERING_PROPOSED: no sourced threshold exists; the numeric margin is fixed at pre-registration from the measured baseline', 'binding': 'VF-001'}
- {'unknown': 'acceptance criterion for verification VF-002', 'reason': 'criterion status ENGINEERING_PROPOSED: no sourced threshold exists; the numeric margin is fixed at pre-registration from the measured baseline', 'binding': 'VF-002'}

- {'unknown': 'acceptance criterion for verification VF-003', 'reason': 'criterion status ENGINEERING_PROPOSED: no sourced threshold exists; the numeric margin is fixed at pre-registration from the measured baseline', 'binding': 'VF-003'}
- {'unknown': 'acceptance criterion for verification VF-004', 'reason': 'criterion status ENGINEERING_PROPOSED: no sourced threshold exists; the numeric margin is fixed at pre-registration from the measured baseline', 'binding': 'VF-004'}
- {'unknown': 'acceptance criterion for verification VF-005', 'reason': 'criterion status ENGINEERING_PROPOSED: no sourced threshold exists; the numeric margin is fixed at pre-registration from the measured baseline', 'binding': 'VF-005'}
- {'unknown': 'acceptance criterion for verification VF-006', 'reason': 'criterion status ENGINEERING_PROPOSED: no sourced threshold exists; the numeric margin is fixed at pre-registration from the measured baseline', 'binding': 'VF-006'}
- {'unknown': 'acceptance criterion for verification VF-007', 'reason': 'criterion status ENGINEERING_PROPOSED: no sourced threshold exists; the numeric margin is fixed at pre-registration from the measured baseline', 'binding': 'VF-007'}
- {'unknown': 'acceptance criterion for verification VF-008', 'reason': 'criterion status ENGINEERING_PROPOSED: no sourced threshold exists; the numeric margin is fixed at pre-registration from the measured baseline', 'binding': 'VF-008'}

17. Decisive Next Experiment

WP	Test Article	Equipment	Measurement	Acceptance	Deliverable	Effort
WP-01	first physical article	maintaining the proposed envelope	constraint-compliance	NOT ESTABLISHED	constraint-compliance	NOT ESTABLISHED
WP-02	constraint-compliance envelope article	per method	Computational modeling of buckling / kink margin	NOT ESTABLISHED	constraint-compliance	NOT ESTABLISHED
WP-03	article for VF-001	per method	run killer experiment: buckling / kink margin	NOT ESTABLISHED	constraint-compliance	NOT ESTABLISHED
WP-04	article for VF-002	per method	fatigue life (run-out) test	NOT ESTABLISHED	constraint-compliance	NOT ESTABLISHED
WP-05	article for VF-003	per method	buckling / kink margin	NOT ESTABLISHED	constraint-compliance	NOT ESTABLISHED
WP-06	article for VF-004	per method	fatigue life (run-out) test	NOT ESTABLISHED	constraint-compliance	NOT ESTABLISHED
WP-07	article for VF-005	per method	buckling / kink margin	NOT ESTABLISHED	constraint-compliance	NOT ESTABLISHED
WP-08	article for VF-006	per method	fatigue run-out at worst stress	NOT ESTABLISHED	constraint-compliance	NOT ESTABLISHED
WP-09	article for VF-007	per method	buckling / kink margin	NOT ESTABLISHED	constraint-compliance	NOT ESTABLISHED
WP-10	article for VF-008	per method	overload margin demonstration	NOT ESTABLISHED	constraint-compliance	NOT ESTABLISHED

18. Buyer Decision Framework

WHY INVEST: Engineering definition exists with 3 governing equations and a 10-step build plan. The next experiment is clearly specified. Target buyer: domain-appropriate technology buyer (to be identified by portfolio owner).

WHY WAIT: No physical prototype exists. Regulatory pathway is unresolved. IP ownership is unverified. Manufacturing process is not qualified.

WHY REJECT: killer experiment fails: falsification_test_from_candidate (decision scores: EIG 0.44573539382492045, EIG-per-cost 0.44573539382492045; the falsification test carries the pre-registered pass/fail before any transfer)

WHAT WOULD CHANGE THE DECISION: A successful bench experiment demonstrating the core mechanism. A qualified regulatory pathway. A confirmed IP position.

19. Kill Condition

KILL_IF: killer experiment fails: falsification_test_from_candidate (decision scores: EIG 0.44573539382492045, EIG-per-cost 0.44573539382492045; the falsification test carries the pre-registered pass/fail before any transfer)

DISCLOSURE

This dossier distinguishes established evidence, external precedent, computational/modelled results, engineering proposals, and unresolved questions. The absence of physical validation should not be interpreted as evidence that the technology will fail; it means the relevant proposition has not yet been established experimentally.

This dossier does not constitute legal, patentability, FTO, regulatory, or investment advice.