PneumoVision Batch Screening Report

AI-Assisted Chest Radiograph Triage & Opacity Telemetry

Generated: 2026-10-15 09:45:20 | **Total Evaluated Scans:** 4 | **Positive (Suspected Pneumonia):** 2 (50.0%) | **Negative (Clear):** 2

Patient ID & Radiograph	Screening Result	Pulmonary Density Index
Patient-PNEU-01 *Image: pneumonia_01_square_1024x1024_severe_pneumonia.png*	**POSITIVE** Suspected Infiltrate	**0.983** Cutoff 0.300 (High Density)
Patient-NORM-01 *Image: normal_01_square_1024x1024_clear_normal.png*	**NEGATIVE** Clear Lung Fields	**0.052** Cutoff 0.300 (Normal Density)
Patient-PNEU-02 *Image: pneumonia_02_portrait_768x1024_dense_opacity.jpg*	**POSITIVE** Suspected Infiltrate	**0.967** Cutoff 0.300 (High Density)
Patient-NORM-02 *Image: normal_02_portrait_800x1024_healthy_normal.jpg*	**NEGATIVE** Clear Lung Fields	**0.170** Cutoff 0.300 (Normal Density)

Medical Protocol Notice: *PneumoVision screening indices reflect DenseNet-121 operational feature activations trained on the RSNA Pneumonia Detection Challenge benchmark. Pulmonary Density Index ≥ 0.300 represents an operational sensitivity of 90.0% (F2=0.841). This automated report is designed for triage assistance and must be correlated with clinical symptoms and confirmed by a certified radiologist or licensed medical doctor.*