UNIVERSITYOFHYDERABAD

(ACentralUniversityestablishedbyanActofParliament,1974)

Prof. C. R. Rao Road, P.O. Central University Campus, Gachibowli,
Hyderabad - 500 046, Telangana, INDIA *website: www.uohyd.ac.in*

School of ***Life Sciences*** *Department/of:* ***Animal Biology***
*Telephone****:9110384948*** *Email:***prasadtammineni@uohyd.ac.in**

Notice Inviting Tender No. UH/SLS/AB/PT/IA/Fluorescence microscopy /2025-26/01 date:13/11/2025

Sealed tenders are invited for supply of the following items under **two bid systems** from the reputed manufacturers or their authorized dealers to the University of Hyderabad, Hyderabad.

Sealed bids should reach the office of the **Dr. Prasad Tammineni, Room no. F-8, Department of Animal Biology, School of Life Sciences,** ***University of Hyderabad, Hyderabad on or before 12/12/2025(date)*** **5.00 pm** ***(time).***

Technical Specifications for High-Resolution and High-Speed Laser Scanning Confocal Microscope

A Inverted Fluorescent Laser Scanning confocal Microscope with Live Cell Capability, with at least three laser lines, along with a super-resolution module for high-resolution imaging applications

Bids are invited for a state-of-the-art imaging platform consisting confocal "Inverted fluorescent microscope with live cell imaging system, with at least three laser lines and a high-resolution module for a wide range of applications in life sciences and interdisciplinary sciences.

OEM/authorized representatives can bid for any one/single part of the imaging platform.

However, preference will be given to OEMs/their authorized representatives that can offer an integrated, robust imaging platform that can be upgradable to applications like, spinning disk, FRET, FRAP, FLIM and 2-camera system etc,. to fit into evolving research questions in the future.

Sealed bids should reach the Tammineni lab, Room F8, School of Life Sciences, University of Hyderabad, Hyderabad on or before 04th Dec 2025 (Date) 5:30pm (time).

Part A: <u>A confocal 'Inverted Fluorescent Microscope with Live Cell Capability</u>

Sl No	Part A: Specification
1	**Microscope:** Fully Motorized Inverted Fluorescence Microscope for BF/Fluorescence with a dedicated remote touch panel/tab/controller and remote focus knob for controlling motorized components of the microscope. The system should be equipped with a built in Infinity port/stratum structure or double deck system for future laser introduction or FRAP upgradation/optical tweezer upgradation etc. The system should have proper drainage system to avoid any leakage from the dish into the precious optics.
2	Widefield tilting Binocular observation head with paired 10X eye pieces with dioptre adjustment. Bright LED light source for BF, DIC imaging applications controlled by imaging software and touch screen panel.
3	Programmable motorized X-Y scanning stage including sample holders for slides, 35/60 mm Petri dish, Labtech chambers and multi well dishes with multipoint, tile and mosaic imaging software (capable of fast 6D imaging).
4	Hardware & IR laser/LED ($\geq$ 790 nm) Based Focus Drift compensation for long term live cell imaging application is mandatory and must be capable of working with plastic and cover glass bottom dishes. The drift compensator should be controlled by the remote touch screen panel with one touch autofocus to instantly focus the specimen mounted on plastic and glass bottom dishes.
5	Bright long working life LED Light Source with direct coupling for Fluorescence imaging, catering to the DAPI, CFP, YFP, FITC/GFP, RFP, mCherry, Cy3 and Cy5 fluorophores to be included. The light source should facilitate in capturing fast events with microsecond switching via TTL inputs via remote triggering (TTL, microsecond), Individual channel triggering (TTL, microsecond) for on/off control of selected channels. Triggering speed <20 µs at full power. A suitable TTL controller board/card should be supplied controlled by imaging software for fast switching of LED lines.
6	Narrow Band pass fluorescent filters cubes should include for DAPI, FITC/GFP, RFP/ mCherry Cy3 and Cy5 observation and imaging.
7	Motorized 6-position nosepiece with slots for DIC, Motorized Condenser with NA 0.55 or better with optical modules for DIC for respective objectives with built in transmitted shutter.
8	Minimum 8 position or higher motorized filter cube turret with built in motorized shutter for accommodating minimum 7 position fluorescent filters for sample visualization and camera-based imaging and one for BF/DIC Imaging
9	High precision Z-focus drive with step size of 10 nm or better.
10	Objectives: Plan super apochromat objectives - 40x/1.4 or better oil, 60/63x/1.42 oil immersion or better, 100X/1.45 oil or better, and confocal compatible lenses for 10X/0.40, Long Distance 20X/0.40 (WD 7.8 mm or better). DIC attachment for 20X, 40X, 60X and 100X should be the part of standard supply. All will be considered) the objectives should be plan Apo chromatically corrected. All the plan apochromatic objectives should have chromatic correction from 400-1000 nm both laterally and axially (<200nm axial between DAPI and Cy5) for precise colocalization studies in XYZ. (Please provide necessary data on website / brochure) The plan Apochromatic objectives should have uniform field of view (FOV) throughout wide FOV of large Chip SCMOS (>21mm). Low Fluorescence Immersion Oil 100ml should be provided (minimum 5 bottles).

11	Stage top CO_2 incubator setup for long-term live cell imaging including incubation system with Temperature, CO_2 and humidity control. The parameters for incubation system should be controlled by software with the ability to record the temperature, CO2 levels at peridics intervals of atleast 10 minutes . Temp collar/Objective heater for High NA objectives should be provided.
12	**Specifications: Lasers** All lasers should be the latest solid-state lasers with following laser wavelengths and powers. a. 477/488 nm: 100 mW or better b. 555/561 nm, 100 mW or better c. 638/640 nm: 100 mW or better All the Lasers should be stable solid-state laser controlled by Direct Modulation/AOTF for precise switching and swift selection of the desired laser lines. Mention the names of manufacturers of Lasers. The Laser Combiner should be from the microscope company itself for better integration and post-sale support. The laser combiner should onsite be upgradable to additional lasers such as 405, 445, 514 and 594nm. All lasers should be combined with high transmission beam splitters and delivered to the confocal scan head through high transmission efficiency optical fiber. The laser control should be through direct Modulation by high speed break out box for ultra-fast ON/OFF control and intensity modulation with individual laser lines, continuously variable (0 % - 100 %, 1 % increments). The breakout box should also be capable of controlling other third-party hardware such as perfusion system, electrophysiology, laser/ LED pulse pickers for advanced image acquisitions. Should be able to synchronize image camera with the image acquisition through TTL trigger for faster image acquisition.

	High Speed Super Resolution module: The super resolution module should be two-step image reconstruction based on High-Speed pattern illumination/high speed structured illumination module with automated optimization for different wavelengths & magnifications/multipoint spinning disc grid pattern. It should be based on low frequency suppression/filtering and high frequency pattern collection to generate high resolution images of 100-120nm in real time with at least 50FPS with full resolution. The image processing for high frequency patten filtering and high-resolution image construction should be online and real time. In the second step the system should have capability to further improve the lateral resolution through image restoration algorithms such as iterative deconvolution algorithm based a dedicated PSF fitting 3D deconvolution module with a dedicated algorithm to fit the Super Res PSF. The system should be able to generate high speed real time super resolution of both monolayer/2D samples and multilayer cell/tissues/spheroids 3D samples (cleared and uncleared specimens). The frame rate of the super resolution should be 100 FPS or better in full resolution and full FOV of the sCMOS camera with at least 2Kx2K resolution. **Note:** The super resolution image should be hardware based through optical theory. Provide white papers/patents/published research article on the techniques used. All the objectives including the 100X and objective should be compatible with Super resolution mode to achieve the highest spatial resolution of upto 100-120nm with high NA objectives. **Detection:** Detection units should be based on high QE Scientific CMOS Camera (Peak QE @560 nm >90%), resolution of 2Kx2K, pixel Size 6.5 micron X 6.5-micron, effective area of 13x 13 mm, Frame rate of up to 100FPS with full resolution and full FOV with Cam Link/Coax cable connection, Dark Current of < 0.06e/p/s and read noise <0.8 with slow scan (30FPS) and <1.0 with high speed scan (100FPS), dynamic range of 35,000:1, full well capacity of 30,000, The camera must be Peltier cooled (-10 Deg absolute) with active air cooling. The camera should have inlet outlet jacket for water cooling. The system should be upgradable to perform two colour simultaneous imaging capabilities in super resolution mode as well with > 1Kx1K and at least 100 FPS with an appropriate beam splitter. The image software should have multi camera control capabilities with an option to trigger the camera, perfusion pumps and other peripherals if required through TTL. **Emission Filter Wheel:** The system should be equipped with an integrated high speed (40 ms or better with adjacent positions), 10 positions fast filter wheel (2 nos). The filter wheels should be equipped with single band pass emission filters for DAPI, CFP, FITC, YFP, TRITC, mCherry, Cy3 and Cy5. All laser lines for confocal Imaging should be useable for imaging in the super resolution mode, GFP/mCherry should be offered for two color fast sequential imaging. Suitable Primary DMs (405/488/561/640) in the scan head should be offered. Quote for dual band pass (GFP/RFP/mCherry) emission filter for faster two- color excitation sequential imaging.
	4. System control and Imaging Software:

1	Software should be capable of controlling every component of the system including microscope, confocal setup. The system should be able to switch seamlessly between confocal and super resolution mode using the software. It should be compatible for 3D, 4D, 5D and 6D imaging (X,Y,Z,t, multi colour, multi point). Image stitching in co-ordination with XYZ stage with dedicated Software module. The system software should have intuitive modules for multidimensional imaging such as experiment manager / jobs / sequence manager. The software should be capable of performing interactive Quantitative data analysis.
2	Advanced 3D image reconstruction with rendering from a Z- stack image series including 3D deconvolution capability in widefield, confocal and Super resolution deconvolution modalities (for all the PSFs such as widefield, Bright Field, point scanning confocal and spinning disc confocal and super resolution. The deconvolution modules should support importing the raw data from other image formats and perform auto-calibration to render high resolution images. It should be able pick up all the optical and wavelength parameters from the meta data and generate hypothetical PSF for 3 D and 2 D deconvolution. It should also be able to import actual PSF data generated from the sub resolution beads and perform realistic deconvolution.
3	Co-localization and histogram analysis with inbuild Parameters.
5	**Hardware control:** All the motorized components of the hardware including XY stage for multi-point and mosaic imaging, hardware based autofocus module, Emission filter wheels for fast sequential imaging, BF-LED light source for intensity and fast shutter control for multi-channel imaging, third party SCMOS cameras for high speed ratio imaging, third party TTL/triggered device control, DAQ card control for external light sources/lasers and perfusion systems, on stage CO2 incubator and other upgradable hardware modules such as TIRF, FRAP, confocal and super resolution systems.
6	**Image Acquisition Modules:** Simultaneous Dual / Triple / Quad Camera acquisition module for multipoint, multidimensional (XYZtʎ) 6D imaging & experimental preview. **(multi point, stitching module to be offered with XY motorised stage)** Over lay of imaging, Mark and find for multipoint and mosaic imaging, Realtime stitching of large samples with higher magnification, Multipoint and multi position imaging with hardware based autofocus. Online/real-time deconvolution. Modules to perform design and execute complicated experiments through drag and drop functions (Graphical experimental manager/Jobs/journals/experiment designer etc.) High Dynamic Range Imaging, Instant Extended focal imaging, simultaneous imaging of two channels with image splitters and multiple cameras, real time/instant EDF. A Real-Time controller: The controller should provide parallel command execution for motorized components such as LED/Laser switching, shuttering, swift movement of filter wheel and Z movement during fast two color XYZ time lapse imaging. The devise should support fast two color near simultaneous imaging using single camera with perfect over lapping of two colors. The process of closing the shutter and moving the filter wheels & Z steps should begin as soon as camera acquisition is done, during camera chip readout to avoid bleaching. Both filter wheel and laser/LED should switch/move in parallel to save time during two color imaging using single camera for perfect overlapping two colors during fast time lapse imaging (for vesicle tracking experiments).

7	**Image Analysis:** Multidimensional intensity measurement and data export to EXCEL format for further statistical analysis. Colocalization analysis with scatter plot display and statistical analysis, FRET and FRAP analysis, spatiotemporal measurement modules. 3D/volume rendering orthogonal slice view of 3D stacks. Calcium/Physiology/Ratio analysis module, background subtraction/correction, bleaching correction, cell count and confluency measurements, Object Classifier/ automated count and measure etc. The imaging software should be able to image large tissues/organoids/spheroids with lower magnifications and automatically detect them switch over to higher magnifications and acquire images and stitch to generate high resolution images. The system software should have integrated Deconvolution module (2D Real Time/2D/3D), and Artificial Intelligence based object classifier/segmentation and counting module. The imaging software should have an automated object detection based on automated, and AI based object detection with trained neural Network. The imaging software should have an automated object tracking of moving objects and analysis module with an ability to measure distance travelled with velocity of each manually or automatically selected objects. Software Package to automatically analyze moving objects which can divide over time, create a tree representation for the split tracks and get the quantitative information from their image data and export data into a data sheet and EXCEL for further analysis.
	5. Computer and Monitor, AV table:
1	Processor : Intel Xeon W-2123, 3.6GHz 4C CPU 8.25MB, 2666 MHz , Chipset : Intel C422 , HDD : 8 TB (2TB x4)., Optical Drive : 9.5mm slim DVD, Memory : 128 GB (DDR4 SDRAM (2666MHz, ECC, Registered), I/O Slot : IEEE1394 board on PCIe slot, RS232C PCIe board (recommended StarTech Make PEX1394B3), Graphics : NVIDIA Quadro RTX3060Ti 8GB or better, Storage : Two internal 2.4"/3.5: bays, Two external 5.25" bays slimline optical bay, Expansion : 2 PCIe x16, 1 PCIe x8, 2 PCIe x4, 2 M.2, Front Ports : 1Mic/Headphone combo jack, 4 USB 3.1 (1 charging), 2 USB 3.1 Type-C, Rear Ports :6 USB 3.1 Gen 1; 2 RJ-45 (1 GbE); 1 line out; 1 line in; 1 PS/2 mouse port; 1 PS/2 keyboard port; 1 serial port (optional), Optional Ports : Thunderbolt3, Operating System : Windows 10 Pro- 64 Bits (English edition)(Intel Xeon 4 Core System), USB Key board, USB Mouse.State of the art Latest 64-bit control computer with minimum Xeon 6 core Processor, ECC RAM 64 GB or more; 512GB SSDs, 16 TB HDD. Gigabit Ethernet, Windows 11 64 bit Pro or better, USB 2.0/ 3.0, Fire wire for controlling the entire system. and USB 3.0, Fire wire. Integrated Intel® UHD Graphics 630 Discrete NVIDIA TX™ A4000 (16 GB GDDR6 dedicated); System Interface PCI Express 3.0 x 16 Max; with 34” monitor. The computer and all the provided license softwares should be upgradable and compatible to newer version of Windows. One additional offline image analysis software licence (similar analysis capabilities as in the main system) should be provided with separate PC and 27-inch monitor.
2	An appropriate High precision vibration isolation table with automated quite compressor pump as follows: size of 3Ft x 4Ft (900X1500 mm) with automated Air Pump for the above system must be the part of the main offer with M6 individually sealed threaded holes.

The bidder should supply/comply with the following items along with the machine:

1. **Training and demonstration:** Training on the usage of the machine (hardware and software) must be demonstrated by the application specialist of the successful bidder at the bidder's cost to the end users at University of Hyderabad. The successful bidder's application specialist must provide additional training in the first two years.
2. The system should be supplied directly by the principal manufacturer/Authorized distributor as turn-key system. All the components in the system, viz., microscope, lasers, confocal unit, and camera, should be the responsibility of the Primary Supplier, including the Third-Party Components.
3. **Warranty:** The complete spinning disc confocal microscope system is to be under a comprehensive warranty period of a minimum of 5 years, including preventive maintenance, spare parts, lasers, and data analysis software from the date of functional installation. The software should be updated as and when these updates are available. If the instrument or any part of the instrument is found to be defective, it has to be replaced or rectified at the cost of the bidder (including the shipping and custom clearance charges plus any other charges if applicable) within 30 days from the date of receipt of written communications from University of Hyderabad during this warranty period. If there is any delay in replacement or rectification, the warranty period should be correspondingly extended.
5. **Delivery and Installation:** The bidder shall provide the lead-time to delivery, installation and to be made functional at Tammineni lab from the date of receipt of purchase order. The system should be delivered, installed and made functional within 180 days from the date of receipt of purchase order. The system's installation will be considered fulfilled only on satisfactory and complete installation of the system by engineers from the principal manufacturer and demonstration of live-cell imaging on this system and complete training of all features/capabilities of the system to the end-users from the principal manufacturer's application specialist. After successful installation, training completion and inspection, the date of taking over of the entire system by Tammineni lab at University of Hyderabad shall be taken as the start of the warranty period.
6. There should be at least two on-site training for a week one immediately after the installation operation training and the other after 3-6 months for follow up training and standardization training for students.
7. The principal company should provide free online training/troubleshooting/ handholding to standardize new experiment protocols to students for a period of two years.
8. All the terms and conditions mentioned above are subject to final concurrence by the UoH-IAO

Terms&conditions

S. No.	Clause	Description
1	Reference Number	The Reference No**. UH/SLS/AB/PT/IA/Fluroscnece Microscope/2025-26/01 date:13/11/2025 of** the Notice Inviting Tender date **13/11/2025** should be clearly mentioned on all envelopes and any correspondence including e-mail s with the University.
2	Mode of submission	Hand-written (manuscript) bids and bids submitted through the-mail shall not be accepted. All pages of the tender document must be signed by the authorized representatives of bidder, in token of having accepted the terms and conditions incorporated in the tender notice.
3	Two Bid System *(This clause is applicable only when NIT is invited involving two bid system)*	As the purchase is a high value plan machinery,etc.,of a complex and technical nature, bids are required to be submitted in two parts as under: ***(i) Technical bid consisting of all technical details along with commercial terms and conditions; and*** ***(ii) Financial bid indicating item-wise price for the items mentioned in the technical bid.*** The technical bid and the financial bid should be sealed by the bidder in separate covers duly super-scribed and both these sealed covers are to be put in a bigger cover which should also be sealed and super-scribed. The technical bids are to be opened by the Purchasing Department / Centre at the first instance and evaluated by a duly constituted Purchase Committee. After such evaluation of technical bids, financial bids of only those firms/suppliers which are found technically suitable will be opened, for determining the lowest bid.
4	Certificate from original manufacturers	If the bid is not from the original equipment manufacturer, the suppliers/ dealers should submit an 'authorized agency certificate' from the manufacturer. Tenders received without Authorization Certificate from the original manufacturer will summarily be rejected. Imported purchases should normally be mad directly from the foreign principal manufacturers. In the event of purchase made through an Indian agent of the foreign company, authorization letter issued by the foreign company / manufacturer should be submitted along with technical bid. No commission shall be payable to Indian Agent.

5	Make and Model	The Vendor/Supplier is required to clearly mention the make, model, size and other specifications along with the unit price. Items with better specifications may also be offered / quoted and the decision of the University in this regard is final. Appropriate brochures / pamphlets / literature (Original and NOT photo copies) for each item providing full technical details must be attached to enable their verification.

6	Bidder Details	The bidder's complete correspondence address, phone numbers and email ids, and website details (if any), must be clearly mentioned in the technical bid documents. (Pamphlet / Brochure depicting the profile and organization structure may be enclosed.) The details of PAN,GSTIN, EPF, ESIC, Registered License Certificate, Labour License etc., may be mentioned and supporting documents shall be enclosed along with the technical bid.
7	Validity of bids	Prices quoted in the bid (financial bid) shall be valid for a minimum period of 90days from the date of opening of the financial bids. No upward revision of the prices will be permitted during the said validity period.

8	Basic Price	The bidder should clearly indicate price per unit on FOR, UoH, Hyderabad basis. In respect of imported items, the prices should be quoted on the basis of CIP / CIF Hyderabad. *If the bidder quotes nil charges / consideration in respect of service contracts, the bid shall be treated as unresponsive and will not be considered.*

9	Taxes: GST& Customs Duty	The bidder should clearly indicate all applicable taxes separately for each item along with **HSN Code** and then arrive at a total price of all the items indicated in the Notice Inviting Tender. (GST is applicable for all indigenous items. However, the University is entitled for concessional rate of GST @ 5% on all items mentioned vide GoI Notifications No. 45/2017-Central Tax (Rate) and 47/2017-Integrated Tax (Rate) dated 14.11.2017 read with GoI Notification No.09/2018 -Central Tax (Rate) dated 25.01.2018 as amended from time to time.) (Customs Duty is applicable for all imported items. However, the University is entitled for exemption/concessional rate (at5.5%ofthe total value) of Customs Duty on all imported items mentioned vide GoI Notification No. 51/96- Customs dated: 23.07.1996, No. 28/2003 -Customs dated: 01.03.2003, No.43/2017 - Customs dated: 30.06.2017 and DSIR Certificate No. TU/V/RG-CDE (58)/2016 dated: 20.02.2018, as amended from time to time.)
10	Other costs/ charges	The bidder should clearly indicate all applicable other charges, if any, separately viz., transportation, delivering, packing, loading, unloading, insurance, installation, commissioning, testing, training, etc.,

11	Earnest Money Deposit(EMD) / Bid Security	The bidders should furnish bid security for Rs. 3.5% (of total values of the Instrument) along with their technical bid in the form of Account Payee Demand Draft / Banker's Cheque from any of the Commercial Banks in an acceptable form, drawn in favor of Finance Officer, University of Hyderabad, payable at Hyderabad. If the bidder furnishes a Fixed Deposit Receipt / Bank Guarantee, the instrument shall be issued in favor of Finance Officer, University of Hyderabad, payable at Hyderabad and shall be valid for a period of 135days from the date of opening technical bids. The MSE units registered with MSME and certificates obtained from NSIC under the Single Point Registration Scheme (SPRS) shall be exempted from payment of Earnest Money Deposit (EMD) on production of requisite proof in respect of valid certification from NSIC.

		(This clause shall be incorporated in all the NIT, where the estimated cost of purchase exceeds Rs. 2.5 lakh.)
12	Warranty	The bidders should offer a minimum period of 5 years warranty for all the quoted items from the date of successful installation of the equipment/item. The quoted price shall be inclusive of warranty costs. AMC/FMC charges for next two years shall be quoted separately (Optional).

13	Comprehensive Annual Maintenance Contract(CAMC)	Bidders shall quote Comprehensive Annual Maintenance (CAMC) charges for a period of 2 years after expiry of the initial warranty period of 5 years, as optional.
14	Agreement	An Agreement in the format prescribed by the University of Hyderabad is required to be entered into by the Suppliers or their authorized Representative with the University in respect of purchase of Goods / Services valuing Rs. 10 lakh and above, within a period of 15 days from the date of issue of Purchase Order / Work Order. In cases of CAMC, agreements are to be entered into on case to case basis duly customizing the format of agreement. In cases of works awarded by University Works Department, template of agreement finalized by Internal Audit Office, UoH is to be adopted.

15	Performance Security / Performance Bank Guarantee	Successful bidder should submit a Performance Security (in all cases of purchases valuing above Rs. 10 lakh in the form of Performance Bank Guarantee for an amount equivalent to 3% of the total basic price of the Goods/Services(exclusive of taxes and duties)withinaperiodof15 days from the date of issue of Purchase Order/ Work Order. Performance Security should remain valid for the entire warranty period plus additional (grace) period of 60 days, from the date of successful installation of the equipment/item. Performance Security shall be submitted in the form of Account Payee Demand Draft / Fixed Deposit Receipt / Banker's Cheque / Bank Guarantee from any of the Commercial Banks in an acceptable form, drawn in favor of Finance Officer, University of Hyderabad, payable at Hyderabad.
16	Payment	**97%** of the total cost will be released after delivery, satisfactory installation, commissioning, testing and training (if any), against submission of valid TAX Invoice (with **HSN** code) with GST Number of the supplier. The remaining **3%** will be released on submission of a Performance Bank Guarantee (PBG) for an equivalent amount.

		In case of purchases from foreign manufacturers/suppliers, Letter of Credit (LC) will be opened for 100% of the purchase value and **97%**shall be released after signing the Agreement and shipment/submission of documents. As regards balance payment, it will be released on submission of PBG for equivalent amount (in Indian currency) or completion of the warranty period as per the terms included in the NIT and Purchase Order.
17	Delivery Location	The items shall be delivered **Lab No. F8, Department of Animal Biology, School of Life Sciences, University of Hyderabad, Gachibowli, Hyderabad-500046.**
18	Delivery Period	The supplier should be able to deliver the items within **30-45** days *(mention the required delivery period by the proposer)* from the date of issue of the purchase order.
19	Liquidated damages for late Delivery	In the event of delay in delivery beyond stipulated period indicated in the Purchase Order, liquidated damages @ 1% per each week of delay subject to a maximum of 5%ofthe cost of the order will believed and collected by the University, by way of deduction from the payments due to the vendor/supplier.

20	Risk Purchase Clause	If the Supplier fails to deliver the ordered materials within the stipulated delivery period specified in the Purchase Order, the University may resort to procure such items and in such a manner as deemed appropriate, goods or services similar to those undelivered, and the Supplier will be liable to reimburse the University for any excess costs for such similar goods or services.
21	Right to reject	The University reserves the right to reject any or all the bids or accept a part of the quotation without assigning any reasons there for. The decision of the University in this regard is final and binding on all the suppliers.
22	Resolution of Disputes	Any disputes arising out of this contract shall be referred to the University, and if any of the parties here to is dissatisfied with the decision, the dispute shall be referred to the decision of the Arbitrator, who should be acceptable to both the parties, to be appointed by the Vice-Chancellor of the University. The decision of such Arbitrator shall be final and binding on both the parties.
23	Jurisdiction of Courts	Disputes, if any, arising during course of execution of the order are to be settled within the jurisdiction of Hyderabad / Ranga Reddy District Courts only.

24	Opening of tenders/bids	The bids will be opened in the office of **Head, Room No.S-4, Department of Animal Biology, School of Life Sciences, University of Hyderabad,Hyderabad.-500046,Telangana, India.**, on tentatively **15/12/2025**(date) at **3.00pm** (time) tentatively in the presence o bidders or their authorized representatives. If any change in the date will be notified to the bidders. The authorized representative shall bring the authorization letter and submit the same to the University before opening of bids. On failure to do so, the Representative loses the right to participate in the bid opening process.
25	Delay in submission	The University will not in any way be responsible for any postal / courier delay. Bids received beyond the stipulated date and time of submission is summarily rejected. Tenders incomplete in any respect will be summarily rejected.
26	Due date of submission of bids	Sealed Bids should be submitted to the office of **Tammineni lab, Room No.F-8, Department of Animal Biology, School of Life Sciences, University ofHyderabad, Hyderabad.-500046,Telangana, India.**, ***on or before 12/12/2025(date)*****5.00 pm** ***(time).***Bids received there after shall not be considered. *(Minimum time to be allowed for submission of bids should be three weeks for procurement of indigenous items and four weeks in case of imported items from the date of publication of the Notice Inviting Tender in CPP Portal and University of Hyderabad Website)*

Contact details for any queries:



Dr. Prasad Tammineni
Assistant Professor
Department of Animal Biology
School of Life Sciences
Email: prasadtammineni@uohyd.ac.in
Phone: +91-9110384948

Note: InternsofGFR-2017,allNITsforpurchasescostingRs.2.5lakhandabove,shallinvariablybe uploaded in the CPP Portal *(http://eprocure.gov.in)* and e-Tendering / e-Wizard System solution through ITI Limited (as directed by MHRD) by the Dean/Head/Director/PI/Faculty In-charge themselves, by having a separate User ID and Password. In case of any doubt, clarification, users may contact the Nodal Office viz., Director, Computer Centre and CNF for thee-procurement. For the purpose of hosting the NIT in the University Website *(https://www.uohyd.ac.in)*,the NIT shall be sent to Webmaster *(webmaster@uohyd.ac.in)*