

SOH OPD MEDICAL EQUIPMENT/FURNITURE TAGGING

Tag No.	Equipment Description	Additional Comments	Brand name	Electrical Specifics	Procurement/ Delivery Schedule	Installation/ Commisioning Schedule	TANIT Comments for A&K
1	PC DESKTOP 15" MONITOR ALL IN ONE WITH UPS 500 VA						To be assumed by Electrical Engineer
2	PRINTER NETWORK BASED LASER B/W						To be assumed by Electrical Engineer
3	TV 40" - LED - WALL MOUNTED - WITH SUPPORT						To be assumed by Electrical Engineer
4	EXAMINATION HALOGEN LIGHT - ON CASTORS			Equipment is mobile on wheels, Power supply: 220V, 50Hz, 20W	Jan-23	Feb-23	Supplier only shared Catalogue. Electrical Specifications indicated here
5	DIAGNOSTIC STATION - WALL MOUNTED - OPTHALMOSCOPE - OTOSCOPE - NIBP - THERMOMETER		Welch Allen	Equipment is wall mounted; Electrical specifications: 100 – 240 V AC, 50 – 60 Hz, 1.5 – 0.8 A	Jan-23	Feb-23	Please check attached Specs for wall mounting height
6	ULTRA SLIM LIGHT BOX 41X36 CM						Pending From Supplier
7	ECG - 12 CHANNEL - ON TROLLEY						Pending From Supplier
8	DEFIBRILLATOR - UNMONITORED		Progetti	Equipment is mobile, uses rechargeable battery. AC charger power supply, Input: 100 ~ 240V AC 50/60Hz max. Battery pack: 14.4V ~ 2Ah Ni-Mh battery (internal rechargeable) charging time maximum 3 hours	Jan-23	Feb-23	Please also check attached
9	TREADMILL ELEVATING WITH PC CONTROL		GE Case + T2100 Treadmill	200 to 240 VAC, 50/60 Hz, single-phase, 20 amperes, NEMA 6-20 R wall socket (dedicated circuit recommended)	Jan-23	Feb-23	Please also check attached
10	ULTRASOUND UNIT - OB/GYN 3/4D VOLUMETRIC WITH DVD RECORDER		Voluson S10	13A Socket outlet wired with >3 core 2.5mm2 cable, Power Consumption: 900VA Earthing system of less than 2ohms, Voltage 220-240VAC. single phase UPS for Ultrasound; Dedicated AC Power outlet must be within reach of the system without extension cords within 1m of the system. Airconditional Availability	Jan-23	Jan-23	Supplier only shared Catalogue. Electrical Specifications indicated here
11	PLETHYSMOGRAPH CABINET - COMPACT SIZE WITH CHAIR						Pending From Supplier
12	ELECTROENCEPHALOGRAPH STATION - 32 CHANNEL TROLLEY MOUNTED - WITH ACCESSORIES		Neurowork	Power frequency (50/60 Hz) magnetic field IEC 61000-4-8	Jan-23	Feb-23	No datasheet/Catalogue shared but electrical specifications indicated here. Please let us know if this is sufficient
13	ENT STATION WITH CHAIR AND INSTRUMENTS - VIDEO MONITOR	Total station inculding Chair	Euroclinic	220/240 V(110/115 VAC)	Jan-23	Feb-23	No datasheet/Catalogue shared but electrical specifications indicated here. Please let us know if this is sufficient
14	ENT MICROSCOPE		Leica	Equipment is mobile on easter wheels. Power supply: 100-240V AC, 50/60Hz. Power consumption: 1200VA	Jan-23	Feb-23	Please also check attached
15	ELECTROSURGICAL UNIT 200 WATTS - ENT	On ENT Station					Pending From Supplier
16	LAMP - DESKTOP						To be assumed by Electrical Engineer
17	WATER DISPENSER						To be assumed by Electrical Engineer
18	FRIDGE - SMALL						To be assumed by Electrical Engineer
18.1	MICROWAVE SMALL						To be assumed by Electrical Engineer
18.2	SCREENER - VISUAL FIELD COMPUTER BASED	May be a TV or a projector depending on the brand					Pending From Supplier
19	CEPHALO SCAN CBCT		Belmont	Power Requirement x-ray will operate on a power supply of 120 + 10 VAC. A three wire GROUNDED circuit, separately connected to the central distribution panel with an over current protection device rate for 20 amperes. Recommended wire size is 12 AWG. But if the wire run distance is to exceed 50 feet, 10 AWG is required. For wire run distance in excess of 75 feet, upto 125 feet, 8 AWG is required. Line voltage regulation must exceed 3 % at 12.5 amps.	Jan-00	Feb-23	Please also check attached
20	DENTAL CHAIR BUILT IN COMPRESSOR COMPLETE		Olsen	Equipment is powered by a Motor: 24V, 46W, 45rpm	Jan-23	Feb-23	Supplier only shared Catalogue. Electrical Specifications indicated here

21	DENTAL X-RAY UNIT WALL MOUNTED		Belmont	Maximum Length Reach of equipment : 84-1/8" with 40.0" arm Power Supply Electrical requirement Power supply of 120VAC ± 10% (108 ~ 132V AC) with a 3 wire GROUNDED circuit, separately connected to the central distribution panel with an over current protection device rated for 15 amperes. Recommended wire size is 14 AWG ; but if the wire run distance is to exceed 50 feet (15m), 12 AWG is required. For wire run distances in excess of 75 feet (23m) 10 AWG is required. Line voltage regulation must be within 0 ~ 3 % at 6.5 amps. Please check drawing on separate sheet	Jan-23	Jan-23	Please also check attached
22	DENTAL CURING LIGHT - CORDLESS - LED		Woodpecker	Equipment is mobile, uses Rechargeable lithium battery. Power input: 100V to 240V, 50/60Hz, 10VA	Jan-23	Feb-23	Please also check attached
23	AUTOClave - DENTAL INSTRUMENTS 24L		Tuttnauer/ 2840EKA D(28L)	Voltage/ Frequency (Hz): 230V/60Hz Power: 1400W 6A Equipment is Table Top with Dimension 630*530*440mm	Jan-23	Feb-23	Supplier only shared Catalogue. Electrical Specifications indicated here
24	LASER - DENTAL SURGICAL/DIODE/TABLETOP FOR SOFT TISSUE		Zolar	Equipment is table-top, uses Battery AC100-240V 50/60Hz Wireless Pedal 1.5V Battery	Jan-23	Feb-23	Please also check attached
25	AMALGAMATOR		Henry Schein	Equipment is Table-top Input Voltage: 110V, 60Hz Power Output: 40W	Jan-23	Feb-23	Please also check attached
26	RETINAL CAMERA / OCT OPHTHALMOSCOPE / EYE F-ANGIO / TABLETOP		Topcon	Equipment is Table-top Input Voltage: 110=240V, 50-60Hz Power Consumption: 400VA (maximum) 100VA (normal) Dimensions W*D*H: (274*508*566mm)	Jan-23	Feb-23	Please also check attached
27	PROJECTOR - VISUAL ACUTY						Pending From Supplier
28	AUTO REFRACTOMETER/KERATOMETER						Pending From Supplier
28.1	LENSOMETER - TABLETOP						Pending From Supplier
29	OPHTHALMOLOGY CHAIR - COMPLETE (RETINOSCOPE - SLIT LAMP - TONOMETER)		Light Box				Pending From Supplier
30	LED VIDEO COLPOSCOPE - 3.75X/7X/15X > 28000 LUX - ON TROLLEY		Ecleris/ C100-FID Colposcope	Equipment is mobile on caster wheels. Power supply: 100-240V AC, 50/60Hz.	Jan-23	Feb-23	Supplier only shared Catalogue. Electrical Specifications indicated here
31	BOOTH - AUDIOMETRY - COMPACT WITH VIEWPORT AND CHAIR - 1.2X1.2M						Pending From Supplier
32	ERGOMETER		GE	Power Supply: 100 to 240 V, 50 to 60 Hz Power Consumption: 60VA max	Jan-23	Feb-23	Please also check attached
33	ELECTROTHERAPY UNIT						Pending From Supplier
33.1	PRESSOTHERAPY UNIT						Pending From Supplier
33.2	VACUUM THERAPY						Pending From Supplier
33.3	INFRARED RADIANT						Pending From Supplier
33.4	MAGNETOTHERAPY						Pending From Supplier
33.5	CONTINUOUS PASSIVE MOTION UPPER LIMBS			Power supply: 220V AC, 50Hz. Input Power: <60VA	Jan-23	Feb-23	No datasheet/Catalogue shared but electrical specifications indicated here. Please let us know if this is sufficient
33.6	CONTINUOUS PASSIVE MOTION LOWER LIMBS						Pending From Supplier
33.7	HYDROCOLLECTOR - COLD COMPRESS			Power Supply: 100 to 240 V, 50 to 60 Hz	Jan-23	Feb-23	Please also check attached
33.8	HYDROCOLLECTOR - HOT COMPRESS			Power Consumption: 1500W max	Jan-23	Feb-23	Please also check attached
34	ULTRASOUND UNIT - CARDIAC - WITH ASSESSMENT SOFTWARE - 3 PROBE		GE Vivid S7	13A Socket outlet wired with >3 core 2.5mm2 cable, Power Consumption: 0.5-1A Earthing system of less than 2ohms, Voltage 240 VAC single phase UPS for Ultrasound Airconditional Availability	Jan-23	Jan-23	Supplier only shared Catalogue. Electrical Specifications indicated here
35	COMPUTER-BASED URODYNAMIC SYSTEM ON CASTORS						Pending From Supplier
36	URODYNAMIC CHAIR						Pending From Supplier
37	REFRIGERATOR 500 L		Haier	Equipment is mobile on caster wheels. Power supply: 100-240V AC, 50/60Hz.	Jan-23	Jan-23	Please also check attached
38	PORCELAIN FURNACE WITH VACUUM PUMP		European				Pending From Supplier
39	BURNOUT FURNACE		European				Pending From Supplier
40	CASTING MACHINE - MANUAL		European				Pending From Supplier
41	SANDBLASTER		European				Pending From Supplier
42	MICRO SANDPLASTER		European				Pending From Supplier

43	VIBRATOR	Supplier to provide exact final positioning	European				Pending From Supplier
44	HANGING MOTOR WITH HANPIECE		European				Pending From Supplier
45	MICRO MOTOR		European				Pending From Supplier
46	VACCUM MIXER		China				Pending From Supplier
47	PINDEX		China				Pending From Supplier
48	MODEL TRIMMER		European				Pending From Supplier
49	STEAM MACHINE		China				Pending From Supplier
50	MOTOR CENTRIFUGE		UK				Pending From Supplier
51	POLISHING MOTOR		European				Pending From Supplier
52	INSTRUMENT CLEANER						Pending From Supplier
53	SEALING MACHINE						Pending From Supplier
54	ORTHODONTIC SPOT WELDER						Pending From Supplier
55	VACUUM FORMER TABLETOP DENTAL LAB						Pending From Supplier