

# Sertifikat Kalibrasi

## Calibration Certificate

NO: 407820

### A. DATA ALAT

#### Instrumen Identification

**Nama** : Anasthesia Workstation  
*Name*  
**Merk** : Oricarw  
*Manufacturer*  
**Tipe** : A9800  
*Type*  
**No.Seri** : YTZW002A9800S001PI  
*Serial Number*

### B. IDENTITAS PELANGGAN

#### Customer Identification

**Nama** : RS Tk. II Udayana  
*Name*  
**Alamat** : Jl. P.B. Sudirman No. 1  
*Address*  
Denpasar, Bali

**Sertifikat Ini Terdiri Dari** : 2 Halaman

*This Certificate comprises of*

**Diterbitkan Tanggal** : 25 Agustus 2022

*Date of Issue*

**Kalibrasi Berikutnya** : 24 Agustus 2023

*Next Calibration*

**Kepala Laboratorium Kalibrasi**

*Head Of Calibration Laboratory*

**PT. GLOBAL PROMEDIKA SERVICES**



**Hamzah, S.T.**

- Tidak diperkenankan mengutip/memperbanyak sertifikat ini tanpa seizin dari PT. Global Promedika Services.
- Sertifikat ini sah apabila telah dibubuhi cap PT. Global Promedika Services dan telah ditandatangani oleh pejabat yang berwenang.
- Hasil kalibrasi ini hanya berlaku untuk alat yang tertera pada Sertifikat ini.

- This certificate shall not be reproduced except in full unless permission for the reproduction of an approved abstract has been obtained in writing from PT. Global Promedika Services.
- This certificate is valid if it has been approved and signed by authorized person.
- This result of calibration to be applied only for unit under test.

Halaman 1 dari 2

**PT. GLOBAL PROMEDIKA SERVICES**

Gedung Bersaudara Lt. 4A, Jl. Penjemihan Raya No. 38 Jakarta Pusat 10210.  
Telp. (021) 5743280, 5701467

[www.globalpromedika.co.id](http://www.globalpromedika.co.id)