protocols.io

🌐 ELISA for quantification of IL-21 in human serum.

Angel A Justiz-Vaillant[1]

[1]University of the West Indies St. Augustine

Aug 21, 2020

In Development   dx.doi.org/10.17504/protocols.io.bj3pkqmn

University of the West Indies   angel.vaillant@sta.uwi.edu

Angel Justiz-Vaillant
University of the West Indies St. Augustine

ABSTRACT

Interleukins (IL) are a type of cytokine first thought to be expressed by leukocytes alone but have later been found to be produced by many other body cells. They play essential roles in the activation and differentiation of immune cells, as well as proliferation, maturation, migration, and adhesion. They also have pro-inflammatory and anti-inflammatory properties. The primary function of interleukins is, therefore, to modulate growth, differentiation, and activation during inflammatory and immune responses. Interleukins consist of a large group of proteins that can elicit many reactions in cells and tissues by binding to high-affinity receptors in cell surfaces.

NK cells and CD4+ T lymphocytes make IL-21. It acts on various immune cells of innate and the adaptive immune systems. IL-21 promotes B and T lymphocyte proliferation and differentiation. It enhances NK cell activity. [1]

Reference

1. Justiz Vaillant AA, Qurie A. Interleukin. In:*StatPearls*. Treasure Island (FL): StatPearls Publishing; June 12, 2019.

DOI

dx.doi.org/10.17504/protocols.io.bj3pkqmn

PROTOCOL CITATION

Angel A Justiz-Vaillant 2020. ELISA for quantification of IL-21 in human serum.. **protocols.io**
https://dx.doi.org/10.17504/protocols.io.bj3pkqmn

CREATED

Aug 21, 2020

LAST MODIFIED

Aug 21, 2020

PROTOCOL INTEGER ID

40783

1   An anti-human IL-21 coating antibody is adsorbed onto the microwells by incubation overnight at 4°C.

2   Add 50 μl of human serum. Human IL-21 present in the serum sample binds to antibodies adsorbed to the microwells.

3   The microplate is blocked with 3% non-fat milk-PBS buffer and later wash to remove unbound proteins.

4   Fifty (50) µl of biotin-conjugated anti-human IL-21 antibody is added. The optimal dilution must be investigated.

5   The microplate is rewashed with PBS-Tween 20 buffer, pH 7.4.

6   One hundred µl of streptavidin-HRP conjugate is added and it binds to the biotin-conjugated anti-human IL-21 antibody. The optimal dilution of this conjugate must be investigated.

7   The plate is washed following incubation to remove the unbound Streptavidin-HRP.

8   Add 100 µl of 3',3',5',5'- tetramethylbenzidine (TMB; Sigma-Aldrich) to each well.

9   Incubate the microwells in the dark for 15 min.

10 A colored product is formed in proportion to the quantity of human IL-21 present in the sample or standard.

11 The reaction is terminated by addition of 100 µl 3M H_2SO_4 and the absorbance is measured at 450 nm.

12 A standard curve is made from 7 human IL-21 standard dilutions and the human IL-21 sample concentration is determined.

13 For better results place the microplate on a microplate shaker in every incubation.