



Aug 19, 2020

Improved protocol of ELISA for determining the serum concentration of IL-23 in humans.

Angel A Justiz-Vaillant¹¹University of the West Indies St. Augustine
1 *Works for me* dx.doi.org/10.17504/protocols.io.bjzakup2e

University of the West Indies

angel.vaillant@sta.uwi.edu

Angel Justiz-Vaillant
University of the West Indies St. Augustine

ABSTRACT

A previous protocol was contaminated with IL-17. This is an improved version of an ELISA for quantification of Interleukin 23. It activates the production of IL-17A, which is a pro-inflammatory cytokine, produced by a group of T helper cell known as T-helper17 lymphocytes in response to the said IL-23 [1,2].

Reference

1. Cho ML, Kang JW, Moon YM, Nam HJ, Jhun JY, Heo SB, Jin HT, Min SY, Ju JH, Park KS, Cho YG, Yoon CH, Park SH, Sung YC, Kim HY (May 2006). Journal of Immunology. 176 (9): 5652–61.
2. Nobee A, Justiz-Vaillant A, Akpaka PE, Poon-King P (2016) Levels of Interleukin 17 and 23 in Patients with Systemic Lupus Erythematosus (SLE) in Trinidad and Tobago. Immunochem Immunopathol 2: 115. doi: 10.4172/2469-9756.1000115

DOI

dx.doi.org/10.17504/protocols.io.bjzakup2e

PROTOCOL CITATION

Angel A Justiz-Vaillant 2020. Improved protocol of ELISA for determining the serum concentration of IL-23 in humans.. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.bjzakup2e>

LICENSE

————— This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

CREATED

Aug 19, 2020

LAST MODIFIED

Aug 19, 2020

PROTOCOL INTEGER ID

40706

MATERIALS

NAME	CATALOG #	VENDOR
Set of one 96-well filter plate with 2 plate sealers	MX-PLATE	Millipore
Interleukin 23 by IBL corporation Germany		

- 1 Ninety-six well ELISA plates are coated with monoclonal anti-human antibodies to interleukin-23 (IL-23).
- 2 Patient serum samples are added to the plates.
- 3 The plate is incubate x 1.30 hour at RT.
- 4 The plate is washed 4 times with PBS-tween 20 buffer.
- 5 The wells are incubated with a biotin conjugated anti-human IL-23 for 1.30 hour at RT.
- 6 The plates are washed again as above.
- 7 To the plate a peroxidase-labeled streptavidin conjugate is added and incubated for 1 hour at RT.
- 8 After a further washing procedure a substrate solution reactive is added and allowed to produced a colored reaction in positive controls.
- 9 The level of IL-23 in the sample is proportional to the colored product developed.
- 10 The addition of 3M H₂SO₄ stops the reaction.
- 11 The absorbance is measured at 450 nm.
- 12 The IL-23 concentration can be calculated by generating an standard curve using lineal regression.