

Sep 03, 2020

© ELISA for quantification of Granulocyte-colony stimulator factor (G-CSF) in human serum or plasma

Angel A Justiz-Vaillant¹, Belkis Ferrer-Cosme²

¹University of the West Indies St. Augustine; ²"Saturnino Lora Torres" Provincial Teaching Clinical Surgical Hospital. Cuba

In Development dx.doi.org/10.17504/protocols.io.bkspkwdn

Angel Justiz-Vaillant University of the West Indies St. Augustine

ABSTRACT

Interleukins (IL) are a type of cytokine first thought to be expressed by leukocytes alone but have later been found to be produced by many other body cells. They play essential roles in the activation and differentiation of immune cells, as well as proliferation, maturation, migration, and adhesion. They also have pro-inflammatory and anti-inflammatory properties. The primary function of interleukins is, therefore, to modulate growth, differentiation, and activation during inflammatory and immune responses. Interleukins consist of a large group of proteins that can elicit many reactions in cells and tissues by binding to high-affinity receptors in cell surfaces. [1]

The immunoregulatory cytokine IL-41 (also known as meteorin-like protein) is expressed at high levels in the synovium of patients with psoriatic arthritis (PsA).[2]

Reference

1. Justiz Vaillant AA, Qurie A. Interleukin. In: Stat Pearls. Treasure Island (FL): Stat Pearls Publishing; June 12, 2019.

2.Onuora S. Novel cytokine, IL-41, linked with PsA. *Nat Rev Rheumatol.* 2019;15(11):636. doi:10.1038/s41584-019-0314-7

DO

dx.doi.org/10.17504/protocols.io.bkspkwdn

PROTOCOL CITATION

Angel A Justiz-Vaillant, Belkis Ferrer-Cosme 2020. ELISA for quantification of Granulocyte-colony stimulator factor (G-CSF) in human serum or plasma. **protocols.io** https://dx.doi.org/10.17504/protocols.io.bkspkwdn

LICENSE

This is an open access protocol distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

CREATED

Sep 03, 2020

LAST MODIFIED

Sep 03, 2020

PROTOCOL INTEGER ID

41519

DISCLAIMER:

DISCLAIMER - FOR INFORMATIONAL PURPOSES ONLY; USE AT YOUR OWN RISK

mprotocols.io

09/03/2020

Citation: Angel A Justiz-Vaillant, Belkis Ferrer-Cosme (09/03/2020). ELISA for quantification of Granulocyte-colony stimulator factor (G-CSF) in human serum or plasma. https://dx.doi.org/10.17504/protocols.io.bkspkwdn

The protocol content here is for informational purposes only and does not constitute legal, medical, clinical, or safety advice, or otherwise; content added to <u>protocols.io</u> is not peer reviewed and may not have undergone a formal approval of any kind. Information presented in this protocol should not substitute for independent professional judgment, advice, diagnosis, or treatment. Any action you take or refrain from taking using or relying upon the information presented here is strictly at your own risk. You agree that neither the Company nor any of the authors, contributors, administrators, or anyone else associated with <u>protocols.io</u>, can be held responsible for your use of the information contained in or linked to this protocol or any of our Sites/Apps and Services.

1	An anti-human granulocyte-colony stimulator factor (G-CSF) coating antibody is adsorbed onto the microwells by incubation overnight at 4°C with carbonate-bicarbonate buffer.
2	Add 50 μ l of human serum. Human G-CSF present in the serum sample binds to antibodies adsorbed into the microwells.
3	The microplate is blocked with 3% non-fat milk-PBS buffer and later wash to remove unbound proteins.
4	Fifty (50) μ I of biotin-conjugated anti-G-CSF antibody is added. The optimal dilution must be investigated.
5	The microplate is rewashed with PBS-Tween 20 buffer, pH 7.4.
6	One hundred μl of streptavidin-HRP conjugate is added and it binds to the biotin-conjugated anti-human G-CSF antibody.
7	The plate is washed following incubation to remove the unbound Streptavidin-HRP.
8	Add 100 µl of 3,3',5,5'- tetramethylbenzidine (TMB; Sigma-Aldrich) into each well.
9	Incubate the microwells in the dark for 15 min.
10	A colored product is formed in proportion to the quantity of G-CSF present in the sample or standard.
11	The reaction is terminated by addition of 100 μ l 3M H2SO4 and the absorbance is measured at 450 nm.

12	A standard curve is made from 7 human G-CSF standard dilutions and the human G-CSF sample concentration is determined.
13	For better results place the microplate on a microplate shaker in every incubation.