



Aug 26, 2020

ELISA for quantification of IL-4 in human serum.

Angel A Justiz-Vaillant¹¹University of the West Indies St. Augustine

In Development

dx.doi.org/10.17504/protocols.io.bkbrksm6

Angel Justiz-Vaillant

University of the West Indies St. Augustine

ABSTRACT

Interleukins (IL) are a type of cytokine first thought to be expressed by leukocytes alone but have later been found to be produced by many other body cells. They play essential roles in the activation and differentiation of immune cells, as well as proliferation, maturation, migration, and adhesion. They also have pro-inflammatory and anti-inflammatory properties. The primary function of interleukins is, therefore, to modulate growth, differentiation, and activation during inflammatory and immune responses. Interleukins consist of a large group of proteins that can elicit many reactions in cells and tissues by binding to high-affinity receptors in cell surfaces. [1]

The immunoregulatory cytokine IL-41 (also known as meteorin-like protein) is expressed at high levels in the synovium of patients with psoriatic arthritis (PsA).[2]

Reference

1. Justiz Vaillant AA, Qurie A. Interleukin. In: *StatPearls*. Treasure Island (FL): StatPearls Publishing; June 12, 2019.
2. Onuora S. Novel cytokine, IL-41, linked with PsA. *Nat Rev Rheumatol*. 2019;15(11):636. doi:10.1038/s41584-019-0314-7

DOI

dx.doi.org/10.17504/protocols.io.bkbrksm6

PROTOCOL CITATION

Angel A Justiz-Vaillant 2020. ELISA for quantification of IL-4 in human serum.. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.bkbrksm6>

LICENSE

— This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

CREATED

Aug 26, 2020

LAST MODIFIED

Aug 26, 2020

PROTOCOL INTEGER ID

41041

DISCLAIMER:

DISCLAIMER – FOR INFORMATIONAL PURPOSES ONLY; USE AT YOUR OWN RISK

The protocol content here is for informational purposes only and does not constitute legal, medical, clinical, or safety advice, or otherwise; content added to protocols.io is not peer reviewed and may not have undergone a formal approval of any kind. Information presented in this protocol should not substitute for independent professional

judgment, advice, diagnosis, or treatment. Any action you take or refrain from taking using or relying upon the information presented here is strictly at your own risk. You agree that neither the Company nor any of the authors, contributors, administrators, or anyone else associated with protocols.io, can be held responsible for your use of the information contained in or linked to this protocol or any of our Sites/Apps and Services.

- 1 An anti-human IL-4 coating antibody is adsorbed onto the microwells by incubation overnight at 4°C with carbonate-bicarbonate buffer.
- 2 Add 50 µl of human serum. Human IL-4 present in the serum sample binds to antibodies adsorbed into the microwells.
- 3 The microplate is blocked with 3% non-fat milk-PBS buffer and later wash to remove unbound proteins.
- 4 Fifty (50) µl of biotin-conjugated anti-human IL-4 antibody is added. The optimal dilution must be investigated.
- 5 The microplate is rewashed with PBS-Tween 20 buffer, pH 7.4.
- 6 One hundred µl of streptavidin-HRP conjugate is added and it binds to the biotin-conjugated anti-human IL-4 antibody. The optimal dilution of this conjugate must be investigated.
- 7 The plate is washed following incubation to remove the unbound Streptavidin-HRP.
- 8 Add 100 µl of 3,3',5,5'-tetramethylbenzidine (TMB; Sigma-Aldrich) into each well.
- 9 Incubate the microwells in the dark for 15 min.
- 10 A colored product is formed in proportion to the quantity of human IL-4 present in the sample or standard.
- 11 The reaction is terminated by addition of 100 µl 3M H₂SO₄ and the absorbance is measured at 450 nm.
- 12 A standard curve is made from 7 human IL-4 standard dilutions and the human IL-4 sample concentration is determined.

13 For better results place the microplate on a microplate shaker in every incubation.