





BGISEQ-500 Sequencing V.2

Jie Huang¹, Xinming Liang², Yuankai Xuan³, Chunyu Geng², Yuxiang Li², Haorong Lu², Shoufang Qu¹, Xianglin Mei³, Hongbo Chen¹, Ting Yu¹, Nan Sun¹, Junhua Rao², Jiahao Wang⁴, Wenwei Zhang², Ying Chen², Sha Liao², Hui Jiang², Xin Liu², Zhaopeng Yang¹, Feng Mu², Shangxian Gao¹

¹National Institutes for Food and Drug Control (NIFDC); ²BGI-Shenzhen; ³State Food and Drug Administration; ⁴BGI-Qingdao

1 Works for me

This protocol is published without a DOI.

BGI



Hongling Zhou

ABSTRACT

BGISEQ-500 is a new desktop sequencer developed by BGI. Using DNA nanoball and combinational probe anchor synthesis developed from Complete Genomics™ sequencing technologies, it generates short reads at a large scale.

EXTERNAL LINK

https://doi.org/10.1093/gigascience/giy144

EXTERNAL LINK

https://doi.org/10.1093/gigascience/giy144

PROTOCOL CITATION

Jie Huang, Xinming Liang, Yuankai Xuan, Chunyu Geng, Yuxiang Li, Haorong Lu, Shoufang Qu, Xianglin Mei, Hongbo Chen, Ting Yu, Nan Sun, Junhua Rao, Jiahao Wang, Wenwei Zhang, Ying Chen, Sha Liao, Hui Jiang, Xin Liu, Zhaopeng Yang, Feng Mu, Shangxian Gao 2020. BGISEQ-500 Sequencing. **protocols.io** https://protocols.io/view/bgiseq-500-sequencing-bngzmbx6

Version created by Hongling Zhou

EXTERNAL LINK

https://doi.org/10.1093/gigascience/giy144

LICENSE

This is an open access protocol distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

CREATED

Oct 15, 2020

LAST MODIFIED

Oct 15, 2020

PROTOCOL INTEGER ID


43257

ABSTRAC1


BGISEQ-500 is a new desktop sequencer developed by BGI. Using DNA nanoball and combinational probe anchor synthesis developed from Complete Genomics™ sequencing technologies, it generates short reads at a large scale

Test the quality of the sequencing library (see protocol for library preparation) by the Qubit® ssDNA Assay Kit and homogenized at 6ng total amounts.

- 2 Carry out Rolling circle amplification MRCAMfor 10 minutes in an 80 ul reaction volume with pure water, buffer and DNB polymerase
 - © 00:10:00
- 3 Add 20 ul DNBs stopping buffer to stop the RCA reaction
- 4 Check the quality of the DNBs using the Qubit® ssDNA Assay Kit, concentration should be above 10 ng/µL



- 5 Add 33ul DNBs loading buffer to the DNBs product from the last step, and place it in the BGIDL-50 (the DNBs loading machine)
- $\begin{tabular}{ll} \textbf{6} & \textbf{install the sequencing chip and selected the DNBs loading process (Version: sample load 2.0) to load the DNBs.} \end{tabular}$



7 after loading, take the sequencing chip out and install in the BGISEQ-500 MDNBSEQ-G50 Msequencing machine

- 8 load the reagent sequencing kit and open the sequence control software (Version 1.1.0.10003), and select the sequence process Version 1.0.06 and Zebracall process Version 0.5.0.13875 for sequencing.
- 9 Sequencing is initiated after the sequencing reagents preloaded and sequencing chip was installed, and this process is finished in about 72 hours. When the whole sequencing was finished, the binary file with bases and quality score were converted into FASTQ format with Phred+33 quality score.