

May 14, 2024

NanoString GeoMx DSP TMA-TNP Phase 4 Protein assay



Forked from NanoString GeoMx DSP TMA-TNP protein assay

DOI

dx.doi.org/10.17504/protocols.io.5qpvokbm7l4o/v1

Jinho Lee^{1,2}, Gabriel Zangirolani², Koei Chin¹, Heidi S Feiler¹, Christopher Corless²

¹Knight Cancer Institute, OHSU; ²Knight Diagnostics Laboratory, OHSU

NCIHTAN

OMS Atlas



Jinho Lee

OHSU

OPEN ACCESS



DOI: dx.doi.org/10.17504/protocols.io.5qpvokbm7l4o/v1

Protocol Citation: Jinho Lee, Gabriel Zangirolani, Koei Chin, Heidi S Feiler, Christopher Corless 2024. NanoString GeoMx DSP TMA-TNP Phase 4 Protein assay. **protocols.io** https://dx.doi.org/10.17504/protocols.io.5qpvokbm7l4o/v1

License: This is an open access protocol distributed under the terms of the **Creative Commons Attribution License**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's

working

Created: June 02, 2022

Last Modified: May 14, 2024

Protocol Integer ID: 99588

Keywords: NanoString, GeoMx, Digital Spatial Profiler, HTAN, TMA, DSP, cell profiling, cell type



Funders Acknowledgement: Human Tumor Atlas Network Grant ID: NIH NCI U2C CA233280

Abstract

This protocol outlines the NanoString GeoMx DSP phase 4 protein assay that was applied in the Human Tumor Atlas Network (HTAN) Tissue MicroArrary (TMA)-TransNetwork Project (TNP).

The TMA-TNP evaluates various characterization and analytics methodologies on a large array of breast tumor samples representing a broad spectrum of disease state and subtype. A commercially available anonymized breast tumor TMA was purchased and serial sections were distributed. Participating HTAN Centers characterized the FFPE specimens using various imaging platforms and generated a spatially resolved cell type/state census using each center's method of choice. Data was recorded in a common repository to enable joint analysis.

The protocol that immediately precedes this one for TMA-TNP Phase 4 can be found at: <u>dx.doi.org/10.17504/protocols.io.ewov1o7wolr2/v1</u>. It describes FFPE block serial sectioning, slide processing and TMA sample distribution.

In this protocol, the DSP protein assay was performed with the following human marker panels: Immune cell profiling core, Immune activation status, Immune-Oncology (IO) Drug target, Cell death, Pan-tumor, MAPK and PI3K/AKT panels including OHSU custom panel (Cell cycle and DNA damage). A total of 85 protein targets were evaluated in the TMA samples. Two compartments (Tumor and Stroma) in each TMA core were analyzed to determine cell-to-cell interactions in the tissues.



Materials

1. Materials

Slide Baking Oven

Antigen retrieval pressure cooker

Humidity chamber

Thermocycler

Centrifuge (up tp 2,000g)

2. Reagents

Xylene

Ethanol

TBS

TBS-T

Citrate buffer pH6.0

10% Neutral Buffered Formalin (NBF)

DEPC-treated water

3. Nanostring reagents

GeoMx Instrument Buffer kit / Collection plate (Item no. 100474)

FFPE slide prep kit (Buffer W / Buffer S) (Item no. 121300312)

Human Immune cell profiling core panel (Item no. 121300101)

Human Immune activation status panel (Item no. 121300103)

Human IO Drug target panel (Item no. 121300102)

Human Cell death panel (Item no. 121300112)

Human MAPK panel (Item no. 121300111)

Human PI3K/AKT panel (Item no. 121300113)

Human Pan-tumor panel (Item no. 121300105)

OHSU SQ-35430 custom protein panel

Solid tumor Morphology kit (Anti-PanCK / Anti-CD45 / SYTO-13) (Item no. 121300301)

Hybridization CodeSets (A-H) (Item no. 121300401)

Master kit / Hybridization buffer (Item no. 100052)



FFPE slide sample preparation 1h 59m 1 FFPE slide preparation 1.1 Bake FFPE slides at \$\mathbb{\ceil} 60 \circ for \chircles 01:00:00 \tag{.} 1h 1.2 Deparaffinize by sequential incubation with Xylene () 00:03:00 twice), 100% EtOH (9m ♠ 00:01:00), 95% EtOH (♠ 00:01:00) and 70% EtOH (♠ 00:01:00). 1.3 Briefly rinse the slides with diH₂O (distilled water) and incubate with 1x Citrate antigen retrieval 15m buffer (Epitope retrieval buffer, ph 6.0) for 00:15:00 at high pressure in a pressure cooker. 1.4 Take out the slides from the pressure cooker and let them cool down to room temperature 25m around (5) 00:25:00 . 1.5 Wash the slides 5 times with TBS-T for 00:02:00 each. 10m Antibody panel incubation 1h 59m 2 DSP protein panel and visualization marker incubation 2.1 Mark off the entire tissue section with a hydrophobic pen to create a reagent boundary. 2.2 Block the slides for 01:00:00 at room temperature using the Nanostring blocking buffer 1h (Buffer W) 2.3 Freshly prepare each commercial protein panel in 1:25 dilution and each visualization marker (PanCK and CD45) in 1:40 dilution with Buffer W to a volume of △ 220 µL per slide.



For the custom antibody panel, the final concentration at 250 ng/ml per antibody is calculated and adjusted to the final volume of 4 220 µL.

2.4 Incubate slides with all antibody panels and visualization markers for 🚫 Overnight at **4** °C .

16h

Post fixation and DNA staining

1h 18m

- 3 Post fixation
- 3.1 After antibody incubation, wash the slides 3 times with TBS-T for (2) 00:03:00 each.

9m

3.2 Fix the slides with 10% NBF (Neutral buffered Formalin) for 00:30:00 at RT.

30m

3.3 Wash 3 times with TBS-T for 00:03:00 each, then stain the samples with SYTO-13, prepared by 1:10 dilution in TBS for 00:15:00 at RT.

18m

3.4 Slides are briefly washed 2 times with TBS-T for 00:03:00 each.

6m

Loading samples to DSP

- 4 DSP run preparation
- 4.1 Scrape the hydrophobic barrier off with a scalpel or a straight-edged razor.
- 4.2 Place slides on the GeoMx slide tray and clean the back of the slide with 70% EtOH.
- 4.3 Once the gasket is sealed, place 6 mL of Buffer S on slides.



4.4 Load the GeoMx slide tray onto the DSP platform.

GeoMx operation 5 Slide scan, ROI (Region of Interest) selection and AOI (Area of Interest) segmentation.

- 5.1 Log onto GeoMx software and start with "New / Continue Run".
- 5.2 After loading slides with the collection plate information, DSP is ready to scan slides.
- 5.3 A slide scan name is created and panel/visualization marker information is selected as below: In the **Probe Reagent Kit** field, in any order select **Human Immune Cell Profiling Protein Core**, **Human Immune Activation Status Protein**, **Human IO Drug Target Protein**, **Human Cell Death Protein**, **Human MAPK Signaling Protein**, **Human Pan-Tumor Protein**, **Human Pl3K/AKT Signaling Protein**, and **SQ-35430 OHSU Custom**
 - 1. Select the FITC/525 nm, Cy5/568 nm and Texas Red/615 nm channels.
 - 2. For **FITC/525 nm, select SYTO 13** as fluorophore, DNA as biological target and enter 50 as exposure time.
 - 3. For **Cy5/568 nm, select Alexa 532 as fluorophore, PanCK** as biological target and enter 300 as exposure time.
 - 4. For **Texas Red/615 nm, select Alexa 594 as fluorophore, CD45** as biological target and enter 300 as exposure time.
 - 5. Select FITC/525 nm as focus channel
- 5.4 When the scan area for each slide has been adjusted with sensitivity setting, select **Scan**.
- 6 ROI selection and AOI segmentation
- 6.1 After scanning is done, each color channel intensity is adjusted to show visualization markers along with tissue or cell line property.
- 6.2 Each ROI is determined and selected by pathologist's guide, and drawn with circle (maximum 660um radius), rectangle (maximum 660x785um) or polygonal shape (maximum 660x785um).

10m

30m

10m



TNP TMA slide contains total 88 cores.

Due to the limitation of scan area in the slide loading slot, only 44 cores per slide can be scanned and collected.

Two TMA slides (88 cores were embedded and slightly shifted to either left or right side of slide to cover half of 88

cores in each slide) are required to collect all cores.

6.3 In the segment menu, 2 segmentation classes (Tumor, stromal and others) are added and parameters are set in the following order:

10m

For Tumor segmentation (Segment 1), Alexa 532 (PanCK) is set to positive ("+") and the others set to ignore ("0") for the tumor collection

For stromal segmentation (Segment 2), Alexa 532 (PanCK), Alexa 594 (CD45) and FITC 525 (SYTO 13) were set to ignore ("0") to collect all other regions

Then click **Generate Segments**.

Once all segments are automatically generated, each channel parameter needs to be manually re-adjusted with pathologist's input to confirm if the segmentation is correctly done.

1h 30m

Caution: less than 20 cells in each segment is removed from collection due to threshold for low signal.

Once all AOI segmentation is complete, the "Exit Scan Workspace" button icon is clicked to approve ROI selection and samples are collected in a 96-well plate.

3h

Hybridization

(18h 10m)

- 7 Hybridization
- 7.1 After completion of sample collection, the 96-well plate is finalized, removed from the DSP and transferred to the PCR thermocycler.

1h

DNA oligo samples in 96-well plate are completely dehydrated at 60 °C for 01:00:00 in a PCR thermocycler.

Caution: the plate lid of the PCR thermocycler should be opened completely to avoid the contamination of DNA oligo from evaporation during this drying step.



7.2	Add $\ \ \ \ \ \ \ \ \ \ \ \ \ $	30m
7.3	(1) Add and mix well Δ 12 μL of each Probe A per panel to a 1.5ml Eppendorf tube (labeled Probe A tube) Since 7 commercial panels and 1 custom panel (consists of 2 Probe A) are used, the total volume is Δ 108 μL (= 12ul x 9 Probe A) in the Probe A tube. (2) Add and mix well Δ 12 μL of Probe B (Universal probe) with Δ 87 μL of DEPC water in a fresh 1.5ml Eppendorf tube (labeled Probe B tube, total volume is Δ 99 μL).	10m
	(3) Calculate and add \blacksquare 80 μ L of hybridization buffer per row into a fresh 1.5ml Eppendorf tube (labeled Master mix tube). For 44 core samples, we collected a total of 8 rows in the collection plate so a total of \blacksquare 640 μ L of hybridization buffer was added (80 x 8 rows).	
	The Probe/Master Mix is made by adding $\ \ \ \ \ \ \ \ \ \ \ \ \ $	
7.4	The Hybridization CodeSet for each row (A-H) is taken out from \$\mathbb{L}^* -80 \circ C\$ and thawed at RT \$\mathbb{L}\$ 84 \mu L of Probe / Master mix is added to each hybridization Codeset and mixed well.	10m
7.5	Prepare the hybridization plate separately (96-well plate)	20m
7.6	Seal the hybridization plate with aluminum foil using the microplate heat sealer and briefly spin it down in a centrifuge at 2000g.	
7.7	Samples are hybridized at 67 °C for 16:00:00 in the PCR Thermocycler and then kept at 4 °C until loading in the nCounter MAX system.	16h

nCounter reading

5h 45m



- 8 nCounter preparation and reading
- 8.1 In the DSP server, the collection plate is finalized and the library preparation file for nCounter loading is downloaded.

15m

The sample loading volume is determined (it will vary) according to the area size calculation of total ROI/AOI in each column.

- The Samples from each row (A-H) are collected and transferred into a 12-well strip PCR tube using a 12-channel multi-pipette.
- 8.2 Briefly, the samples are spun down in the 12-well PCR strip and then transferred to Nanostring's MAX Prep-station to load samples onto the cartridge with the standard sensitivity setting.

3h

8.3 The cartridge is transferred to nCounter to read the counts with defined CDF (CodeSet Design **Form)** setting downloaded from the DSP plate information.

2h 30m

8.4 The RCC (Reporter Code Count) file from nCounter is downloaded to your PC and imported into the DSP server.

DSP data analysis

- 9 QC DSP data and analysis
- 9.1 Select and queue the slides to analyze using "New Analysis" in the DSP server.
- 9.2 Determine the New Analysis file name and save it in the designated folder.
- 9.3 Open an analysis file and perform the **QC** with preset parameters.
- 9.4 QC passed samples are processed and the QC file (CSV) is downloaded to a PC for further analysis.

Comment: All documents related with GeoMx DSP run can be found at https://university.nanostring.com/page/document-library