

SEP 23, 2023

Microscopy-based bead assay

Elias Adriaenssens¹

¹Sascha Martens lab, University of Vienna, Max Perutz Labs - Vienna

ASAP Collaborative Research Network

Sascha Martens lab, University of Vienna, Max Perutz Labs - Vienna



Elias Adriaenssens

Sascha Martens lab, University of Vienna, Max Perutz Labs - ...

ABSTRACT

This protocol describes the microscopy-based bead assay.

MATERIALS

Materials

- Glutathione Sepharose 4B beads (GE Healthcare)
- dH₂0
- MgCl₂
- FIP200-GFP
- SINTBAD-mCherry
- TBK1
- SINTBAD-GFP
- mCherry-OPTN
- Zeiss LSM 700 confocal microscope

Bead assay buffer

Α	В
Tris-HCl pH 7.4	25 mM
NaCl	150 mM
DTT	1 mM

OPEN ACCESS



DOI:

dx.doi.org/10.17504/protocol s.io.14egn38pzl5d/v1

Protocol Citation: Elias Adriaenssens 2023. Microscopy-based bead assay. protocols.io https://dx.doi.org/10.17504/p rotocols.io.14egn38pzl5d/v1

License: This is an open access protocol distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working We use this protocol and it's working

Created: Jul 07, 2023

Last Modified: Sep 23,

2023

PROTOCOL integer ID: 84651

04031

Keywords: Microscopybased bead assay

Microscopy-based bead assay

6h 30m

We use Glutathione Sepharose 4B beads (GE Healthcare) to bind GST-tagged bait proteins.

Oct 23 2023

buffer. 3 Then, resuspend beads in 🔼 40 µL bead assay buffer, to which bait proteins are added at a final concentration of [м] 5 micromolar (µМ) Incubate the Beads with the bait proteins for 50 01:00:00 at \$\mathbb{4}\$ of at a horizontal tube roller. 5 Wash the beads three times to remove unbound GST-tagged bait proteins and resuspend in A 30 µL bead assay buffer. Where needed, also add MgCl₂ and ATP to the buffer to allow the phosphorylation of targets by TBK1 or other kinases. Prepare glass-bottom 384-well microplates (Greiner Bio-One) with 🚨 20 µL samples containing prey proteins at the concentrations described below and dilute in bead assay buffer, and add 🔼 3 µL of beads per well. 8 For the experiments in which full-length FIP200-GFP is recruited to GST-4xUb beads in presence of NDP52 and/or SINTBAD-mCherry, use NDP52 at a final concentration of [M] 50 nanomolar (nM), FIP200-GFP, SINTBAD-mCherry, and TBK1 are used at a final concentration of [м] 100 nanomolar (nM) 9 For recruitment of SINTBAD-GFP to GST-LC3/GABARAP beads, use a final concentration of [M] 200 nanomolar (nM) SINTBAD-GFP. 10 For the TBK1-binding competition experiment between OPTN and NAP1, use mCherry-OPTN and GFP-TBK1 at a final concentration of [м] 500 nanomolar (nM) , and use NAP1 from [м] 100 nanomolar (nM) to [м] 10 micromolar (µM) 11 Incubate the beads with the prey proteins for 00:30:00 prior to imaging, with the exception of experiments where full-length FIP200 5h 30m is recruited, where proteins are co-incubated for 04:00:00 before imaging. For recruitment of autophagy components to GST-LC3/GABARAP beads, proteins are co-incubated for 01:00:00 before imaging.

12

Image samples with a Zeiss LSM 700 confocal microscope equipped with Plan Apochromat 20X/0.8 WD 0.55 mm objective.

Note

Three biological replicates were performed for each experimental condition.

For the quantification, we employ an artificial intelligence (AI) script that automatically quantifies signal intensities from microscopy images by drawing line profiles across beads and recording the difference between the minimum and maximum grey values along the lines.

Note

The AI was trained to recognize beads employing cellpose [60].

- Processing is composed of two parts, with the first operating in batch mode. Multichannel input images are split into individual TIFF images and passed to cellpose (running in a Python environment).
- 15 The labeled images produced by cellpose are re-assembled into multichannel images.
- 16 Circular regions of interest (ROIs) are fitted to the segmented particles, and a pre-defined number of line profiles (here set to 20) are drawn automatically, starting at the center of the ROI and extending beyond the border of the circular ROI.
- This results in line profiles from the center of the bead into the inter-bead space of the well, allowing us to quantify the signal intensities at the rim of the beads.
- 18 To prevent line profiles from protruding into adjacent beads, a combined ROI containing all beads is used.
- 19 Inspect the Al-generated results manually for undetected beads, incorrect line profiles, or false-assigned bead structures.
- 20 For each bead, a mean fluorescence and standard deviation are obtained based on the 20 line profiles per bead.
- 21 Either exclude or subject the beads with standard deviations equal to or greater than half the mean value to manual inspection for correction.
- 22 If needed, correct for inter-experiment variability in absolute values by dividing the mean values for each bead by the average bead

intensity of the control condition, to obtain relative values.

These values are then plotted and subjected to statistical significance calculations.