

Sep 16, 2024

## T-RFLP Sephadex purification and sample preparation

DOI

**[dx.doi.org/10.17504/protocols.io.36wgqwyogk57/v1](https://dx.doi.org/10.17504/protocols.io.36wgqwyogk57/v1)**

Eva Petrova<sup>1</sup>, Roey Angel<sup>1</sup>

<sup>1</sup>Soil and Water Research Infrastructure

SoWa RI Anaerobic and Molecular Microbiology (public)  
Tech. support email: [eva.petrova@bc.cas.cz](mailto:eva.petrova@bc.cas.cz)



Eva Petrova

Soil and Water Research Infrastructure

OPEN  ACCESS



DOI: **[dx.doi.org/10.17504/protocols.io.36wgqwyogk57/v1](https://dx.doi.org/10.17504/protocols.io.36wgqwyogk57/v1)**

**Protocol Citation:** Eva Petrova, Roey Angel 2024. T-RFLP Sephadex purification and sample preparation. **protocols.io**  
**<https://dx.doi.org/10.17504/protocols.io.36wgqwyogk57/v1>**

**License:** This is an open access protocol distributed under the terms of the **[Creative Commons Attribution License](#)**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

**Protocol status:** Working

**We use this protocol and it's working**

**Created:** May 24, 2018

**Last Modified:** September 16, 2024

**Protocol Integer ID:** 12442

### Abstract

Purification of PCR-product after restriction digest. Samples proceeded on the capillary sequencer.



## Guidelines

The Sephadex solution is stored in a glass bottle in the fridge. Before you start the purification let the Sephadex solution come to room temperature and shake well before each usage!

## Materials

### MATERIALS

☒ MultiScreen-HV Filter Plate, 0.45 µm, clear, non-sterile **Merck Millipore Catalog #MAHVN4550**

☒ Sephadex G-50 M DNA Grade, 100 g **Ge Healthcare Catalog #17004502**

## Before start


### Materials:

- digested PCR-product (100-200ng in 20ul)
- 96 well filter plates
- 96 well collection plates (IBL 96\*0.2ml frosted Subkirted thin well)
- Sephadex G50 (GE, 37.5g into 500ml autoclaved MilliQ H<sub>2</sub>O)
- Multipette 10ml/1ml
- Multichanel 200ul/10ul
- HiDi Formamide (Prepare a master mix: 10ul HiDi Formamide + 0.3ul Size Standard, GS\_1200 LIZ, AB per sample nebo 165ul HiDi Formamide + 5ul Size Standard for one run = 16 samples)



## Sephadex purification

- 1
  - Load 200µl Sephadex solution onto the 96 well filter plates (shake the Sephadex solution well before each usage!) (use the Eppendorf Multipipette with a 10ml tip), the no. of wells you load depends on the no. of samples,
  - load filter plate on 96 well collection plate,
  - place the lit on top of this 'plate sandwich' and
  - place in swing out rotor base.


 200 µL Sephadex solution


- 2 Spin at 230rcf(g) for 1 min 15sec (Eppendorf centrifuge 5810). Don't forget to balance the centrifuge by placing a dummy 'plate sandwich' on the rotor.


Command

**420 x g**

CENTRIFUGE

 00:01:00

- 3 Discard flow through.
- 4 Repeat step 2. (Shake Sephadex!)  
 go to step #2 repeat
- 5 Place Sephadex G-50 containing 96 well filter plate on the collection plate.
- 6 Load 20µl sample in the centre of the Sephadex column. Just drop the 20ul on tj top of the gel.  
**Avoid to destroy the Sd. gel by penetrating it with the pipette tip.**
- 7 Spin at 420 rcf (g) for 1 min (Eppendorf centrifuge 5810). Note: There will be some extra buffer that comes through from the Sephadex resulting in a total sample volume of ~ 35-40 µl.






 00:01:00

**Command****420 x g**

CENTRIFUGE

- 8 Transfer the purified samples back to the original caps.
- 9 Remove the Sephadex from the filter plate and rinse the filter plate and collection plate as well (with MQ). Both plates could be reused several times.
- 10 Go immediately to sample preparation or store the samples at -20°C.

## Sample preparation for capillary sequencing

- 11 Aliquot 10µl of HiDi Formamide master mix into a **new** 96 well plate.
- 12 Add the 4µl of your purified sample per well.  
 4 µL purified sample
- 13 Denature samples for 3min at 95°C.  
 00:03:00 denaturation  
 95 °C
- 14 Transfer samples immediately on ice and let the chill for 5 min.  
 0 °C on ice  
 00:05:00
- 15 Store samples in the fridge, prepare the sample sheet and run the samples at ABI 3130XL Capillary Sequencer.