

Jun 24, 2024

BAF_Protocol_012_Lipidomics LC-MS(/MS): Vanquish UPLC and Orbitrap ID-X

DOI

dx.doi.org/10.17504/protocols.io.6qpvr885olmk/v1

Nicholas Sherman¹

¹University of Virginia Biomolecular Analysis Facility Core

Biomolecular Analysis Fac...



Dilza Silva

UVA

OPEN  ACCESS



DOI: dx.doi.org/10.17504/protocols.io.6qpvr885olmk/v1

Protocol Citation: Nicholas Sherman 2024. BAF_Protocol_012_Lipidomics LC-MS(/MS): Vanquish UPLC and Orbitrap ID-X. protocols.io <https://dx.doi.org/10.17504/protocols.io.6qpvr885olmk/v1>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

We use this protocol and it's working

Created: June 05, 2024

Last Modified: June 24, 2024

Protocol Integer ID: 101282

Keywords: mass spectrometry, lipidomics, LC-MS, Orbitrap ID-X

Abstract

This protocol is the basic LC and MS running parameters for lipids or very hydrophobic molecules. Most are singly charged. Lipids can be more challenging as more mass isoforms exist.



Guidelines

These are our general settings as a starting point for untargeted lipidomics experiments. Specific adjustments need to be made to meet specific samples or type of data required.

Materials

Thermo Orbitrap ID-X - FETD1-10001

Thermo Vanquish Duo UHPLC

Sigma - Ascentis Express 90A C18 2.7 μ m, 10 cm x 2.1 mm - 53823-U

Thermo Optima 0.1% FA (formic acid) in water - LS118-4

Thermo Optima Methanol - A456-212

Thermo Optima Isopropanol - A461-4

Thermo Optima Acetonitrile- A955-4

Thermo Autosampler glass amber vials 2 mL – 6PK1655


WHEATON MicroLiter insert, 300 μ L, conical with spring Part number: 11-0000-100

Fisherbrand Standard Pipette Tips (200 μ L - Yellow) – 53503-065

SPLASH LipidoMIX Internal Standard - Part number 330707, Avanti Polar Lipids, Inc.





Prepare samples for injection

- 1 Add 10 μL of Splash Mix standard to the extracted lipids. Dry samples under liquid nitrogen at 40C.
- 2 Suspend dried extracted samples with  100 μL of 1:1 IPA/ACN
- 3 Mix well using a “figure-eight” movement of the vials in the bench. Transfer the sample to a new vial with borosilicate inserts.
- 4 Make sure there is no air in the bottom of the inserts add the vials into the autosampler of the UPLC Vanquish system.
- 5 In the sequence setup view of the Xcalibur software include each sample in a row filling with: file name, sample ID, folder directory to save results, directory for the acquisition method, vial position and injection volume. Normally use 10% of each sample for injection.

LC parameters 15 min gradient

- 6



Speed limits: draw speed 1 $\mu\text{L/s}$; dispense speed 1 $\mu\text{L/s}$.
Injection wash parameters: Wash mode: both; Wash time: 10.0s; Wash speed: 10.0 $\mu\text{L/s}$.
Column temperature:  55 $^{\circ}\text{C}$
Run time:  00:33:00
Solvents: %A1: 60% ACN in 10 mM Ammonium Formate with 0.1%FA.
 %A2: 90% IPA, 10% ACN, 10 mM Ammonium Formate with 0.1% FA.
Upper limit pressure: 1500 bar
Flow = 0.260 ml/min
Equilibration: -3min, 32% B, curve = 5.
Gradient (all steps curve = 5): 0 -1.5 min 32%B; 4 min 45% B; 5 min 52% B; 8 min 58% B; 11min 66% B; 14min 70%B; 18 min 75%B; 21min 97%B; 25 min 97%B; 25.01 min 32% B, 25.01-33 min 32%B.

33m

General Instrument (Orbitrap ID-X) – Positive and Negative modes

- 7 Ion Source Type: H-ESI
Spray Voltage Positive 3500V
Spray Voltage Negative 2500V



Ion Transfer Tube Temperature  275 °C , Sheath gas 35, Aux Gas 7, Sweep
gas 1, Vaporizer Temp Temperature  350 °C
Default charge state: 1

Full scan parameters

- 8 Scan 200-2000 m/z, resolution 120K, Quadrupole isolation: True, normalized AGC 50%, RF lens 40%, Max. Inj. Time 50ms, 1 microscan, Data Type Profile, Polarity Positive or Negative (create two methods one for Positive acquisition and another for Negative acquisition)

MS/MS parameters

- 9 Min intensity 5E4, Dynamic exclusion - 1 time, 5s, 10ppm, isolation 1.5 m/z, resolution 15K, HCD NCE stepped 25, 30, 35%, Max. Inj. Time 50ms, normalized AGC 50%, 1 microscan, Data Type Centroid.