

•



Jul 09. 2020

Scanning electron microscopy (SEM) protocol for imaging living materials

Hans Priks¹, Tobias Butelmann¹

¹Institute of Technology, University of Tartu

1 Works for me dx.doi.org/10.17504/protocols.io.bekcjcsw

Hans Priks

ABSTRACT

Scanning electron microscopy (SEM) can be used to image cells and colonies immobilized inside hydrogels after supercritical carbon dioxide (CO_2) extraction. Supercritical CO_2 extraction can also be used on suspension cells after filtering the sample onto a 0.2 μ M filter attached into the extractors carriers. This protocol gives an overview on how different techniques can be used to characterize triblock copolymer hydrogels.

EXTERNAL LINK

https://doi.org/10.1101/2020.03.25.004887

THIS PROTOCOL ACCOMPANIES THE FOLLOWING PUBLICATION

Priks, H.; Butelmann, T.; Illarionov, A.; Johnston, T. G.; Fellin, C.; Tamm, T.; Nelson, A.; Kumar, R.; Lahtvee, P.-J. Physical Confinement Impacts Cellular Phenotypes within Living Materials. ACS Appl. Bio Mater. 2020. https://doi.org/10.1021/acsabm.0c00335.

DOI

dx.doi.org/10.17504/protocols.io.bekcjcsw

PROTOCOL CITATION

Hans Priks, Tobias Butelmann 2020. Scanning electron microscopy (SEM) protocol for imaging living materials. **protocols.io**

dx.doi.org/10.17504/protocols.io.bekcjcsw

MANUSCRIPT CITATION please remember to cite the following publication along with this protocol

Priks, H.; Butelmann, T.; Illarionov, A.; Johnston, T. G.; Fellin, C.; Tamm, T.; Nelson, A.; Kumar, R.; Lahtvee, P.-J. Physical Confinement Impacts Cellular Phenotypes within Living Materials. ACS Appl. Bio Mater. 2020. https://doi.org/10.1021/acsabm.0c00335.

EXTERNAL LINK

https://doi.org/10.1101/2020.03.25.004887

KEYWORDS

SEM, living materials, hydrogels, triblock copolymers, immobilized microbial cells, yeast, polymer degradation, supercritical carbon dioxide extraction, scanning electron microscopy

LICENSE

This is an open access protocol distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

CREATED

Apr 03, 2020

 LAST MODIFIED

Jul 09, 2020

PROTOCOL INTEGER ID

35172

MATERIALS TEXT

Reagents:

- formaldehyde
- Ethanol, 99.5 100 %
- MilliQ water, sterile
- 0.2 M phosphate buffer (20.44 g of Na2HPO4 and 6.72 g of NaH2PO4 per litre)
- liquid N₂

Supplies:

- falcon tubes/ependorfs/glass vials
- scalpels
- 12.5 mm aluminum SEM pin stubs
- conductive double sided carbon tabs/tape
- sharpie marker

SAFETY WARNINGS

Formaldehyde (FA) is toxic and should handled accordingly. Wear protective gear!

N₂ cooled scalpels can break during sample cutting. Wear protective eyewear!

Supercritical CO_2 extraction involves high pressure. Do not leave extractor unattended while chamber temperature is rising.

2 Submerge sample into the fixation solution (M3.7 % volume formaldehyde in M10.1 Molarity (M) phosphate buffer) 1 Prepare fixation solution (M3.7 % volume formaldehyde in M10.1 Molarity (M) phosphate buffer) 2 Submerge sample into the fixation solution and incubate at & Room temperature for © 24:00:00 1 d Sample Dehydration 1d 0h 30m 4 Prepare ethanol (EtOH) dilutions in milli-Q water as indicated below. 3 Samples are dehydrated at & Room temperature in an ascending EtOH series (40 – 90 %, 10 % steps; 96 %, 99.5 %). Submerge sample into EtOH solution, let it incubate (minimum © 02:00:00 per step), change the EtOH solution.

protocols.io
2
07/09/2020

Citation: Hans Priks, Tobias Butelmann (07/09/2020). Scanning electron microscopy (SEM) protocol for imaging living materials. https://dx.doi.org/10.17504/protocols.io.bekcjcsw

EtOH steps:

- 1. [M]40 % volume EtOH
- 2. [M]50 % volume EtOH
- 3. [M]60 % volume EtOH
- 4. [M]70 % volume EtOH
- 5. [M]80 % volume EtOH
- 6. [M]90 % volume EtOH
- 7. [M]96 % volume EtOH
- 8. [M] 99.5 % volume EtOH ((Overnight)
- 9. [M] 99.5 % volume (for storage)

Supercritical CO₂ extraction

7h 25m

- 6 Cool the critical point dryer (E3100, Quorum Technologies) to § 15 °C with a thermostat (Proline RP 1845, LAUDA) using thermostat external temperature probe.
- 7 Connect the critical point dryer outlet to a bottle containing EtOH (half full) under fume hood (it is used to capture residues during extraction and to estimate the gas realese speed).
- 8 Open the critical point dryer and mount the samples. Close the critical point dryer according to producers instructions.
- 9 Open CO₂ inlet and fill the chamber with liquid CO₂.

2m

10 Slightly open the outlet and purge the chamber for © 00:05:00 (bubbling inside the external EtOH bottle should not be too intensive).

After purging close the outlet first then the inlet (to avoid pressure drop inside the chamber).

- 11 The chamber should be purged with fresh CO₂ 6 8 times in 30 60 min intervals (let the EtOH diffuse out of the structure and purge it out of the chamber, use shorter intervals at the beginning of this process).
 - open the CO₂ inlet, then slightly open the outlet
 - purge for **© 00:05:00**
 - close the outlet, then close the CO₂ inlet
 - repeat 6-8 times in 30 60 min intervals

protocols.io
3
07/09/2020

Citation: Hans Priks, Tobias Butelmann (07/09/2020). Scanning electron microscopy (SEM) protocol for imaging living materials. https://dx.doi.org/10.17504/protocols.io.bekcjcsw

12 Increase the thermostat temperature to § 37 °C (inlet and outlet of the chamber should be closed at this point).

Do not leave the critical point dryer unattended while the temperature is rising as the pressure can exceed the safety limit of the chamber.

Control the internal pressure so it does not exceed 110 bar by opening the chamber outlet (should be done slowly as too fast gas release can cool the reactor and turn supercritical state back to liquid state).

Adjust the outlet so that the pressure gauge stays stable around 105 bar as the temperature is rising

Leave the outlet open as it is, when § 37 °C is achieved (do not open the outlet more, as the faster gas release can cool the reactor).

Step 12 includes a Step case.

SLOW

FAST

step case

SLOW




Figure 1 Slow pressure release (A) vs. fast pressure release (B)

Leave the outlet open until the chamber is ready to be opened (© **Overnight**).

12h

As the pressure drops so does the bubbling.

 $Adjust\ the\ outlet\ so\ that\ there\ is\ always\ slight\ bubbling\ (do\ not\ over\ do\ it\ as\ it\ can\ result\ in\ pore\ formation\ figure\ 1B)$

Before opening the chamber remove outlet tube from EtOH bottle that is situated under the hood (to avoid sucking EtOH into the chamber while opening it).

 $Remove \ samples \ from \ the \ chamber \ and \ store \ in \ a \ sealable \ container \ (ependorf, \ glass \ vial, \ falcon \ tube)$

Sample cutting and mounting

15 Attache conductive double sided carbon tabs/tape on aluminum SEM pin stubs and then lable them with a sharpie 5m marker




Figure 3: Sample cutting - exposing cell-material interactions in hydrogels. Varying the sample and the blade temperature together with the speed of cutting can be used to demonstrate various aspects of LMs. A combination of sample and scalpel cooling (~20 s) together with fast incisions results in the most accurate SEM images in terms of polymeric material and colony localization (A, B), but with this technique it is impossible to evaluate the colony size and shape because of the unknown location of the obtained cross-section in respect to the colony. A shorter duration of sample and scalpel cooling (~10 s) together with slow incision highlights biologically relevant information such as cell-polymer encapsulations (Figure 5 A - C) and colony size and shape (C, D) but results in cutting marks across the polymer (D). Different sample cuttings and resulting images: samples prepared with longer cooling of sample and scalpel showing relatively smooth cuts (A, B). Samples prepared with short sample and scalpel cooling showing clear colonies (C, D).

16.1 Fast incision (Figure 3: A, B) - for acquiring artifact-free cross-sections

20s

Immerse the sample with forceps and scalpel into liquid N_2 for © 00:00:20 and instantly cut with fast incision (N_2 cooled scalpels can break during sample cutting. Wear protective eyewear!).

16.2 Slow incision (Figure 3: C, D) - for acquiring information of colony-material interactions, colony size and shape

15s

Immerse the sample with forceps and scalpel in liquid N_2 for © 00:00:10 and cut after © 00:00:03 at room temperature with slow incision.

17 Using forceps, pick up the cut sample and gently press it onto the two-sided carbon tape.

1m

Sputter Coating

18 Coat the sample with a +7.5 nm gold layer using a high vacuum sputter coater (EM ACE600, Leica Microsystems).

SEM imaging 1d

19 Gold-coated samples were imaged with a tabletop scanning electron microscope (TM3000, Hitachi).

1d

 The imaging was done under a high vacuum and 15 kV accelerating voltage.

