



Jun 16, 2021

MojoSort™ Whole Blood Human Neutrophil Isolation Kit Column Protocol V.2

Ken Lau¹¹BioLegend

Other



Share

This protocol is published without a DOI.

BioLegend

Tech. support email: tech@biolegend.com

Ken Lau

ABSTRACT

MojoSort™ Whole Blood Human Neutrophil Isolation Kit Column Protocol

Please note that BioLegend is no longer updating protocols.io. For the most up-to-date protocols, visit our website: <https://www.biolegend.com/en-us/technical-protocols>

EXTERNAL LINK

<https://www.biolegend.com/en-us/protocols/mojosort-whole-blood-human-neutrophil-isolation-kit-column-protocol>

PROTOCOL CITATION

Ken Lau 2021. MojoSort™ Whole Blood Human Neutrophil Isolation Kit Column Protocol. **protocols.io**
<https://protocols.io/view/mojosort-whole-blood-human-neutrophil-isolation-ki-bvwhn7b6>
Version created by Ken Lau



KEYWORDS

MojoSort, cell separation, magnetic beads, BioLegend, magnetic columns, nanobeads, neutrophils

LICENSE

This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

CREATED

Jun 16, 2021

LAST MODIFIED

Jun 16, 2021

PROTOCOL INTEGER ID

50857

GUIDELINES

Introduction: BioLegend MojoSort™ nanobeads work in commonly used separation columns, based on our internal research as well as validation by external testing by academic labs. This simple protocol consists of following the MojoSort™ protocol to label the cells with **pre-diluted** MojoSort™ reagents and using the columns as indicated by the manufacturer.

Note: Due to the properties of our beads, it may be possible to use far fewer beads than with other commercial suppliers. We recommend a titration to find the best dilution factor. However, as a general rule, dilutions ranging from 1:3 to 1:20 for the Nanobeads can be used. Please contact BioLegend Technical Service (tech@biolegend.com) if further assistance is needed.

Important Note: MojoSort™ magnetic particles can be used with other commercially available magnetic separators, both free standing magnets and column-based systems. Because MojoSort™ protocols are optimized for the MojoSort™ separator, the protocols may need to be adjusted for other systems. Please contact BioLegend Technical Service (tech@biolegend.com) for more information and guidance. We do not recommend using MojoSort™ particles for BD's iMag™ or Life Technologies' DynaMag™.

MATERIALS TEXT

- MojoSort™ Buffer (5X) (Cat. No. [480017](#))
- Human TruStain FcX™ (Cat. No. [422301](#)).
- Adjustable pipettes
- 70 µm filters (one per sample)
- 5 mL (12 x 75mm) or 14 mL (17 x 100 mm) polypropylene tubes
- Reagents for sample preparation
- Reagents and instruments (flow cytometer) to determine yield and purity

Protocol 55m

- Collect whole blood in collection tube that has anticoagulant, preferably EDTA.
Note: Keep MojoSort™ Buffer on ice throughout the procedure.
- Aliquot 1 mL of human whole blood into a 5 mL (12 x 75 mm) polypropylene tube. **Add 5 µL of the Biotin-Antibody Cocktail.** Mix well and **incubate on ice for 15 minutes.** Scale up or down the volume accordingly if separating more blood. For example, add 10 µL of biotinylated antibody cocktail for separating 2 mL blood. 15m
- Resuspend the beads by vortexing, maximum speed, 5 touches. **Add 10 µL of Streptavidin Nanobeads.** Mix well and **incubate on ice for 15 minutes.** Scale up the volume accordingly if separating larger starting volume of whole blood. For example, add 20 µL of pre-diluted Nanobeads for separating 2 mL of whole blood. 15m
- Wash the cells by adding MojoSort™ Buffer up to 4 mL. Centrifuge the cells at 300 x g for 5 minutes. Remove supernatant by Pipet aid instead of pouring. 5m
- Proceed to separation on column as indicated by the manufacturer. 15m
Note: There are several types of commercially available columns, depending on your application. Choose the one that fits best your experiment. *This kit requires the use of a column compatible with processing of human whole blood.*
- Column:** 5m

A	B	C	D	E
	Max. volume of whole blood	Column rinse volume	Cell wash volume	Elution volume
Whole Blood Column	15 mL	3 mL	3x3 mL	5 mL

Example of magnetic separation with whole blood columns:

- Place the column in a magnetic separator that fits the column.
- Rinse the column with 3 mL of cell separation buffer.

3. Add the labeled whole blood to the column through a 30 μ m filter and collect the fraction containing the unlabeled cells.
4. Wash the cells in the column **3 times** with 3 mL of buffer and collect the fraction containing the unlabeled cells. Combine with the collected fraction from step 3. These cells may be useful as controls, to monitor purity/yield, or other purposes.
5. Take away the column from the magnet and place it on a tube. Then add 5 mL of elution buffer and flush out the magnetically labeled fraction with a plunger or supplied device. These are the positively isolated cells of interest; do not discard.
6. (*Optional*) Lyse remaining erythrocytes using 20 mL room temperature 1X Red Blood Cell Lysis Buffer (10X Cat. No. 420301/420302). Incubate 15 min, in dark at room temperature followed by filtering through a cell strainer (40-70 μ m). Wash twice with FACS wash buffer.