




MAR 12, 2023

🌐 Preparing Agarose Gel

alyssa.jeng¹

¹University of Colorado



alyssa.jeng

ABSTRACT

N/A

OPEN  ACCESS

DOI:

dx.doi.org/10.17504/protocols.io.kqdg391q1g25/v1

Protocol Citation: alyssa.jeng 2023. Preparing Agarose Gel. **protocols.io**

<https://dx.doi.org/10.17504/protocols.io.kqdg391q1g25/v1>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

We use this protocol and it's working

Created: Mar 10, 2023

Last Modified: Mar 12, 2023

PROTOCOL integer ID:

78544

Materials

- 1
 - 1X TAE Buffer
 - SYBRsafe

- Agarose powder
- Weigh boat
- 500mL Erlenmeyer flask
- Gel Electrophoresis tray
- Gel Electrophoresis comb
- Tape

Procedure

- 2 Tape up the gel electrophoresis tray and add well comb
- 3 Add 200mL of 1X TAE buffer to the Erlenmeyer flask
- 4 Weigh the right amount of agarose powder - depends on the expected PCR products: smaller than 500 bp: 2% agarose gel – 4 gr/200mL; bigger than 500 bp: 1% agarose gel – 2 gr/200mL).
- 5 Microwave for 2-3 minutes
- 6 Carefully remove flask from microwave and add 20uL of SYBRsafe to the gel solution. Swirl to mix.
- 7 Slowly pour the mixture in to the electrophoresis tray
- 8 Let cool for at least 30 minutes (no longer than 2 hours)



9 Clean the Erlenmeyer flask to remove excess agarose gel