

#### **VERSION 2**

JAN 16, 2024

# OPEN ACCESS



#### DOI:

dx.doi.org/10.17504/protocol s.io.5jyl8p876g2w/v2

Protocol Citation: Dusan Velickovic, Chris Anderton 2024. Spatial N-glycomics with MALDI-MSI for human lung tissue. protocols.io https://dx.doi.org/10.17504/p rotocols.io.5jyl8p876g2w/v2V ersion created by Dusan Velickovic

# **MANUSCRIPT CITATION:** https://doi.org/10.1021/jasms

https://doi.org/10.1021/jasms .2c00120

License: This is an open access protocol distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

# Spatial N-glycomics with MALDI-MSI for human lung tissue V.2

Version 1 is forked from Spatial N-glycomics with MALDI-MSI for human kidney tissue

Dusan Chris

Velickovic<sup>1</sup>, Anderton<sup>1</sup>

<sup>1</sup>PNNL

**URMC Pryhuber Lab** 



Dusan Velickovic

#### **DISCLAIMER**

Protocol is adapted from "MALDI Imaging Mass Spectrometry of N-glycans and Tryptic Peptides from the Same Formalin-Fixed, Paraffin-Embedded Tissue Section"

Angel, P. M.;Mehta, A.;Norris-Caneda, K.;Drake, R. R. *Methods Mol. Biol.* **2017**, *1788*, 225–241, DOI: 10.1007/7651\_2017\_81

Protocol utilized in "Controlled Humidity Levels for Fine Spatial Detail Information in Enzyme-Assisted N-Glycan MALDI MSI"

Veličković, D.; Sharma, K.; Alexandrov, T.; Hodgin, J.B.; Anderton, C.R. *J. Am. Soc. Mass Spectrom.* **2022**, 33, 1577–1580, DOI: 10.1021/jasms.2c00120

#### **ABSTRACT**

This protocol describes the procedure to obtain high quality MALDI mass spectrometry images of N-linked glycans from formalin-fixed paraffin embedded tissue. This protocol is optimized for human lung tissue as part of the HubMAP Project.

Protocol status: Working

We use this protocol and it's working

8 Coplin jars

5 slide mailers with side opening

Created: Jan 15, 2024 pH strips (optimized for acidic pH)

**MATERIALS** 

Vegetable steamer (antigen retrieval device) - example: AROMA 8-Cup Cool-Touch

Last Modified: Jan 16, 2024 Rice Cooker

Flatbed color scanner - example: EPSON PERFECTION V500 Photo

**PROTOCOL** integer ID: Zeiss PALM MicroBeam 93558

Syringe pump capable of 25 µl/min

Keywords: spatial omics, nglycans, glycosylation, lung, MALDI imaging, glycoprotein

Pump capable of 100 µl/min Automated matrix sprayer - example: TM, M3, or M5-Sprayer for HTX imaging Home-designed chamber for incubation: Note: This consists of a rubber gasket

sealed glass container (example: the jar with lid, KORKEN, IKEAof Sweden. Diameter 11 cm; high 10.5 cm, volume 0.5 L) in which a 50 ml glass beaker and a set of

weights is placed. The weights are required to keep the glass beaker from floating.

**Funders Acknowledgement:** 

NIH Human Biomolecular Atlas Program

Grant ID: U54HL165443

NIH LungMAP

Grant ID: U01HL148860

NIH LunaMAP

Grant ID: U01HL122700

Chemicals & Enzymes:

**Xylenes** 

200 proof Ethanol

Water

Citraconic anhydrous buffer

KN03 1M HCL

PNGase F Enzyme (PRIME, 50 U/µg (lyophilized))

## Scope

1 This protocol describes the procedure to obtain high quality matrix-assisted laser desorption/ionization (MALDI) mass spectrometry images of N-linked glycans from formalin-fixed paraffin embedded tissue.

## **Health and Safety**

2 Wear nitrile gloves and safety glasses. Follow standard laboratory safety procedures.

## **Equipment**

3 **Equipment Required:** 

## 3.1 8 Coplin jars

3.2




5 slides mailer with side opening

- **3.3** pH strips (optimized for acidic pH)
- 3.4 Vegetable steamer (antigen retrieval device) example: AROMA 8-Cup Cool-Touch Rice Cooker
- 3.5 Flatbed color scanner example: EPSON PERFECTION V500 PHOTO)
- 3.6 Zeiss PALM MicroBeam
- 3.7 Syringe pump capable of 25 µl/min

3.8	Pump capable of 100 μl/min
3.9	Automated matrix sprayer – example: TM, M3, or M5-Sprayer from HTX imaging
3.10	Home-designed chamber for incubation:  Note: This consists of a rubber gasket sealed glass container (example: the jar with lid, KORKEN, IKEA of Sweden. Diameter 11 cm; high 10.5 cm, volume 0.5 L) in which a 50 ml glass beaker and a set of weights is placed. The weights are required to keep the glass beaker from floating.
	Chemicals & Enzymes
4	Chemicals & Enzymes:
4.1	Xylenes
4.2	200 proof ethanol
4.3	Water
4.4	Citraconic anhydrous buffer

- 4.5  $KNO_3$ 4.6 1 M HCl 4.7 PNGase F enzyme (PRIME, 50 U/µg (lyophilized)) 12h 30m **Procedure** 5 Formalin fixed tissue should be sectioned at 5-7 µm and mounted on indium-tin-oxide (ITO) glass slic 30m (25 mm x 75 mm). 5.1 12h Slides should be dried overnight at 37°C (in slide mailer) 10m **Antigen Retrieval Prep** 6 10m Prepare citraconic buffer: 6.1 25 ml distilled water or HPLC grade water into a 50 ml falcon tube
  - 6.2 Add 25 μl of citraconic buffer to the water



6.10	Use the Coplin jars for dewaxing and washing tissues (by submerging slide mounted tissues)	21m
6.11	Xylenes 3 minutes, repeating a total of two times.	6m
6.12	100% ethanol 1 minute, repeating a total of two times.	2m
6.13	95% ethanol 1 minute	1m
6.14	70% ethanol 1 minute	1m
6.15	Distilled water 3 minutes, repeating a total of two times.	6m
6.16	Dry slides in desiccator 5 minutes	5m
6.17	Scan each slide, minus the surrounding sample holder, at a minimum of 1200 ppi resolution usi flatbed scanner (as we use for scanning paper document). This will be needed for image regist in FlexImaging during imaging acquisition. Samples for higher resolution will require a higher resolution scanned image. For example, images acquired with a ≤ 50 µm step size require a 24 scanned image.	ration
	Obtaining high-resolution (5x-20x) brightfield images usin	g P

6.18	Turn on power supply, and key switch on PALM control unit.
6.19	Mount the slide in the slide holder and place it in the microscope
6.20	Run the PALMRobo software
6.21	In "View" Tab find "Navigation Window".
6.22	In the small screen display at the microscope, select 10x Objective and adjust focusing by turning the knob on the microscope.
6.23	In the "Navigation window" find top left corner of the tissue and select: "Set ROI top left. Next, find bottom right corner of the tissue and select "Set ROI bottom right"
6.24	Click "Scan" 30m
6.25	Save tile images after scanning is done.
	Antigen Retrieval Pren

7	Heat slides in vegetable steamer:
,	rieut sildes in vegetable steamer.
7.1	Preheat the vegetable steamer to generate steam by pressing the "cook" switch prior to retrieval procedure (example, preheating takes ~15 min)
7.2	Add ~ 10 ml of the buffer to a 5-slide mailer with side opening
7.3	Place no more than 3 slides per 5-slide mailer with side opening. Slides should be placed with tissue facing outward to the solution in positions 1 and 5, NOT facing the slide mailer walls. Position 3 may face either way
7.4	Completely fill the slide mailer the rest of the way with buffer
7.5	If the mailer has no holes punched in the lid, only snap close one corner of the mailer
7.6	Place the mailer in the corner of the vegetable steamer
7.7	Maintain "cook" option for 30 minutes
8	Cool the slides after antigen retrieval:

8.1	Remove mailer and place in a tub with cool water from the faucet. Water should not go over the top of the mailer
8.2	Allow to cool for 5 minutes 5m
8.3	Remove half the buffer from the mailer and replace with distilled water.
8.4	Allow to cool 5 minutes on countertop  5m
8.5	Repeat removal of half the buffer two more times, each time with 5 minutes of cooling
8.6	Complete by rinsing in 100% distilled water
8.7	Dry the slides 5 minutes in the desiccator
8.8	Check to ensure scanning of the slides has been performed

**8.9** For scanning, scan one slide each at 1200 dpi

## **Application of PNGase F Solution**

9	Prepare PNGase F solution:
9.1	Prepare 0.1 μg/μl PNGase F in water: resuspend lyophilized enzyme in 1 mL water
9.2	Ensure that enough solution is prepared, e.g. three full slides takes approximately 1 ml of solution, spraying at 25 µl/min
10	Spray the PNGase F solution:
10.1	Using the syringe dedicated to PNGase F enzyme solution, rinse the syringe with water by screwing in the needle tip, filling with 3 or more ml of water, and aspirating into waste
10.2	Fill with PNGase F solution ensuring that there are no bubbles in syringe. Tip: After loading all the PNGase F solution required, pull a small volume of air into the syringe. Gently dispense the syringe until the large air bubble is gone.
10.3	Remove the needle tip and fasten the syringe to the TM-Sprayer line used for PNGaseF. Place the syringe onto the red syringe pump. Check that the syringe head is snug against the dispense head of the syringe pump. Ensure that the diameter is set appropriate to the syringe and the rate is set at 25 µl/min. Do not start the pump at this time.

10.4	Place the samples in the matrix sprayer (e.g., TM-Sprayer) tray, fastening them with tape.
10.5	Set up TM-Sprayer, referring to the guide for the TM-Sprayer. Temperature should be set to 45°C with 15 passes, velocity of 1200, and 3 mm offset. Set temp of the stage to be 40°C
10.6	Pressure reading on the front of the TM-Sprayer should be 10 psi.
10.7	Start the syringe pump.
10.8	Use a dummy slide to check the TM-Sprayer nozzle for spraying of solution. It generally takes abou 3 minutes (100 µL) to start spraying.
10.9	Once moisture is detected on the dummy slide, press Start on the TM-Sprayer. PNGase F solution will be applied in a thin layer onto target tissue.
11	Incubation PNGase F digest: To prevent liquid from evaporating too fast and the enzyme from becoming inactive, a wet atmosphere is maintained by placing the ITO slide into a sealed incubation chamber filled with 150 ml saturated KNO $_3$ solution and pre-incubated at 37.5 °C.
11.1	After application of PNGase F onto the slide, place it on top of a 50 ml glass beaker in the incubation chamber.
11.2	Incubate 2 hours at 37.5 °C

11.3	After incubation, remove the slide from the incubation chamber and let dry in the desiccator (15 n 15m
11.4	Store the slide in a 5-slide mailer to protect the released glycans. If matrix cannot be sprayed the same day, store at -20°C. It is preferred to immediately spray matrix onto the slide.
	Application of Matrix
12	Prepare the CHCA Matrix:
12.1 12.2	Prepare CHCA matrix at 7 mg/ml in 50% acetonitrile/0.1% TFA.Add 0.042 g CHCA to 6 ml 50% acetonitrile/0.1% TFA. Prepare fresh each time in a 15 ml falcon tube.  Vortex briefly and sonicate 5 minutes.
12.3	Small chunks may remain in the bottom of the falcon tube. Make sure that there are not loaded into the TM-Sprayer loop as they will clog components of the TM-Sprayer.
12.4	Filter CHCA solution using Millex (Millipore) 0.2 µm syringe filter.
13	Spray the CHCA Matrix:
rotocols	io l

13.1 Fill the glass-5ml syringe with CHCA solution ensuring that there are no bubbles in syringe. Tip: After loading all the solution required, pull a small volume of air into the syringe. Gently dispense the syringe until the large air bubble is gone. 13.2 Remove the needle tip and fasten the syringe to the TM-Sprayer line going to the 6-port valve. 13.3 Move the switch to "LOAD" and steadily depress the syringe until all the sample is loaded. Note: Do not load air bubbles or undissolved matrix. 13.4 Make sure that the pump is flowing at 0.1 ml/minute. Pump pressure should be 30-40 psi when flowing at 0.1 ml/min. 13.5 Place the sample in the TM-Sprayer tray, fasten them with tape. 13.6 Set up TM-Sprayer referring to the guide for the TM-Sprayer. Temperature should be set to 80°C with 10 passes, velocity of 1300, and 2.5 mm offset. 13.7 Pressure reading on the front of the TM-Sprayer should be 10 psi. 13.8 Move the 6-port valve switch to "Spray".

13.9 Use a dummy slide to check the TM-Sprayer nozzle for spraying of solution. It generally takes about one minute to start spraying matrix. 13.10 Once matrix is detected as an opaque solution on the dummy slide, press Start on the TM-Sprayer. CHCA solution will be applied in a thin layer onto target tissue. 13.11 When finished, matrix coated slides may be imaged immediately or stored in a desiccator. **MALDI Imaging MS Acquisition** 14 Put the slide in the MALDI holder and load it in the FTICR-MS instrument 15 Load the method for N-glycan analysis Parameters of the method: m/z range 800-4,000; R=512k, laser power=28-32%, laser focus=minimum, laser shots=200, frequency=2,000 16 Teach plate 17 Select measurement region in FlexImaging 18 Run acquisition through FlexImaging

#### **MALDI Imaging Analysis**

	MALDI imaging Analysis
19	Open SCiLS Lab
19.1	Load data to SCILS lab
19.2	Convert data to imZML using complete spectra
20	Upload imZML files to METASPACE
20.1	Under Annotation settings select: "NGlycDB-v1" for Metabolite database
20.2	Under Annotation settings select: "+Na" as Adducts