

Oct 04, 2021

© Diagnostic accuracy of Watch-peripheral arterial tone for sleep apnea syndrome: protocol for a systematic review and meta-analysis

Masahiro Ichikawa¹, Tomoaki Akiyama¹, Yasushi Tsujimoto², Keisuke Anan³,

Tadashi Yamakawa¹, Yasuo Terauchi⁴

¹Department of Endocrinology and Diabetes, Yokohama City University Medical Center;

²Department of Nephrology and Dialysis, Kyoritsu Hospital; ³Division of Respiratory Medicine, Saiseikai Kumamoto Hospital;

⁴Department of Endocrinology and Metabolism, Graduate School of Medicine, Yokohama City University

1 Works for me Share

dx.doi.org/10.17504/protocols.io.byrepv3e

Masahiro Ichikawa

Department of Endocrinology and Diabetes, Yokohama City Univ...

DISCLAIMER

DISCLAIMER - FOR INFORMATIONAL PURPOSES ONLY; USE AT YOUR OWN RISK

The protocol content here is for informational purposes only and does not constitute legal, medical, clinical, or safety advice, or otherwise; content added to protocols.io is not peer reviewed and may not have undergone a formal approval of any kind. Information presented in this protocol should not substitute for independent professional judgment, advice, diagnosis, or treatment. Any action you take or refrain from taking using or relying upon the information presented here is strictly at your own risk. You agree that neither the Company nor any of the authors, contributors, administrators, or anyone else associated with protocols.io, can be held responsible for your use of the information contained in or linked to this protocol or any of our Sites/Apps and Services.

ABSTRACT

INTRODUCTION

Rationale

Obstructive sleep apnea (OSA) is common, and may affect approximately 1 billion adults aged 30-69 years worldwide. Of these, an estimated 425 million suffer from moderate or severe disease (1). Evidence has suggested OSA is associated with an increased incidence of hypertension, type 2 diabetes, coronary artery disease, and stroke, and early diagnosis is important to prevent these complications. The gold standard for diagnosing OSA is in-laboratory polysomnography (PSG). However, PSG is expensive, labor intensive, limited in available facilities, and can result in long waiting times(2, 3). The home sleep apnea testing (HSAT) has been developed to solve these problems. The American Academy of Sleep medicine (AASM) guidelines recommend HSAT with at least nasal pressure, chest and abdominal breath-guided plethysmography, and oximetry or peripheral arterial tonometry (PAT) with oximetry and actigraphy (4). Sleep studies are classified as Type I-IV, with Type III recording two respiratory variables, oxygen saturation, and one cardiac variable. (4)A meta-analysis of HSAT suggested that HSAT alone could not be the gold standard test for OSA due to the low sensitivity and specificity. However, the meta-analysis did not include studies that uses HSAT with PAT (5). Watch-peripheral arterial tone (WatchPAT), one of the HSATs, is a device that uses PAT. can be worn like a wristwatch and continuously records PAT, heart rate, oxygen saturation, and actigraphy at night. PAT signals are measured continuously in the peripheral arterial bed of the fingertip. WatchPAT can indirectly detect apneic/hypopneic events by monitoring PAT because PAT reflects physiological changes in pulsatile arterial volume that are modulated by sympathetic nerve activity (6). While other HSAT cannot accurately measure total sleep time, WatchPAT can accurately measure actual total sleep time (7), which may increase sensitivity. WatchPAT is simple and can be an alternative diagnostic tool to PSG if it has sufficient diagnostic accuracy. This review will investigate the diagnostic accuracy of watchPAT and determine whether watchPAT are replaceable for PSG.

OBJECTIVE

This systematic review and meta-analysis aims to evaluate the diagnostic accuracy of WatcPAT for OSA among

Citation: Masahiro Ichikawa, Tomoaki Akiyama, Yasushi Tsujimoto, Keisuke Anan, Tadashi Yamakawa, Yasuo Terauchi (10/04/2021). Diagnostic accuracy of Watch-peripheral arterial tone for sleep apnea syndrome: protocol for a systematic review and meta-analysis. https://dx.doi.org/10.17504/protocols.io.byrepv3e

patients suspected of having OSA and to determine if PSG can be replaced by WatchPAT. Secondary objective is whether WatchPAT can classify the severity of OSA according to AASM guideline.

METHODS

We will follow the Preferred Reporting Items for a Systematic Review and Meta-analysis of Diagnostic Test Accuracy Studies (PRISMA-DTA) statement (8).

Criteria for considering this review

Types of studies

We will include all studies on the diagnostic accuracy of WatchPAT for evaluating OSA. We will include the studies designed as follows: prospective or retrospective observational (cohort or cross-sectional) studies, secondary analysis of randomized controlled trials (RCTs), and diagnostic case-control studies. We will not exclude studies based on language, countries, year of publication, or publication status. We will exclude case series, case reports, case control study of two-gate design, reviews, editorials, or studies for which the true positive, true negative, false positive, and false negative cannot be calculated.

Participants

The target participants are patients suspected of having OSA based on medical and sleep history, and physical examination.

Target condition

The target condition is OSA.

Settina

We expect studies conducted in outpatient settings; however, we will accept any type of settings.

Index tests

The index test assessed in this review is WatchPAT. The findings are generally categorized as Mild OSA (AHI 5 to 14.9), moderate OSA (AHI 15 to 29.9), and severe OSA (AHI 30 or more) (9). Our primary interest is whether AHI of 5 or more determined by WatchPAT can diagnose mild or more severe OSA accurately, as compared with PSG. We also investigate if WatchPAT can classify the severity of OSA according to AASM guideline. We will also include other positive findings of each imaging modality as the authors defined. We will allow the judgment of these positive findings by any physicians in included studies.

Reference standard

The reference standard is in-laboratory PSG. For the severity using AHI, we will use the same definition as WatchPAT.

Search methods for identification of studies

Electronic searches

We will conduct a literature search to identify all published and unpublished eligible studies. We will translate non-English articles and thoroughly assess them for potential inclusion as necessary. We will search in the following electronic databases: MEDLINE via Pubmed, EMBASE via ProQuest Dialog, and Cochrane Central Register of Controlled Trials (CENTRAL). We will also search for the World Health Organization international clinical trials Platform Search Portal (ICTRP) and ClinicalTrials.gov to find ongoing and unpublished studies. We will check the reference lists of studies, including international guidelines as well as the reference lists of eligible studies and articles citing eligible studies.

Data collection and analysis

Selection of studies

The two independent reviewers (MI and TA) will screen the results of the search according to our search strategy, with initial title and abstract screening followed by full-text screening. If two reviewers disagree, they will discuss to resolve the disagreement or consult with a third reviewer.

Data extraction and management

Two independent reviewers (MI and TA) will conduct independent data extraction of the included studies using sheet including the following information.

The sheet includes the information on the first author's name, year of publication, country, study design, sample

protocols.io
2
10/04/2021

Citation: Masahiro Ichikawa, Tomoaki Akiyama, Yasushi Tsujimoto, Keisuke Anan, Tadashi Yamakawa, Yasuo Terauchi (10/04/2021). Diagnostic accuracy of Watch-peripheral arterial tone for sleep apnea syndrome: protocol for a systematic review and meta-analysis. https://dx.doi.org/10.17504/protocols.io.byrepv3e

size, setting, participant's demographic data (mean age and body mass index and sex) and information of index tests and reference standard. We will calculate true positives, false positives, true negatives, and false negatives from 2×2 table. When only receiver operating characteristic curves were reported, WebPlotDigitizer 4.4 was used to extract sensitivity and specificity information (10).

Assessment of methodological quality

Two independent reviewers (MI and TA) will independently assess the risk of bias and applicability using the QUADAS-2 tool (11). We will revise the signaling questions in the QUADAS-2 according to this review question.

Statistical analysis and data synthesis

We will compare the results of watchPAT to that of the reference standard. We will extract data for 2×2 tables of watchPAT against the reference standard from each study. We will present a summary of findings table. We will show estimates of sensitivity and specificity with 95% confidence intervals (CI) of watchPAT per study using forest plots.

We will conduct estimates of the accuracy of each study using a scatter plot in a receiver operating characteristic (ROC). We will conduct a meta-analysis for watchPAT. In the meta-analysis, we will use a bivariate model to report the summary point estimate of sensitivity and specificity with 95% CI. For better interpretation of our result, we will also tabulate the number of false negatives and false positives when low-, moderate-, and high-prevalence of OSA are applied to the pooled sensitivity and specificity. We will perform all analyses using Stata version 16.1 for Windows (Stata Corporation, College Station, TX, USA) and Review Manager 5.4.1 (Cochrane Collaboration, London, United Kingdom).

Investigations of heterogeneity

Whenever possible, we will conduct subgroup analyses on the following groups. 1. Type of inspection location (home versus laboratory) (5)2. Type of patients (patients with significant comorbidities as defined by AASM guideline versus patients without significant comorbidities (4) 3. Type of risk of OSA (patients with signs suggestive of moderate to severe OSA as defined by AASM guideline (4) versus patients without signs). 4. Type of the definition of a hypopnea event (Changes in flow associated with 3% oxygen deprivation or cortical arousal versus Changes in flow associated with 4% oxygen deprivation without considering cortical arousal (4) or the author's own definition). 5. Type of index test (WatchPAT 100 versus WatchPAT 200). To examine the subgroup differences, we will perform a meta-regression, and present the relative diagnostic odds ratios and the 95% confidence intervals. If the number of studies is limited and the above analysis is not possible, we will present each SROC curve for each subgroup and qualitatively assess the heterogeneity.

Sensitivity analysis

We will assess the robustness to explore diagnostic accuracy for watchPAT by excluding studies judged to be at high risk of bias in the domains of index test or reference standard in QUADAS-2 assessment.

ATTACHMENTS

protocol ver.final.docx

DOI

dx.doi.org/10.17504/protocols.io.byrepv3e

PROTOCOL CITATION

Masahiro Ichikawa, Tomoaki Akiyama, Yasushi Tsujimoto, Keisuke Anan, Tadashi Yamakawa, Yasuo Terauchi 2021. Diagnostic accuracy of Watch-peripheral arterial tone for sleep apnea syndrome: protocol for a systematic review and meta-analysis. **protocols.io**

https://dx.doi.org/10.17504/protocols.io.byrepv3e

KEYWORDS

sleep apnea syndrome, watchPAT, peripheral arterial tonometry, polysomnography

protocols.io
3
10/04/2021

LICENSE

______ This is an open access protocol distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

CREATED

Oct 04, 2021

LAST MODIFIED

Oct 04, 2021

PROTOCOL INTEGER ID

53766