



Feb 28, 2021

Protein-A_Protein-LA Sandwich ELISA

Angel A Justiz-Vaillant¹¹University of the West Indies St. Augustine**1** Works for me dx.doi.org/10.17504/protocols.io.bsu6neze

Angel Justiz-Vaillant

University of the West Indies St. Augustine

SUBMIT TO PLOS ONE

ABSTRACT

This ELISA was used to study the interactions between Staphylococcal protein-A (SpA) and protein-LA (SpLA) with different immunoglobulin preparations of mammalian and avian species.

DOI

dx.doi.org/10.17504/protocols.io.bsu6neze

PROTOCOL CITATION

Angel A Justiz-Vaillant 2021. Protein-A_Protein-LA Sandwich ELISA. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.bsu6neze>

LICENSE

————— This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

CREATED

Feb 28, 2021

LAST MODIFIED

Feb 28, 2021

PROTOCOL INTEGER ID

47742

- 1 This ELISA was used to study the interactions between Staphylococcal protein-A (SpA) and protein-LA (SpLA) with different immunoglobulin preparations of mammalian and avian species. The 96 well microtiter plate was coated overnight at 4°C with 2 µg/µl per well of SpA in carbonate-bicarbonate buffer pH 9.6.
- 2 The plate was then treated with bovine serum albumin solution and washed 4X with PBS-Tween. 50 µl of immunoglobulins (1 mg/ml) is added and incubated for 1h at room temperature, and the microplate is rewashed 4X with PBS-Tween.
- 3 Then, 50 µL of peroxidase-labeled SpLA conjugate diluted 1:5000 in PBS-non-fat milk was added to each well and incubated for 1h at RT. The plate was washed 4X with PBS-Tween.
- 4 Then, 50 µL of o-phenylenediamine solution (4 mg/mL) was added, and the plate was incubated for 15 min at RT in the dark. The reaction was stopped with 50 µL of a 3M H2SO4 solution.

- 5 The plate was visually assessed for color development and read on a microplate reader at 492 nm. The cut-off value was set to 0.32.