



AUG 14, 2023

## CAMbank: cfDNA BCT Field Processing v1

Eliah G

Overbey<sup>1,2</sup>, Krista A Ryon<sup>1</sup>, jak<sup>1</sup>, chm2042<sup>1,2</sup>

<sup>1</sup>Weill Cornell Medicine; <sup>2</sup>BioAstra



Eliah G Overbey

Weill Cornell Medicine, BioAstra

### ABSTRACT

Field processing of cfDNA BCTs for the Cornell Aerospace Medicine Biobank (CAMbank).

Instructions for preserving: plasma and RBC Pellets.

### MATERIALS

Tube Type: Streck Cell-Free DNA BCT<sup>®</sup> RUO (Streck: #230470)

OPEN  ACCESS



**Protocol Citation:** Eliah G Overbey, Krista A Ryon, jak, chm2042 2023. CAMbank: cfDNA BCT Field Processing v1. **protocols.io**  
<https://protocols.io/view/cambank-cfdna-bct-field-processing-v1-cynsxvee>

**License:** This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

**Protocol status:** Working  
We use this protocol and it's working

**Created:** Aug 14, 2023

**Last Modified:** Aug 14, 2023

**PROTOCOL integer ID:**  
86450

**Keywords:** CAMbank,  
astronaut, SOMA, cfDNA BCT,  
Streck, plasma

## Perform Venipuncture

5m

- 1 After venipuncture, invert the tubes gently 8 to 10 times to fully mix tube anticoagulant with blood sample.

5m

Store the tube upright at room temperature until centrifugation.

## Centrifuge Settings: Plasma Separation

25m

- 2 Note: A **swing bucket** centrifuge is required.

3m

Set centrifuge:

- acceleration: 9
- deceleration: 9
- temperature: RT
- duration: 20 minutes
- speed: 300xg

- 3 Place the cfDNA BCTs in the centrifuge.

22m

Place a protective cover over the swing buckets in case of tube leakage.

Start the centrifuge.

Stand by the centrifuge until the centrifuge reaches max speed. Listen for signs of imbalance or compromised tube integrity.

## Separate Plasma

7m

- 4 Carefully remove cfDNA BCTs from the centrifuge and inspect for separation of red blood vs plasma layers.

2m

Transfer cfDNA BCTs to a sanitized laminar flow hood.

- 5 Remove the upper plasma layer and transfer to a new 15mL conical tube.

5m

Cap the conical tubes and return to the centrifuge.

## Centrifuge Settings: Plasma Purification

15m

- 6 Note: A **swing bucket** centrifuge is required.

3m

Set centrifuge:

- acceleration: 9
- deceleration: 9
- temperature: RT
- duration: 10 minutes
- speed: 5000xg

- 7** Place the conical tubes in the centrifuge. 12m
- Place a protective cover over the swing buckets in case of tube leakage.
- Start the centrifuge.
- Stand by the centrifuge until the centrifuge reaches max speed. Listen for signs of imbalance or compromised tube integrity.

## Freeze Red Blood Cell Pellets 5m

- 8** Recap and freeze the cfDNA tubes to preserve DNA in the red blood cell pellet 5m

## Aliquot Plasma 17m

- 9** Carefully remove conical tubes from centrifuge. 2m
- Transfer conical tubes to a sanitized laminar flow hood.
- 10** Using a P1000, carefully aliquot plasma into 2D barcoded tubes at **500µL each**. 15m
- Return 2D tubes with plasma to the rack and place in the -80C freezer.