

Sep 29, 2025

UroMOCA and StimPod Daily Testing in Pig

DOI

dx.doi.org/10.17504/protocols.io.dm6gp3d2pvzp/v1

Dennis Bourbeau^{1,2}, Brett Hanzlicek³

¹MetroHealth Medical Center, Cleveland, Ohio; ²Louis Stokes VA Hospital, Cleveland, Ohio;

³Advanced Platform Technology Center, Louis Stokes VA Hospital, Cleveland, Ohio



Brett Hanzlicek

OPEN  ACCESS



DOI: <https://dx.doi.org/10.17504/protocols.io.dm6gp3d2pvzp/v1>

Protocol Citation: Dennis Bourbeau, Brett Hanzlicek 2025. UroMOCA and StimPod Daily Testing in Pig. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.dm6gp3d2pvzp/v1>

License: This is an open access protocol distributed under the terms of the **Creative Commons Attribution License**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

We use this protocol and it's working

Created: September 27, 2023

Last Modified: September 29, 2025

Protocol Integer ID: 88485

Keywords: stimpod daily testing in pig, daily testing of stimpod, stimpod daily testing, data from the stimpod, stimpod, pig, uromoca, daily testing, protocol, session

Abstract

This protocol is for the daily testing of StimPods implanted in a pig. Data from the StimPods and UroMOCA are recorded during these sessions.



Tolerance Testing

- 1 Attach BluMOCA receiver/antenna to harness/jacket
- 2 Charge StimPods, then confirm wireless connection to computer
- 3 Wake up UroMOCA
- 4 Stimulation permutations
 - 4.1 20 Hz, 500 Hz, 10 kHz
 - 4.2 Stim0, Stim1
- 5 Stimulation duration: 1-5 seconds ON, >30 s rest
- 6 For each stimulation permutation, determine upper limit of stimulus intensity/amplitude without distress or discomfort to animal
- 7 Document subject's responses

Recruitment Curve

- 8 Attach BluMOCA receiver/antenna to harness/jacket
- 9 Charge StimPods, then confirm wireless connection to computer
- 10 Wake up UroMOCA



11 Administer stimulation while recording bladder pressure and EMG

11.1 Stimulus duration: 5 seconds on, > 30 s rest

11.2 Stimulus amplitude range: [0:25 μ A:tolerance limit]

11.3 Stimulus amplitudes randomized within range

11.4 For each stimulation permutation

11.5 Within tolerance range for that permutation

12 Document subject's responses

Simultaneous Stimulation

13 Attach BluMOCA receiver/antenna to harness/jacket

14 Charge StimPods, then confirm wireless connection to computer

15 Wake up UroMOCA

16 Stimulus duration: 5-10 seconds on

17 Amplitude for 20 Hz stimulation: 90% of full recruitment



- 18 Amplitude for 500 Hz or 10 kHz stimulation: minimum amplitude (threshold) to achieve maximum EMG reduction
- 19 Measure bladder pressure and EMG
- 20 Document subject's responses