

May 16, 2024 Version 3

Nextera XT at 0.2X On the Mantis V.3

DOI

dx.doi.org/10.17504/protocols.io.j8nlk57y1l5r/v3

Allen Institute for Brain Science¹

¹Allen Institute

Allen Institute for Brain Science
Tech. support
[Click here to message tech. support](#)



Allen Institute

Allen Institute

OPEN  ACCESS



DOI: **dx.doi.org/10.17504/protocols.io.j8nlk57y1l5r/v3**

Protocol Citation: Allen Institute for Brain Science 2024. Nextera XT at 0.2X On the Mantis. **protocols.io**

<https://dx.doi.org/10.17504/protocols.io.j8nlk57y1l5r/v3>Version created by **[Allen Institute](#)**

License: This is an open access protocol distributed under the terms of the **[Creative Commons Attribution License](#)**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

We use this protocol and it's working

Created: January 12, 2021

Last Modified: May 16, 2024

Protocol Integer ID: 99989

Keywords: Nextera XT, Nextera, Formulatrix Mantis, Mantis

Abstract

Protocol to generate Nextera libraries using 0.2x reagents in a 96-well PCR plate using the Formulatrix Mantis instrument.

Note: Research reported in this publication was supported by the National Institute Of Mental Health of the National Institutes of Health under Award Number U19MH114830. The content is solely the responsibility of the authors and does not necessarily represent the official views of the National Institutes of Health.

Attachments



HT_031_Mantis_Chip_C..

31KB



HT_043_How_To_Clean

17KB



MB0196_Nextera_XT_at..

1.9MB

Materials

- 1.1. Associated Library paper work
- 1.2. 96-well plate with normalized cDNA at 50 pg/μl (Stored at 4°C)
- 1.3. Library plate labels
- 1.4. 5x Eppendorf twin.tec® 96-well PCR Plates (VWR 47744-106)
- 1.5. 4x Formulatrix molecular grade HV chips for Mantis instrument (MCHSVM96)
 1. Filtered pipette tips (GP-L200F, GP-L1000F)
- 1.7. Nextera XT DNA Sample Preparation Kit for 96 Samples (Illumina FC-131-1096)
 - 1.7.1. Tagment DNA Buffer (TD) (15027866)
 - 1.7.2. Amplicon Tagmentation Mix (ATM) (15031561)
 - 1.7.3. Nextera PCR Master Mix (NPM) (15027920)
 - 1.7.4. Neutralize Tagment Buffer (NT) (15031559)
- 1.8. Microseal 'B' Adhesive seal (BioRad MSB1001)
- 1.9. VWR Aluminum Foil seals for PCR, Sterile (VWR 60941-076)
- 1.10. Index primers combined at 2.5 μM each in 96 well plate format.
- 1.11. Small Kimwipes (VWR 21905-026)

